

LB92263

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/19/2017 16:21

Test: Phenolics-mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.982	0.0	0.117	
ICB1	0.001	0.0	0.014	
CCV1	1.032	0.0	0.123	
CCB1	0.002	0.0	0.014	
PB105180BL	0.001	0.0	0.014	
PB105180BS	0.979	0.0	0.117	
I6981-01	-0.002	0.0	0.013	
I6981-01DUP	-0.000	0.0	0.014	
I6981-01MS	0.858	0.0	0.104	
I6981-01MSD	0.910	0.0	0.110	
I6981-04	0.003	0.0	0.014	
I6981-07	0.002	0.0	0.014	
I6981-10	0.002	0.0	0.014	
CCV2	1.014	0.0	0.121	
CCB2	0.004	0.0	0.014	

N	15
Mean	0.386
SD	0.4889
CV%	126.69

AS.

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 19 15:42:30 2017

Tue Dec 19 16:18:41 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0513	mg/l	12/19/2017 15:42:30	
0.1PPM	A	Phenolics-	P	0.0994	mg/l	12/19/2017 15:42:31	
0.25PPM	A	Phenolics-	P	0.2513	mg/l	12/19/2017 15:42:32	
0.50PPM	A	Phenolics-	P	0.4996	mg/l	12/19/2017 15:42:33	
1.0PPM	A	Phenolics-	P	0.997	mg/l	12/19/2017 15:42:34	
2.0PPM	A	Phenolics-	P	2.0014	mg/l	12/19/2017 15:42:35	
ICV1	S	Phenolics-	P	0.9816	mg/l	12/19/2017 16:11:49	
ICB1	S	Phenolics-	P	0.0014	mg/l	12/19/2017 16:11:50	
CCV1	S	Phenolics-	P	1.0318	mg/l	12/19/2017 16:11:52	
CCB1	S	Phenolics-	P	0.002	mg/l	12/19/2017 16:11:53	
PB105180BL	S	Phenolics-	P	0.0008	mg/l	12/19/2017 16:11:55	
PB105180BS	S	Phenolics-	P	0.979	mg/l	12/19/2017 16:11:56	
I6981-01	S	Phenolics-	P	-0.0016	mg/l	12/19/2017 16:11:57	
I6981-01DUP	S	Phenolics-	P	0	mg/l	12/19/2017 16:11:58	
I6981-01MS	S	Phenolics-	P	0.8584	mg/l	12/19/2017 16:11:59	
I6981-01MSD	S	Phenolics-	P	0.9103	mg/l	12/19/2017 16:18:34	
I6981-04	S	Phenolics-	P	0.0031	mg/l	12/19/2017 16:18:35	
I6981-07	S	Phenolics-	P	0.0018	mg/l	12/19/2017 16:18:36	
I6981-10	S	Phenolics-	P	0.0022	mg/l	12/19/2017 16:18:37	
CCV2	S	Phenolics-	P	1.0139	mg/l	12/19/2017 16:18:38	
CCB2	S	Phenolics-	P	0.0036	mg/l	12/19/2017 16:18:39	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/19/2017 15:44


Test Phenolics-

Accepted 12/19/2017 15:44

Factor 9.457
Bias 0.014

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.019	0.0513	0.0500	
2	Phenol-2	0.024	0.0994	0.1000	
3	Phenol-2	0.040	0.2513	0.2500	
4	Phenol-2	0.066	0.4996	0.5000	
5	Phenol-2	0.119	0.9970	1.0000	
6	Phenol-2	0.225	2.0014	2.0000	