



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92264

Temp.: 18.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	12/19/2017	17:00
2	CAL2	1	Water			7.01	12/19/2017	17:01
3	CAL3	1	Water			10.02	12/19/2017	17:02
4	ICV	1	Water			7.00	12/19/2017	17:04
5	CCV1	1	Water			2.02	12/19/2017	17:05
6	I6999-01	1	Water			7.96	12/19/2017	17:06
7	I6999-01DUP	1	Water			7.97	12/19/2017	17:07
8	CCV2	1	Water			12.02	12/19/2017	17:09

LB92264

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6999

WorkList ID : 106750

Date : 12/19/2017 4:45:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2017	Water	I6999-01- 1 ²	pH	Cool 4 deg C	WAYN02	F12	WASTEWATER	12/19/2017	SM 4500-PH B

Date/Time

12-19-17 4:45 PM

Received by:

JD

Relinquished by:

Date/Time

12-19-17 5:30 PM

Received by:

CP

Relinquished by:

JD