



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92271

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/19/2017	18:35
2	CAL2	1	Water			7.01	12/19/2017	18:36
3	CAL3	1	Water			10.03	12/19/2017	18:37
4	ICV	1	Water			6.99	12/19/2017	18:39
5	CCV1	1	Water			2.01	12/19/2017	18:41
6	I6681-02	1	Solid	20.02	20	8.20	12/19/2017	18:43
7	I6681-02DUP	1	Solid	20.02	20	8.22	12/19/2017	18:45
8	I6681-04	1	Solid	20.02	20	8.60	12/19/2017	18:47
9	I6681-06	1	Solid	20.02	20	8.57	12/19/2017	18:49
10	I6681-08	1	Solid	20.02	20	7.86	12/19/2017	18:51
11	I6988-03	1	Solid	20.02	20	8.31	12/19/2017	18:53
12	I6990-02	1	Solid	20.02	20	11.80	12/19/2017	18:55
13	I6993-03	1	Solid	20.02	20	10.75	12/19/2017	18:57
14	I6993-04	1	Solid	20.02	20	8.57	12/19/2017	18:59
15	I6995-03	1	Solid	20.02	20	10.95	12/19/2017	19:02
16	CCV2	1	Water			2.01	12/19/2017	19:04
17	I6995-04	1	Solid	20.02	20	8.20	12/19/2017	19:06
18	CCV3	1	Water			12.01	12/19/2017	19:08

JB 9227)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6990

WorkList ID : 106747

Date : 12/19/2017 2:41:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2017	Solid	I6681-02	Corrosivity	Cool 4 deg C	PSEG05	F52	PL-01-121917-A	12/19/2017	9045C
12/19/2017	Solid	I6681-04	Corrosivity	Cool 4 deg C	PSEG05	F52	PL-01-121917-B	12/19/2017	9045C
12/19/2017	Solid	I6681-06	Corrosivity	Cool 4 deg C	PSEG05	F52	PL-01-121917-C	12/19/2017	9045C
12/19/2017	Solid	I6681-08	Corrosivity	Cool 4 deg C	PSEG05	F52	PL-01-121917-D	12/19/2017	9045C
12/19/2017	Solid	I6988-03	Corrosivity	Cool 4 deg C	PSEG01	F52	RT4008-RT3867-RT3075	12/19/2017	9045C
12/19/2017	Solid	I6990-02	Corrosivity	Cool 4 deg C	PSEG01	F52	MH-CONCRETE	12/19/2017	9045C
12/14/2017	Solid	I6993-03	Corrosivity	Cool 4 deg C	BEMS01	F61	ROLL-OFF-3	12/14/2017	9045C
12/14/2017	Solid	I6993-04	Corrosivity	Cool 4 deg C	BEMS01	F61	ROLL-OFF-4	12/14/2017	9045C
12/19/2017	Solid	I6995-03	Corrosivity	Cool 4 deg C	PKFM01	F22	TP-1	12/19/2017	9045C
12/19/2017	Solid	I6995-04	Corrosivity	Cool 4 deg C	PKFM01	F22	TP-2	12/19/2017	9045C

Date/Time 12-19-17 2:14 PM
 Received by: J
 Relinquished by:

Date/Time 12-19-17 7:10 PM
 Received by: CD
 Relinquished by: J