

7470A


LB92272

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34692



A= 0.0000e+000

B= 2.9019e-004

C= 3.3101e-002

Rho= 0.9994053

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	22	0.000	22				
0.2	0.200	0.189	-0.011	538	0.0 %	538				
2.5	2.500	2.445	-0.055	8311	0.0 %	8311				
5.0	5.000	5.163	0.163	17676	0.0 %	17676				
7.5	7.500	7.263	-0.237	24916	0.0 %	24916				
10.0	10.000	10.100	0.100	34692	0.0 %	34692				

LB92272 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Dec 2017 11:06:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	22	-	0.00007470A	PPB		20 Dec 2017 11:09:02	S	Std
0.2 - 1	50.2	538	-	0.20007470A	PPB		20 Dec 2017 11:11:11	S	Std
2.5 - 1	52.5	8311	-	2.50007470A	PPB		20 Dec 2017 11:13:19	S	Std
5.0 - 1	55	17676	-	5.00007470A	PPB		20 Dec 2017 11:15:28	S	Std
7.5 - 1	57.5	24916	-	7.50007470A	PPB		20 Dec 2017 11:17:36	S	Std
10.0 - 1	510	34692	-	10.00007470A	PPB		20 Dec 2017 11:22:11	S	Std
ICV17 - 1	ICV17	12558	3.6773	-7470A	PPB		20 Dec 2017 11:24:51	U	SMPL
ICB17 - 1	ICB17	-65	0.0142	-7470A	PPB		20 Dec 2017 11:26:57	U	SMPL
CCV01 - 1	CCV01	17744	5.1822	-7470A	PPB		20 Dec 2017 11:29:04	U	SMPL
CCB01 - 1	CCB01	-71	0.0125	-7470A	PPB		20 Dec 2017 11:31:10	U	SMPL
CRA - 1	CRA	530	0.1869	-7470A	PPB		20 Dec 2017 11:33:18	U	SMPL
HighStd - 1	HighStd	34863	10.1500	-7470A	PPB		20 Dec 2017 11:35:24	U	SMPL
ChkStd - 1	ChkStd	24335	7.0949	-7470A	PPB		20 Dec 2017 11:37:31	U	SMPL
PB105184BL - 1	PBW	-115	-0.0003	-7470A	PPB		20 Dec 2017 11:39:40	U	SMPL
PB105184BS - 1	LCSW	12526	3.6680	-7470A	PPB		20 Dec 2017 11:41:49	U	SMPL
PB105184TB - 1	PB105143TB	-75	0.0113	-7470A	PPB		20 Dec 2017 11:43:56	U	SMPL
I6827-17 - 1	WC-4539-01	26	0.0406	-7470A	PPB		20 Dec 2017 11:46:04	U	SMPL
I6835-02 - 1	WC-4539-02	33	0.0427	-7470A	PPB		20 Dec 2017 11:48:13	U	SMPL
I6836-02 - 1	WC-70-01	25	0.0404	-7470A	PPB		20 Dec 2017 11:50:21	U	SMPL
I6837-02 - 1	WC-69-01	55	0.0491	-7470A	PPB		20 Dec 2017 11:52:28	U	SMPL
CCV02 - 1	CCV02	17513	5.1152	-7470A	PPB		20 Dec 2017 11:54:35	U	SMPL
CCB02 - 1	CCB02	-50	0.0186	-7470A	PPB		20 Dec 2017 11:56:41	U	SMPL
I6838-02 - 1	WC-7317-01	65	0.0520	-7470A	PPB		20 Dec 2017 11:58:49	U	SMPL
I6839-02 - 1	WC-41446854-01	26	0.0406	-7470A	PPB		20 Dec 2017 12:00:56	U	SMPL
I6840-02 - 1	WC-41446854-02	41	0.0450	-7470A	PPB		20 Dec 2017 12:03:02	U	SMPL
I6841-02 - 1	WC-01-01	51	0.0479	-7470A	PPB		20 Dec 2017 12:05:10	U	SMPL
I6842-02 - 1	WC-04-01	53	0.0485	-7470A	PPB		20 Dec 2017 12:07:17	U	SMPL
I6843-02 - 1	WC-04-02	57	0.0496	-7470A	PPB		20 Dec 2017 12:09:25	U	SMPL
I6844-02 - 1	WC-9974103-01	57	0.0496	-7470A	PPB		20 Dec 2017 12:11:32	U	SMPL
I6845-02 - 1	WC-10257-01	52	0.0482	-7470A	PPB		20 Dec 2017 12:13:41	U	SMPL
I6846-02 - 1	WC-34972326-01	35	0.0433	-7470A	PPB		20 Dec 2017 12:15:50	U	SMPL
I6847-02 - 1	WC-646507-01	85	0.0578	-7470A	PPB		20 Dec 2017 12:17:57	U	SMPL
CCV03 - 1	CCV03	17406	5.0842	-7470A	PPB		20 Dec 2017 12:20:03	U	SMPL
CCB03 - 1	CCB03	-63	0.0148	-7470A	PPB		20 Dec 2017 12:22:11	U	SMPL
I6848-02 - 1	WC-9871-01	40	0.0447	-7470A	PPB		20 Dec 2017 12:24:18	U	SMPL
I6848-02DUP - 1	WC-9871-01DUP	42	0.0453	-7470A	PPB		20 Dec 2017 12:26:25	U	SMPL
I6848-02MS - 1	WC-9871-01MS	12898	3.7760	-7470A	PPB		20 Dec 2017 12:28:33	U	SMPL
I6848-02MSD - 1	WC-9871-01MSD	12414	3.6355	-7470A	PPB		20 Dec 2017 12:30:39	U	SMPL
PB105185BL - 1	PBW	-70	0.0128	-7470A	PPB		20 Dec 2017 12:32:47	U	SMPL
PB105185BS - 1	LCSW	13701	4.0090	-7470A	PPB		20 Dec 2017 12:40:25	U	SMPL
PB105185TB - 1	PB105144TB	-50	0.0186	-7470A	PPB		20 Dec 2017 12:42:32	U	SMPL
I6676-09 - 1	OK-01-121817	68	0.0528	-7470A	PPB		20 Dec 2017 12:44:40	U	SMPL
I6682-02 - 1	NB-01-121817-A	45	0.0462	-7470A	PPB		20 Dec 2017 12:46:49	U	SMPL
I6682-04 - 1	NB-01-121817-B	36	0.0435	-7470A	PPB		20 Dec 2017 12:48:56	U	SMPL
CCV04 - 1	CCV04	17492	5.1091	-7470A	PPB		20 Dec 2017 12:51:03	U	SMPL
CCB04 - 1	CCB04	-55	0.0171	-7470A	PPB		20 Dec 2017 12:53:10	U	SMPL
I6682-06 - 1	NB-01-121817-C	-4	0.0319	-7470A	PPB		20 Dec 2017 12:55:18	U	SMPL
I6973-03 - 1	B11(0-10)-COMP	26	0.0406	-7470A	PPB		20 Dec 2017 12:57:25	U	SMPL
I6975-02 - 1	72-12000	21	0.0392	-7470A	PPB		20 Dec 2017 12:59:32	U	SMPL
I6975-04 - 1	72-11989	35	0.0433	-7470A	PPB		20 Dec 2017 13:01:40	U	SMPL
I6981-13 - 1	PHMHA-WCS	56	0.0494	-7470A	PPB		20 Dec 2017 13:03:47	U	SMPL
I6981-14 - 1	PHMHA-WCD	42	0.0453	-7470A	PPB		20 Dec 2017 13:05:55	U	SMPL
I6981-15 - 1	PHMHB-WCS	74	0.0546	-7470A	PPB		20 Dec 2017 13:08:03	U	SMPL
I6981-16 - 1	PHMHB-WCD	115	0.0665	-7470A	PPB		20 Dec 2017 13:10:11	U	SMPL
I6981-16DUP - 1	PHMHB-WCDDUP	88	0.0586	-7470A	PPB		20 Dec 2017 13:12:20	U	SMPL
I6981-16MS - 1	PHMHB-WCDMS	12653	3.7049	-7470A	PPB		20 Dec 2017 13:14:27	U	SMPL
CCV05 - 1	CCV05	17564	5.1300	-7470A	PPB		20 Dec 2017 13:16:34	U	SMPL
CCB05 - 1	CCB05	-51	0.0183	-7470A	PPB		20 Dec 2017 13:18:41	U	SMPL
I6981-16MSD - 1	PHMHB-WCDMSD	12798	3.7470	-7470A	PPB		20 Dec 2017 13:20:49	U	SMPL
I6983-02 - 1	PHMHB-WCW	-52	0.0180	-7470A	PPB		20 Dec 2017 13:22:56	U	SMPL
PB105186BL - 1	PBW	19	0.0386	-7470A	PPB		20 Dec 2017 13:25:04	U	SMPL
PB105186BS - 1	LCSW	13499	3.9504	-7470A	PPB		20 Dec 2017 13:27:11	U	SMPL
I6983-01 - 1	PHMHB-WCW	14687	4.2951	-7470A	PPB		20 Dec 2017 13:29:18	U	SMPL
I6983-01DUP - 1	PHMHB-WCWDUP	13960	4.0842	-7470A	PPB		20 Dec 2017 13:31:25	U	SMPL
I6983-01MS - 1	PHMHB-WCWMS	29978	8.7324	-7470A	PPB		20 Dec 2017 13:33:33	U	SMPL
I6983-01MSD - 1	PHMHB-WCWMSD	28474	8.2960	-7470A	PPB		20 Dec 2017 13:35:41	U	SMPL
CCV06 - 1	CCV06	17374	5.0749	-7470A	PPB		20 Dec 2017 13:37:51	U	SMPL
CCB06 - 1	CCB06	-71	0.0125	-7470A	PPB		20 Dec 2017 13:39:59	U	SMPL
I6848-02LX5 - 1		37	0.0438	-7470A	PPB		20 Dec 2017 13:42:07	U	SMPL
I6848-02A - 1		13082	3.8294	-7470A	PPB		20 Dec 2017 13:44:14	U	SMPL
I6981-16LX5 - 1		-1	0.0328	-7470A	PPB		20 Dec 2017 13:46:22	U	SMPL
I6981-16A - 1		12946	3.7899	-7470A	PPB		20 Dec 2017 13:48:29	U	SMPL
I6983-01LX5 - 1		2275	0.6933	-7470A	PPB		20 Dec 2017 13:50:36	U	SMPL
I6983-01A - 1		28406	8.2763	-7470A	PPB		20 Dec 2017 13:52:43	U	SMPL
CCV07 - 1	CCV07	17326	5.0609	-7470A	PPB		20 Dec 2017 13:54:51	U	SMPL
CCB07 - 1	CCB07	-79	0.0102	-7470A	PPB		20 Dec 2017 13:56:59	U	SMPL