



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB92274
Analysis Date: 12/20/2017
BalanceID: WC SC-4
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 12/20/2017
Extraction IN Time: 12:20
Extraction OUT Time: 13:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB92274BL	LB92274BL	WATER	2	1000	100	2.9411	2.9411	3.02	2.9411	2.9411	0	0
2	LB92274BS	LB92274BS	WATER	2	1000	100	2.8647	2.8647	3.01	2.8832	2.8832	0.0185	18.5
3	LB92274BSD	LB92274BSD	WATER	2	1000	100	2.8157	2.8157	3.04	2.8342	2.8342	0.0185	18.5
4	I6999-01	WASTEWATER	WATER	2	1000	100	2.9827	2.9827	3.05	2.9883	2.9883	0.0056	5.6
5	I7003-01	STORMDRAIN-1	WATER	2	1000	100	3.035	3.035	3.02	3.0357	3.0357	0.0007	0.7
6	I7003-01DUP	STORMDRAIN-1DUP	WATER	2	1000	100	3.0147	3.0147	3.03	3.0155	3.0155	0.0008	0.8
7	I7003-02	STORMDRAIN-2	WATER	2	1000	100	2.9952	2.9952	3.05	2.9957	2.9957	0.0005	0.5
8	I7003-02MS	STORMDRAIN-2	WATER	2	1000	100	3.1697	3.1697	3.01	3.1901	3.1901	0.0204	20.4



QC Batch# LB92274 Test: TPH Analysis Date: 12/20/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2303
pH Paper 0-14	W2287
Sodium Sulfate	EP1795
1:1 HCL	WP59249
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD	10.00 ML	WP58006

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0021 In OVEN TEMP : 70 Dessicator Time In : 14:15
1.0000 gram Balance: 1.0005 In Time: 13:40
Bal Check Time: 12:40

After Analysis

0.0020 gram Balance: 0.0018 Out OVEN TEMP: 71 Dessicator Time Out: 16:10
1.0000 gram Balance: 1.0003 Out Time: 15:15
Bal Check Time: 16:45

1392274

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I7003 WorkList ID : 106787 Date : 12/20/2017 11:49:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/26/2017	Water	I6999-01-D	TPH	Conc H2SO4 to	WAYN02	F12	WASTEWATER	12/19/2017	1664A
12/26/2017	Water	I7003-01-E, F	TPH	Conc H2SO4 to	KWAR01	F52	STORMDRAIN-1	12/19/2017	1664A
12/26/2017	Water	I7003-02-E, F	TPH	Conc H2SO4 to	KWAR01	F52	STORMDRAIN-2	12/19/2017	1664A

Date/Time 12-20-17 11:55 AM
Received by: JP
Relinquished by: JP

Date/Time 12-20-17 2:00 PM
Received by: CP
Relinquished by: JP