


245.1

LB94283
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34911



A= 0.0000e+000

B= 2.8564e-004

C= -1.3570e-002

Rho= 0.9999168

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	24	0.000	24				
0.05	0.050	0.050	-0.000	221	0.0 %	221				
0.2	0.200	0.187	-0.013	701	0.0 %	701				
2.5	2.500	2.547	0.047	8965	0.0 %	8965				
5.0	5.000	4.931	-0.069	17310	0.0 %	17310				
7.5	7.500	7.584	0.084	26599	0.0 %	26599				
10.0	10.000	9.958	-0.042	34911	0.0 %	34911				

LB92283 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Dec 2017 14:44:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	245.1	PPB	20 Dec 2017 14:46:16	S	Std
0.05 - 1	50.05	221	-	0.0500	245.1	PPB	20 Dec 2017 14:48:25	S	Std
0.2 - 1	50.2	701	-	0.2000	245.1	PPB	20 Dec 2017 14:50:33	S	Std
2.5 - 1	52.5	8965	-	2.5000	245.1	PPB	20 Dec 2017 14:52:41	S	Std
5.0 - 1	55	17310	-	5.0000	245.1	PPB	20 Dec 2017 14:54:49	S	Std
7.5 - 1	57.5	26599	-	7.5000	245.1	PPB	20 Dec 2017 14:56:57	S	Std
10.0 - 1	510	34911	-	10.0000	245.1	PPB	20 Dec 2017 14:59:05	S	Std
ICV18 - 1	ICV18	13510	3.8454	-245.1		PPB	20 Dec 2017 15:06:28	U	SMPL
ICB18 - 1	ICB18	-63	-0.0316	-245.1		PPB	20 Dec 2017 15:08:34	U	SMPL
CCV08 - 1	CCV08	17407	4.9585	-245.1		PPB	20 Dec 2017 15:10:40	U	SMPL
CCB08 - 1	CCB08	-78	-0.0359	-245.1		PPB	20 Dec 2017 15:12:47	U	SMPL
CRA - 1	CRA	696	0.1852	-245.1		PPB	20 Dec 2017 15:14:54	U	SMPL
HighStd - 1	HighStd	34678	9.8918	-245.1		PPB	20 Dec 2017 15:17:01	U	SMPL
ChkStd - 1	ChkStd	23555	6.7146	-245.1		PPB	20 Dec 2017 15:19:07	U	SMPL
PB105187BL - 1	PBW	-111	-0.0453	-245.1		PPB	20 Dec 2017 15:21:17	U	SMPL
PB105187BS - 1	LCSW	13640	3.8825	-245.1		PPB	20 Dec 2017 15:23:26	U	SMPL
I6956-01 - 1	121417-DBRI-E-D-M	0	-0.0136	-245.1		PPB	20 Dec 2017 15:25:33	U	SMPL
I6956-02 - 1	121417--DBRI-E-U-M	62	0.0041	-245.1		PPB	20 Dec 2017 15:27:41	U	SMPL
I6956-03 - 1	121417-TIDO-E-D-B	131	0.0238	-245.1		PPB	20 Dec 2017 15:29:50	U	SMPL
I6956-04 - 1	121417-TIDO-E-D-S	146	0.0281	-245.1		PPB	20 Dec 2017 15:31:58	U	SMPL
I6956-05 - 1	121417-TIDO-E-U-B	72	0.0070	-245.1		PPB	20 Dec 2017 15:34:05	U	SMPL
CCV09 - 1	CCV09	17051	4.8569	-245.1		PPB	20 Dec 2017 15:36:11	U	SMPL
CCB09 - 1	CCB09	-74	-0.0347	-245.1		PPB	20 Dec 2017 15:38:18	U	SMPL
I6956-06 - 1	121417-TIDO-E-U-S	86	0.0110	-245.1		PPB	20 Dec 2017 15:40:25	U	SMPL
I6956-07 - 1	121417-JETT-E-D-B	80	0.0093	-245.1		PPB	20 Dec 2017 15:42:32	U	SMPL
I6956-08 - 1	121417-JETT-E-D-S	76	0.0081	-245.1		PPB	20 Dec 2017 15:44:38	U	SMPL
I6956-09 - 1	121417-JETT-E-U-B	70	0.0064	-245.1		PPB	20 Dec 2017 15:46:46	U	SMPL
I6956-10 - 1	121417-JETT-E-U-S	59	0.0033	-245.1		PPB	20 Dec 2017 15:48:53	U	SMPL
I6956-10DUP - 1	121417-JETT-E-U-SDUP	79	0.0090	-245.1		PPB	20 Dec 2017 15:51:01	U	SMPL
I6956-10MS - 1	121417-JETT-E-U-SMS	15219	4.3336	-245.1		PPB	20 Dec 2017 15:53:09	U	SMPL
I6956-10MSD - 1	121417-JETT-E-U-SMSD	14941	4.2542	-245.1		PPB	20 Dec 2017 15:55:17	U	SMPL
PB105188BL - 1	PBW	-30	-0.0221	-245.1		PPB	20 Dec 2017 15:57:25	U	SMPL
PB105188BS - 1	LCSW	13307	3.7874	-245.1		PPB	20 Dec 2017 15:59:32	U	SMPL
CCV10 - 1	CCV10	16901	4.8140	-245.1		PPB	20 Dec 2017 16:01:40	U	SMPL
CCB10 - 1	CCB10	-68	-0.0330	-245.1		PPB	20 Dec 2017 16:03:46	U	SMPL
I6971-01 - 1	121517-TIDO-F-D-B	61	0.0039	-245.1		PPB	20 Dec 2017 16:05:54	U	SMPL
I6971-02 - 1	121517-TIDO-F-D-S	79	0.0090	-245.1		PPB	20 Dec 2017 16:08:01	U	SMPL
I6971-03 - 1	121517-TIDO-F-U-B	73	0.0073	-245.1		PPB	20 Dec 2017 16:10:08	U	SMPL
I6971-04 - 1	121517-TIDO-F-U-S	73	0.0073	-245.1		PPB	20 Dec 2017 16:12:15	U	SMPL
I6971-05 - 1	121517-JETT-E-D-B	115	0.0193	-245.1		PPB	20 Dec 2017 16:14:23	U	SMPL
I6971-06 - 1	121517-JETT-E-D-M	64	0.0047	-245.1		PPB	20 Dec 2017 16:16:30	U	SMPL
I6971-07 - 1	121517-JETT-E-D-S	96	0.0139	-245.1		PPB	20 Dec 2017 16:18:38	U	SMPL
I6971-08 - 1	121517-JETT-E-U-B	95	0.0136	-245.1		PPB	20 Dec 2017 16:20:46	U	SMPL
I6971-09 - 1	121517-JETT-E-U-M	77	0.0084	-245.1		PPB	20 Dec 2017 16:22:55	U	SMPL
I6971-10 - 1	121517-JETT-E-U-S	80	0.0093	-245.1		PPB	20 Dec 2017 16:25:02	U	SMPL
CCV11 - 1	CCV11	17011	4.8454	-245.1		PPB	20 Dec 2017 16:27:09	U	SMPL
CCB11 - 1	CCB11	-53	-0.0287	-245.1		PPB	20 Dec 2017 16:29:16	U	SMPL
I6971-11 - 1	121517-DBRI-E-D-B	96	0.0139	-245.1		PPB	20 Dec 2017 16:31:24	U	SMPL
I6971-12 - 1	121517-DBRI-E-D-S	74	0.0076	-245.1		PPB	20 Dec 2017 16:33:32	U	SMPL
I6971-13 - 1	121517-DBRI-E-U-B	212	0.0470	-245.1		PPB	20 Dec 2017 16:35:39	U	SMPL
I6971-14 - 1	121517-DBRI-E-U-S	82	0.0099	-245.1		PPB	20 Dec 2017 16:37:46	U	SMPL
I6971-15 - 1	121517-TIFR-E-D-B	98	0.0144	-245.1		PPB	20 Dec 2017 16:39:54	U	SMPL
I6971-16 - 1	121517-TIFR-E-D-M	85	0.0107	-245.1		PPB	20 Dec 2017 16:42:02	U	SMPL
I6971-17 - 1	121517-TIFR-E-D-S	80	0.0093	-245.1		PPB	20 Dec 2017 16:44:10	U	SMPL
I6971-18 - 1	121517-TIFR-E-U-B	84	0.0104	-245.1		PPB	20 Dec 2017 16:46:18	U	SMPL
I6971-19 - 1	121517-TIFR-E-U-M	61	0.0039	-245.1		PPB	20 Dec 2017 16:48:27	U	SMPL
I6971-20 - 1	121517-TIFR-E-U-S	217	0.0484	-245.1		PPB	20 Dec 2017 16:50:34	U	SMPL
CCV12 - 1	CCV12	16935	4.8237	-245.1		PPB	20 Dec 2017 16:52:41	U	SMPL
CCB12 - 1	CCB12	-74	-0.0347	-245.1		PPB	20 Dec 2017 16:54:48	U	SMPL
I6971-20DUP - 1	121517-TIFR-E-U-SDUP	106	0.0167	-245.1		PPB	20 Dec 2017 16:56:58	U	SMPL
I6971-20MS - 1	121517-TIFR-E-U-SMS	14803	4.2147	-245.1		PPB	20 Dec 2017 17:01:35	U	SMPL
I6971-20MSD - 1	121517-TIFR-E-U-SMSD	14137	4.0245	-245.1		PPB	20 Dec 2017 17:03:42	U	SMPL
I6956-10LX5 - 1		-44	-0.0261	-245.1		PPB	20 Dec 2017 17:05:49	U	SMPL
I6956-10A - 1		15042	4.2830	-245.1		PPB	20 Dec 2017 17:07:57	U	SMPL
I6971-20LX5 - 1		-15	-0.0179	-245.1		PPB	20 Dec 2017 17:10:05	U	SMPL
I6971-20A - 1		14535	4.1382	-245.1		PPB	20 Dec 2017 17:12:13	U	SMPL
CCV13 - 1	CCV13	16887	4.8100	-245.1		PPB	20 Dec 2017 17:14:21	U	SMPL
CCB13 - 1	CCB13	-84	-0.0376	-245.1		PPB	20 Dec 2017 17:16:30	U	SMPL