


ISM02.4

LB 92285
INSTRUMENT ID: CVL

Linear

 μ Abs.:

44669



A= 0.0000e+000

B= 2.2589e-004

C= -1.1598e-002

Rho= 0.9998357

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.001	-0.001	45	0.000	45					-
0.2	0.200	0.196	-0.004	919	0.0 %	919					-2
2.5	2.500	2.587	0.087	11502	0.0 %	11502					3
5.0	5.000	4.932	-0.068	21886	0.0 %	21886					-1
7.5	7.500	7.408	-0.092	32844	0.0 %	32844					-1
10.0	10.000	10.079	0.079	44669	0.0 %	44669					1

NB
R/2017

LB92285
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	45	-		0 ISM02.4	PPB	12/20/2017 17:32	Std
	0.2 S0.2	919	-		0.2 ISM02.4	PPB	12/20/2017 17:34	Std
	2.5 S2.5	11502	-		2.5 ISM02.4	PPB	12/20/2017 17:36	Std
	5 S5	21886	-		5 ISM02.4	PPB	12/20/2017 17:38	Std
	7.5 S7.5	32844	-		7.5 ISM02.4	PPB	12/20/2017 17:41	Std
	10 S10	44669	-		10 ISM02.4	PPB	12/20/2017 17:46	Std
ICV	ICV	16831	3.7904	-	ISM02.4	PPB	12/20/2017 17:49	SMPL
ICB	ICB	-67	-0.0267	-	ISM02.4	PPB	12/20/2017 17:51	SMPL
CCV014	CCV014	21827	4.919	-	ISM02.4	PPB	12/20/2017 17:53	SMPL
CCB014	CCB014	-89	-0.0317	-	ISM02.4	PPB	12/20/2017 17:56	SMPL
PB105216BL	PBS004	121	0.0157	-	ISM02.4	PPB	12/20/2017 17:58	SMPL
I6774-01	MA4Z99	1978	0.4352	-	ISM02.4	PPB	12/20/2017 18:00	SMPL
I6774-02	MA4ZA0	2967	0.6586	-	ISM02.4	PPB	12/20/2017 18:02	SMPL
I6774-03	MA4ZA1	16349	3.6815	-	ISM02.4	PPB	12/20/2017 18:05	SMPL
I6774-04	MA4ZA2	20160	4.5424	-	ISM02.4	PPB	12/20/2017 18:07	SMPL
I6774-05	MA4ZA3	3344	0.7438	-	ISM02.4	PPB	12/20/2017 18:09	SMPL
I6774-06	MA4ZA4	6000	1.3438	-	ISM02.4	PPB	12/20/2017 18:12	SMPL
I6774-07	MA4ZA5	3542	0.7885	-	ISM02.4	PPB	12/20/2017 18:14	SMPL
I6774-08	MA4ZA6	3283	0.73	-	ISM02.4	PPB	12/20/2017 18:16	SMPL
I6774-09	MA4ZA7	6920	1.5516	-	ISM02.4	PPB	12/20/2017 18:18	SMPL
I6774-10	MA4ZA8	4693	1.0485	-	ISM02.4	PPB	12/20/2017 18:21	SMPL
I6774-11	MA4ZA9	2594	0.5744	-	ISM02.4	PPB	12/20/2017 18:24	SMPL
I6774-12	MA4ZB0	6008	1.3456	-	ISM02.4	PPB	12/20/2017 18:27	SMPL
I6774-13	MA4ZB1	2695	0.5972	-	ISM02.4	PPB	12/20/2017 18:29	SMPL
I6774-17	MA4ZC6	2923	0.6487	-	ISM02.4	PPB	12/20/2017 18:33	SMPL
I6774-18	MA4ZC7	2490	0.5509	-	ISM02.4	PPB	12/20/2017 18:36	SMPL
I6774-19	MA4ZC8	6827	1.5306	-	ISM02.4	PPB	12/20/2017 18:38	SMPL
I6774-20	MA4ZC8D	6936	1.5552	-	ISM02.4	PPB	12/20/2017 18:40	SMPL
I6774-21	MA4ZC8S	16543	3.7254	-	ISM02.4	PPB	12/20/2017 18:42	SMPL
CCV015	CCV015	21653	4.8797	-	ISM02.4	PPB	12/20/2017 18:45	SMPL
CCB015	CCB015	-172	-0.0505	-	ISM02.4	PPB	12/20/2017 18:47	SMPL
I6774-22	MA4ZC9	24778	5.5856	-	ISM02.4	PPB	12/20/2017 18:49	SMPL
I6774-23	MA4ZD0	13337	3.0011	-	ISM02.4	PPB	12/20/2017 18:52	SMPL
I6774-24	MA4ZD1	15233	3.4294	-	ISM02.4	PPB	12/20/2017 18:54	SMPL
PB105217BL	PBS004	103	0.0117	-	ISM02.4	PPB	12/20/2017 18:56	SMPL
I6800-03	MC0298	383	0.0749	-	ISM02.4	PPB	12/20/2017 18:58	SMPL
I6800-04	MC02A1	2783	0.6171	-	ISM02.4	PPB	12/20/2017 19:01	SMPL
I6800-09	MC0299	978	0.2093	-	ISM02.4	PPB	12/20/2017 19:03	SMPL
I6800-10	MC02B3	9128	2.0504	-	ISM02.4	PPB	12/20/2017 19:08	SMPL
I6800-11	MC02B3D	9058	2.0345	-	ISM02.4	PPB	12/20/2017 19:10	SMPL
I6800-12	MC02B3S	19298	4.3477	-	ISM02.4	PPB	12/20/2017 19:12	SMPL
I6800-14	MC02B7	1320	0.2866	-	ISM02.4	PPB	12/20/2017 19:15	SMPL
I6800-16	MC02E1	1085	0.2335	-	ISM02.4	PPB	12/20/2017 19:17	SMPL
PB105218BL	PBS004	212	0.0363	-	ISM02.4	PPB	12/20/2017 19:19	SMPL
I6807-01	MA4ZD2	40260	9.0829	-	ISM02.4	PPB	12/20/2017 19:22	SMPL
I6807-02	MA4ZD3	17637	3.9725	-	ISM02.4	PPB	12/20/2017 19:24	SMPL

INSTRUMENT ID : CV1

I6807-03	MA4ZD4	11898	2.6761 -	ISM02.4	PPB	12/20/2017 19:26	SMPL
I6807-04	MA4ZD5	11835	2.6619 -	ISM02.4	PPB	12/20/2017 19:28	SMPL
I6807-05	MA4ZD6	4837	1.081 -	ISM02.4	PPB	12/20/2017 19:31	SMPL
I6807-06	MA4ZD7	374	0.0729 -	ISM02.4	PPB	12/20/2017 19:33	SMPL
I6807-07	MA4ZD8	9756	2.1922 -	ISM02.4	PPB	12/20/2017 19:35	SMPL
CCV016	CCV016	21463	4.8368 -	ISM02.4	PPB	12/20/2017 19:38	SMPL
CCB016	CCB016	-78	-0.0292 -	ISM02.4	PPB	12/20/2017 19:40	SMPL
I6807-09	MA4ZE3	14390	3.239 -	ISM02.4	PPB	12/20/2017 19:42	SMPL
I6807-10	MA4ZE4	4779	1.0679 -	ISM02.4	PPB	12/20/2017 19:44	SMPL
I6807-11	MA4ZE5	9612	2.1597 -	ISM02.4	PPB	12/20/2017 19:47	SMPL
I6807-12	MA4ZE6	770	0.1623 -	ISM02.4	PPB	12/20/2017 19:49	SMPL
I6807-13	MA4ZE7	552	0.1131 -	ISM02.4	PPB	12/20/2017 19:51	SMPL
I6807-14	MA4ZE8	13296	2.9919 -	ISM02.4	PPB	12/20/2017 19:54	SMPL
I6807-15	MA4ZE8D	13179	2.9655 -	ISM02.4	PPB	12/20/2017 19:56	SMPL
I6807-16	MA4ZE8S	24581	5.5411 -	ISM02.4	PPB	12/20/2017 19:58	SMPL
PB105219BL	PBS004	21	-0.0069 -	ISM02.4	PPB	12/20/2017 20:00	SMPL
I6735-01	MA4Z49	1020	0.2188 -	ISM02.4	PPB	12/20/2017 20:03	SMPL
I6735-02	MA4Z50	46518	10.4965 -	ISM02.4	PPB	12/20/2017 20:05	SMPL
I6735-03	MA4Z51	5796	1.2977 -	ISM02.4	PPB	12/20/2017 20:07	SMPL
I6735-04	MA4Z52	8559	1.9218 -	ISM02.4	PPB	12/20/2017 20:10	SMPL
I6735-05	MA4Z53	4051	0.9035 -	ISM02.4	PPB	12/20/2017 20:12	SMPL
I6735-06	MA4Z54	5716	1.2796 -	ISM02.4	PPB	12/20/2017 20:14	SMPL
I6735-07	MA4Z54S	16669	3.7538 -	ISM02.4	PPB	12/20/2017 20:16	SMPL
I6735-08	MA4Z54D	5573	1.2473 -	ISM02.4	PPB	12/20/2017 20:19	SMPL
CCV017	CCV017	21584	4.8641 -	ISM02.4	PPB	12/20/2017 20:21	SMPL
CCB017	CCB017	-103	-0.0349 -	ISM02.4	PPB	12/20/2017 20:23	SMPL
I6774-14	MA4ZB5	13401	3.0156 -	ISM02.4	PPB	12/20/2017 20:26	SMPL
I6735-02DLX2	MA4Z50	23447	5.2849 -	ISM02.4	PPB	12/20/2017 20:30	SMPL
CCV018	CCV018	21424	4.8279 -	ISM02.4	PPB	12/20/2017 20:32	SMPL
CCB018	CCB018	-104	-0.0351 -	ISM02.4	PPB	12/20/2017 20:35	SMPL