



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92291

Temp.: 19.4 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/20/2017	17:10
2	CAL2	1	Water			6.99	12/20/2017	17:11
3	CAL3	1	Water			10.01	12/20/2017	17:12
4	ICV	1	Water			7.00	12/20/2017	17:13
5	CCV1	1	Water			2.01	12/20/2017	17:15
6	I7007-01	1	Water			7.66	12/20/2017	17:16
7	I7007-01DUP	1	Water			7.67	12/20/2017	17:19
8	CCV2	1	Water			12.01	12/20/2017	17:20

1392291

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I7007

WorkList ID : 106799

Date : 12/20/2017 4:53:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/20/2017	Water	I7007-01-13	pH	Cool 4 deg C	PSEG01	F52	SV1854L	12/20/2017	SM 4500-PH B

Date/Time 12-20-17 4:51:00 PM
Received by: JF
Relinquished by:

Date/Time 12-20-17 6:00 PM
Received by: CP
Relinquished by: JA