



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92293

Temp.: 19.4 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/20/2017	17:10
2	CAL2	1	Water			6.99	12/20/2017	17:11
3	CAL3	1	Water			10.01	12/20/2017	17:12
4	ICV	1	Water			7.00	12/20/2017	17:13
5	CCV1	1	Water			2.01	12/20/2017	17:24
6	I6683-08	1	Solid	20.02	20	8.48	12/20/2017	17:25
7	I6683-08DUP	1	Solid	20.03	20	8.49	12/20/2017	17:26
8	I6683-10	1	Solid	20.04	20	7.33	12/20/2017	17:27
9	I6683-12	1	Solid	20.01	20	6.85	12/20/2017	17:28
10	CCV2	1	Water			12.01	12/20/2017	17:30

LB 92293

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6907			WorkList ID : 106798		Date : 12/20/2017 4:52:13 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/20/2017	Solid	I6683-08	Corrosivity	Cool 4 deg C	PSEG05		EO-02-122017-A	12/20/2017	9045C
12/20/2017	Solid	I6683-10	Corrosivity	Cool 4 deg C	PSEG05		EO-02-122017-B	12/20/2017	9045C
12/20/2017	Solid	I6683-12	Corrosivity	Cool 4 deg C	PSEG05		EO-02-122017-C	12/20/2017	9045C

Date/Time 12.20.17 5:00 PM
Received by: JL
Relinquished by: CA

Date/Time 12.20.17 6:30 PM
Received by: CD
Relinquished by: JP