



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92311

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	12/21/2017	09:15
2	I7007-02		1	>100.0	>212.0	12/21/2017	09:40
3	I7007-02DUP		1	>100.0	>212.0	12/21/2017	10:05

Result = (Celsius * 1.8) + 32

LB92311

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP122117

WorkList ID : 106820

Date : 12/21/2017 8:16:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/24/2017	Water	I7007-02 C	Flash Point	Cool 4 deg C	PSEG01	F52	SV1854L	12/20/2017	1010A

Date/Time 12/21/17 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 12/21/17 10:15 AM
Received by: CP
Relinquished by: KS