


7470A

LB 92316
INSTRUMENT ID: CYL

Linear

 μ Abs.:

35644



A= 0.0000e+000

B= 2.7998e-004

C= 1.5222e-002

Rho= 0.9999557

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	21	0.000	21				
0.2	0.200	0.199	-0.001	655	0.0 %	655				
2.5	2.500	2.502	0.002	8883	0.0 %	8883				
5.0	5.000	4.934	-0.066	17568	0.0 %	17568				
7.5	7.500	7.549	0.049	26908	0.0 %	26908				
10.0	10.000	9.995	-0.005	35644	0.0 %	35644				

LB92316 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Dec 2017 15:11:41

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	21	-	0.00007470A	PPB	21 Dec 2017 15:14:07	S	Std
0.2 - 1	50.2	655	-	0.20007470A	PPB	21 Dec 2017 15:16:16	S	Std
2.5 - 1	52.5	8883	-	2.50007470A	PPB	21 Dec 2017 15:18:25	S	Std
5.0 - 1	55	17568	-	5.00007470A	PPB	21 Dec 2017 15:20:33	S	Std
7.5 - 1	57.5	26908	-	7.50007470A	PPB	21 Dec 2017 15:22:41	S	Std
10.0 - 1	510	35644	-	10.00007470A	PPB	21 Dec 2017 15:24:49	S	Std
ICV19 - 1	ICV19	13647	3.8362	-7470A	PPB	21 Dec 2017 15:27:32	U	SMPL
ICB19 - 1	ICB19	-99	-0.0125	-7470A	PPB	21 Dec 2017 15:29:39	U	SMPL
CCV27 - 1	CCV27	18016	5.0594	-7470A	PPB	21 Dec 2017 15:31:45	U	SMPL
CCB27 - 1	CCB27	-101	-0.0131	-7470A	PPB	21 Dec 2017 15:33:52	U	SMPL
CRA - 1	CRA	651	0.1975	-7470A	PPB	21 Dec 2017 15:36:00	U	SMPL
HighStd - 1	HighStd	35593	9.9807	-7470A	PPB	21 Dec 2017 15:38:06	U	SMPL
ChkStd - 1	ChkStd	24424	6.8535	-7470A	PPB	21 Dec 2017 15:40:13	U	SMPL
PB105233BL - 1	PBW	-126	-0.0201	-7470A	PPB	21 Dec 2017 15:42:22	U	SMPL
PB105233TB - 1	PB105181TB	11	0.0183	-7470A	PPB	21 Dec 2017 15:46:38	U	SMPL
I6681-02 - 1	PL-01-121917-A	9	0.0177	-7470A	PPB	21 Dec 2017 15:48:46	U	SMPL
I6681-02DUP - 1	PL-01-121917-ADUP	100	0.0432	-7470A	PPB	21 Dec 2017 15:50:55	U	SMPL
CCV28 - 1	CCV28	17917	5.0317	-7470A	PPB	21 Dec 2017 15:57:16	U	SMPL
CCB28 - 1	CCB28	-84	-0.0083	-7470A	PPB	21 Dec 2017 15:59:23	U	SMPL
I6681-04 - 1	PL-01-121917-B	0	0.0152	-7470A	PPB	21 Dec 2017 16:01:30	U	SMPL
I6681-06 - 1	PL-01-121917-C	23	0.0217	-7470A	PPB	21 Dec 2017 16:03:37	U	SMPL
I6681-08 - 1	PL-01-121917-D	10	0.0180	-7470A	PPB	21 Dec 2017 16:05:44	U	SMPL
I6979-01 - 1	SAU-1-E(3-4)	886	0.2633	-7470A	PPB	21 Dec 2017 16:07:51	U	SMPL
I6979-02 - 1	SAU-4-E(7-8)	12	0.0186	-7470A	PPB	21 Dec 2017 16:09:58	U	SMPL
I6988-03 - 1	RT4008-RT3867-RT3075	23	0.0217	-7470A	PPB	21 Dec 2017 16:12:06	U	SMPL
I6990-02 - 1	MH-CONCRETE	21	0.0211	-7470A	PPB	21 Dec 2017 16:14:14	U	SMPL
I6993-03 - 1	ROLL-OFF-3	41	0.0267	-7470A	PPB	21 Dec 2017 16:16:22	U	SMPL
I6993-04 - 1	ROLL-OFF-4	34	0.0247	-7470A	PPB	21 Dec 2017 16:18:31	U	SMPL
I6995-03 - 1	TP-1	24	0.0219	-7470A	PPB	21 Dec 2017 16:20:38	U	SMPL
CCV29 - 1	CCV29	17758	4.9872	-7470A	PPB	21 Dec 2017 16:22:45	U	SMPL
CCB29 - 1	CCB29	-93	-0.0108	-7470A	PPB	21 Dec 2017 16:24:52	U	SMPL
I6995-04 - 1	TP-2	15	0.0194	-7470A	PPB	21 Dec 2017 16:26:59	U	SMPL
PB105233BS - 1	LCSW	13754	3.8661	-7470A	PPB	21 Dec 2017 16:29:06	U	SMPL
I6681-02MS - 1	PL-01-121917-AMS	13770	3.8706	-7470A	PPB	21 Dec 2017 16:31:13	U	SMPL
I6681-02MSD - 1	PL-01-121917-AMSD	13633	3.8322	-7470A	PPB	21 Dec 2017 16:33:20	U	SMPL
I6681-02LX5 - 1		-76	-0.0061	-7470A	PPB	21 Dec 2017 16:35:28	U	SMPL
I6681-02A - 1		13945	3.9196	-7470A	PPB	21 Dec 2017 16:37:35	U	SMPL
CCV30 - 1	CCV30	17483	4.9102	-7470A	PPB	21 Dec 2017 16:39:43	U	SMPL
CCB30 - 1	CCB30	-68	-0.0038	-7470A	PPB	21 Dec 2017 16:41:52	U	SMPL