



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92330

Temp.: 23.04 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/21/2017	18:45
2	CAL2	1	Water			6.99	12/21/2017	18:46
3	CAL3	1	Water			10.02	12/21/2017	18:47
4	ICV	1	Water			7.00	12/21/2017	18:48
5	CCV1	1	Water			2.01	12/21/2017	18:50
6	I7047-01	1	Water			5.76	12/21/2017	18:52
7	I7047-01DUP	1	Water			5.77	12/21/2017	18:54
8	I7048-01	1	Water			5.21	12/21/2017	18:56
9	CCV2	1	Water			12.01	12/21/2017	18:58

WORKLIST(Hardcopy Internal Chain)

1392230

WorkList Name : PH I7047

WorkList ID : 106859

Date : 12/21/2017 6:35:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2017	Water	I7047-01-A	pH	Cool 4 deg C	STER15	J11	GILROY-OMEGA-CELL	12/16/2017	SM 4500-PH B
12/18/2017	Water	I7048-01 - J	pH	Cool 4 deg C	STER04	G12	POOL-WATER	12/18/2017	SM 4500-PH B

Date/Time

12-21-17 6:34 PM

Received by:

Jp

Relinquished by:

CS

Date/Time

12-21-17

7:30 PM

Received by:

CS

Relinquished by:

Jp