


7470A

LB 92337
INSTRUMENT 12: CV2

Linear

 μ Abs.:

34714



A= 0.0000e+000

B= 2.8722e-004

C= 1.9507e-002

Rho= 0.9999606

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	19	0.000	19				
0.2	0.200	0.216	0.016	683	0.0 %	683				
2.5	2.500	2.433	-0.067	8402	0.0 %	8402				
5.0	5.000	5.007	0.007	17363	0.0 %	17363				
7.5	7.500	7.530	0.030	26149	0.0 %	26149				
10.0	10.000	9.990	-0.010	34714	0.0 %	34714				

LB92337 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Dec 2017 10:21:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7470A	PPB	22 Dec 2017 10:21:54	S	Std
0.2 - 1	50.2	683	-	0.2000	7470A	PPB	22 Dec 2017 10:26:28	S	Std
2.5 - 1	52.5	8402	-	2.5000	7470A	PPB	22 Dec 2017 10:28:37	S	Std
5.0 - 1	55	17363	-	5.0000	7470A	PPB	22 Dec 2017 10:30:45	S	Std
7.5 - 1	57.5	26149	-	7.5000	7470A	PPB	22 Dec 2017 10:32:53	S	Std
10.0 - 1	50	34714	-	10.0000	7470A	PPB	22 Dec 2017 10:35:01	S	Std
ICV20 - 1	ICV20	13091	3.7795	-	7470A	PPB	22 Dec 2017 10:37:37	U	SMPL
ICB20 - 1	ICB20	-82	-0.0040	-	7470A	PPB	22 Dec 2017 10:39:43	U	SMPL
CCV31 - 1	CCV31	17548	5.0597	-	7470A	PPB	22 Dec 2017 10:41:50	U	SMPL
CCB31 - 1	CCB31	-124	-0.0161	-	7470A	PPB	22 Dec 2017 10:43:56	U	SMPL
CRA - 1	CRA	685	0.2163	-	7470A	PPB	22 Dec 2017 10:46:04	U	SMPL
HighStd - 1	HighStd	34898	10.0429	-	7470A	PPB	22 Dec 2017 10:48:11	U	SMPL
ChkStd - 1	ChkStd	24607	7.0871	-	7470A	PPB	22 Dec 2017 10:50:17	U	SMPL
PB105267BL - 1	PBW	-152	-0.0242	-	7470A	PPB	22 Dec 2017 10:52:26	U	SMPL
PB105267BS - 1	LCSW	12303	3.5532	-	7470A	PPB	22 Dec 2017 10:54:35	U	SMPL
PB105267TB - 1	PB105230TB	-35	0.0095	-	7470A	PPB	22 Dec 2017 10:56:42	U	SMPL
I6683-08 - 1	EO-02-122017-A	18	0.0247	-	7470A	PPB	22 Dec 2017 10:58:51	U	SMPL
I6683-08DUP - 1	EO-02-122017-ADUP	9	0.0221	-	7470A	PPB	22 Dec 2017 11:00:59	U	SMPL
I6683-08MS - 1	EO-02-122017-AMS	12768	3.6867	-	7470A	PPB	22 Dec 2017 11:03:08	U	SMPL
I6683-08MSD - 1	EO-02-122017-AMSD	12096	3.4937	-	7470A	PPB	22 Dec 2017 11:05:15	U	SMPL
CCV32 - 1	CCV32	17671	5.0950	-	7470A	PPB	22 Dec 2017 11:07:21	U	SMPL
CCB32 - 1	CCB32	-107	-0.0112	-	7470A	PPB	22 Dec 2017 11:09:28	U	SMPL
I6683-10 - 1	EO-02-122017-B	-1	0.0192	-	7470A	PPB	22 Dec 2017 11:11:36	U	SMPL
I6683-12 - 1	EO-02-122017-C	31	0.0284	-	7470A	PPB	22 Dec 2017 11:13:42	U	SMPL
I6979-22 - 1	SAU-2-E(4-5)	52	0.0344	-	7470A	PPB	22 Dec 2017 11:15:49	U	SMPL
I7007-02 - 1	SV1854L	14	0.0235	-	7470A	PPB	22 Dec 2017 11:17:56	U	SMPL
PB105293BL - 1	PBW	2	0.0201	-	7470A	PPB	22 Dec 2017 11:20:03	U	SMPL
PB105293BS - 1	LCSW	12317	3.5572	-	7470A	PPB	22 Dec 2017 11:22:11	U	SMPL
I7007-01 - 1	SV21854L	259	0.0939	-	7470A	PPB	22 Dec 2017 11:28:53	U	SMPL
I7007-01DUP - 1	SV21854LDUP	227	0.0847	-	7470A	PPB	22 Dec 2017 11:33:26	U	SMPL
I7007-01MS - 1	SV21854LMS	11350	3.2795	-	7470A	PPB	22 Dec 2017 11:35:34	U	SMPL
I7007-01MSD - 1	SV21854LMSD	11727	3.3877	-	7470A	PPB	22 Dec 2017 11:37:41	U	SMPL
CCV33 - 1	CCV33	16435	4.7400	-	7470A	PPB	22 Dec 2017 11:39:49	U	SMPL
CCB33 - 1	CCB33	-94	-0.0075	-	7470A	PPB	22 Dec 2017 11:41:56	U	SMPL
I7019-02 - 1	FB-03	-10	0.0166	-	7470A	PPB	22 Dec 2017 11:44:04	U	SMPL
I6683-08LX5 - 1		17	0.0244	-	7470A	PPB	22 Dec 2017 11:46:11	U	SMPL
I6683-08A - 1		12805	3.6974	-	7470A	PPB	22 Dec 2017 11:48:18	U	SMPL
I7007-01LX5 - 1		-8	0.0172	-	7470A	PPB	22 Dec 2017 11:50:25	U	SMPL
I7007-01A - 1		12734	3.6770	-	7470A	PPB	22 Dec 2017 11:52:32	U	SMPL
CCV34 - 1	CCV34	17060	4.9195	-	7470A	PPB	22 Dec 2017 11:54:40	U	SMPL
CCB34 - 1	CCB34	-126	-0.0167	-	7470A	PPB	22 Dec 2017 11:56:48	U	SMPL