

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB92338

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP60515
calibration std. phosphate 0.5 ppm	WP60514
calibration std. phosphate 0.3 ppm	WP60513
calibration std. 0 ppm	WP60510
phosphate CCV std.	WP60516
calibration std. phosphate 0.1 ppm	WP60512
calibration std. phosphate 0.05 ppm	WP60511
Combined reagent	WP60520
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036
phosphate ICV std.	WP60517

Intercept: -0.0007Slope: 0.6498Regression: 0.999842

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/22/2017	11:36
2	CAL2	0.05	1	50	50	0.033	12/22/2017	11:36
3	CAL3	0.10	1	50	50	0.066	12/22/2017	11:37
4	CAL4	0.30	1	50	50	0.186	12/22/2017	11:37
5	CAL5	0.50	1	50	50	0.329	12/22/2017	11:38
6	CAL6	1.00	1	50	50	0.649	12/22/2017	11:38



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB92338

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	12/22/2017	11:39
2	ICB		1	50	50	0.000	0.001	12/22/2017	11:39
3	CCV1	0.50	1	50	50	0.326	0.503	12/22/2017	11:40
4	CCB1		1	50	50	0.000	0.001	12/22/2017	11:40
5	PB105295BL		1	50	50	0.000	0.001	12/22/2017	11:41
6	PB105295BS	0.50	1	50	50	0.314	0.484	12/22/2017	11:41
7	I7003-01		1	50	50	0.016	0.026	12/22/2017	11:42
8	I7003-01DUP		1	50	50	0.014	0.023	12/22/2017	11:42
9	I7003-01MS	0.50	1	50	50	0.315	0.486	12/22/2017	11:43
10	I7003-01MSD	0.50	1	50	50	0.316	0.487	12/22/2017	11:43
11	I7003-02		1	50	50	0.029	0.046	12/22/2017	11:44
12	I7027-01		1	50	50	2.086	3.211	12/22/2017	11:45
13	CCV2	0.50	1	50	50	0.326	0.503	12/22/2017	11:46
14	CCB2		1	50	50	0.000	0.001	12/22/2017	11:46
15	I7027-01		10	50	50	0.516	0.795	12/22/2017	11:50
16	CCV3	0.50	1	50	50	0.327	0.504	12/22/2017	11:50
17	CCB3		1	50	50	0.000	0.001	12/22/2017	11:51