

245.1


LB92345

INSTRUMENT ID: CV2

Linear

 μ Abs.:

3444E



A= 0.0000e+000

B= 2.9024e-004

C= 2.3866e-004

Rho= 0.9999986

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	19	0.000	19				
0.05	0.050	0.053	0.003	182	0.0 %	182				
0.2	0.200	0.198	-0.002	681	0.0 %	681				
2.5	2.500	2.493	-0.007	8587	0.0 %	8587				
5.0	5.000	4.992	-0.008	17200	0.0 %	17200				
7.5	7.500	7.511	0.011	25876	0.0 %	25876				
10.0	10.000	9.998	-0.002	34445	0.0 %	34445				

LB92345 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Dec 2017 13:06:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.000245.1		PPB	22 Dec 2017 13:11:03	S	Std
0.05 - 1	50.05	182	-	0.0500245.1		PPB	22 Dec 2017 13:13:11	S	Std
0.2 - 1	50.2	681	-	0.2000245.1		PPB	22 Dec 2017 13:15:20	S	Std
2.5 - 1	52.5	8587	-	2.5000245.1		PPB	22 Dec 2017 13:17:28	S	Std
5.0 - 1	55	17200	-	5.0000245.1		PPB	22 Dec 2017 13:19:36	S	Std
7.5 - 1	57.5	25876	-	7.5000245.1		PPB	22 Dec 2017 13:21:43	S	Std
10.0 - 1	511	34445	-	10.0000245.1		PPB	22 Dec 2017 13:26:14	S	Std
ICV21 - 1	ICV21	13242	3.8437	-245.1		PPB	22 Dec 2017 13:43:13	U	SMPL
ICB21 - 1	ICB21	-73	-0.0209	-245.1		PPB	22 Dec 2017 13:45:20	U	SMPL
CCV35 - 1	CCV35	16894	4.9036	-245.1		PPB	22 Dec 2017 13:47:26	U	SMPL
CCB35 - 1	CCB35	-101	-0.0291	-245.1		PPB	22 Dec 2017 13:49:33	U	SMPL
CRA - 1	CRA	667	0.1938	-245.1		PPB	22 Dec 2017 13:51:40	U	SMPL
HighStd - 1	HighStd	34524	10.0206	-245.1		PPB	22 Dec 2017 13:53:47	U	SMPL
ChkStd - 1	ChkStd	23458	6.8088	-245.1		PPB	22 Dec 2017 13:55:54	U	SMPL
PB105257BL - 1	PBW	-136	-0.0392	-245.1		PPB	22 Dec 2017 13:58:03	U	SMPL
PB105257BS - 1	LCSW	12542	3.6405	-245.1		PPB	22 Dec 2017 14:00:12	U	SMPL
I7008-01 - 1	121817-DBRI-F-D-B	-37	-0.0105	-245.1		PPB	22 Dec 2017 14:02:20	U	SMPL
I7008-02 - 1	121817-DBRI-F-D-S	2	0.0008	-245.1		PPB	22 Dec 2017 14:04:28	U	SMPL
I7008-03 - 1	121817-DBRI-F-U-B	26	0.0078	-245.1		PPB	22 Dec 2017 14:06:36	U	SMPL
I7008-04 - 1	121817-DBRI-F-U-S	18	0.0055	-245.1		PPB	22 Dec 2017 14:08:45	U	SMPL
I7008-05 - 1	121817-TIDO-F-D-B	79	0.0232	-245.1		PPB	22 Dec 2017 14:10:52	U	SMPL
CCV36 - 1	CCV36	16645	4.8314	-245.1		PPB	22 Dec 2017 14:12:58	U	SMPL
CCB36 - 1	CCB36	-87	-0.0250	-245.1		PPB	22 Dec 2017 14:15:05	U	SMPL
I7008-06 - 1	121817-TIDO-F-D-S	35	0.0104	-245.1		PPB	22 Dec 2017 14:17:12	U	SMPL
I7008-07 - 1	121817-TIDO-F-U-B	37	0.0110	-245.1		PPB	22 Dec 2017 14:19:18	U	SMPL
I7008-08 - 1	121817-TIDO-F-U-S	7	0.0023	-245.1		PPB	22 Dec 2017 14:21:25	U	SMPL
I7008-09 - 1	121817-TIFR-E-D-B	24	0.0072	-245.1		PPB	22 Dec 2017 14:23:32	U	SMPL
I7008-10 - 1	121817-TIFR-E-D-M	20	0.0060	-245.1		PPB	22 Dec 2017 14:25:39	U	SMPL
I7008-11 - 1	121817-TIFR-E-D-S	27	0.0081	-245.1		PPB	22 Dec 2017 14:27:47	U	SMPL
I7008-12 - 1	121817-TIFR-E-U-B	6	0.0020	-245.1		PPB	22 Dec 2017 14:29:55	U	SMPL
I7008-13 - 1	121817-TIFR-E-U-M	15	0.0046	-245.1		PPB	22 Dec 2017 14:32:03	U	SMPL
I7008-14 - 1	121817-TIFR-E-U-S	28	0.0084	-245.1		PPB	22 Dec 2017 14:34:12	U	SMPL
I7008-14DUP - 1	121817-TIFR-E-U-SDUP	20	0.0060	-245.1		PPB	22 Dec 2017 14:36:19	U	SMPL
CCV37 - 1	CCV37	16633	4.8279	-245.1		PPB	22 Dec 2017 14:38:25	U	SMPL
CCB37 - 1	CCB37	-84	-0.0241	-245.1		PPB	22 Dec 2017 14:40:33	U	SMPL
I7008-14MS - 1	121817-TIFR-E-U-SMS	14542	4.2210	-245.1		PPB	22 Dec 2017 14:42:41	U	SMPL
I7008-14MSD - 1	121817-TIFR-E-U-SMSD	14267	4.1412	-245.1		PPB	22 Dec 2017 14:44:48	U	SMPL
PB105258BL - 1	PBW	-107	-0.0308	-245.1		PPB	22 Dec 2017 14:46:55	U	SMPL
PB105258BS - 1	LCSW	13011	3.7766	-245.1		PPB	22 Dec 2017 14:49:02	U	SMPL
I7009-01 - 1	121917-JETT-F-D-B	-55	-0.0157	-245.1		PPB	22 Dec 2017 14:51:09	U	SMPL
I7009-02 - 1	121917-JETT-F-D-M	6	0.0020	-245.1		PPB	22 Dec 2017 14:53:17	U	SMPL
I7009-03 - 1	121917-JETT-F-D-S	85	0.0249	-245.1		PPB	22 Dec 2017 14:55:25	U	SMPL
I7009-04 - 1	121917-JETT-F-U-B	29	0.0087	-245.1		PPB	22 Dec 2017 14:57:33	U	SMPL
I7009-05 - 1	121917-JETT-F-U-M	26	0.0078	-245.1		PPB	22 Dec 2017 14:59:42	U	SMPL
I7009-06 - 1	121917-JETT-F-U-S	22	0.0066	-245.1		PPB	22 Dec 2017 15:01:49	U	SMPL
CCV38 - 1	CCV38	16888	4.9019	-245.1		PPB	22 Dec 2017 15:03:56	U	SMPL
CCB38 - 1	CCB38	-57	-0.0163	-245.1		PPB	22 Dec 2017 15:06:03	U	SMPL
I7009-07 - 1	121917-TIDO-F-D-B	13	0.0040	-245.1		PPB	22 Dec 2017 15:08:12	U	SMPL
I7009-08 - 1	121917-TIDO-F-D-S	30	0.0089	-245.1		PPB	22 Dec 2017 15:10:19	U	SMPL
I7009-09 - 1	121917-TIDO-F-U-B	10	0.0031	-245.1		PPB	22 Dec 2017 15:12:26	U	SMPL
I7009-10 - 1	121917-TIDO-F-U-S	17	0.0052	-245.1		PPB	22 Dec 2017 15:14:33	U	SMPL
I7009-10DUP - 1	121917-TIDO-F-U-SDUP	43	0.0127	-245.1		PPB	22 Dec 2017 15:16:41	U	SMPL
I7009-10MS - 1	121917-TIDO-F-U-SMS	14410	4.1827	-245.1		PPB	22 Dec 2017 15:18:48	U	SMPL
I7009-10MSD - 1	121917-TIDO-F-U-SMSD	13980	4.0579	-245.1		PPB	22 Dec 2017 15:20:56	U	SMPL
PB105259BL - 1	PBW	-99	-0.0285	-245.1		PPB	22 Dec 2017 15:23:04	U	SMPL
PB105259BS - 1	LCSW	12719	3.6919	-245.1		PPB	22 Dec 2017 15:25:13	U	SMPL
I7010-01 - 1	122017-DBRI-F-D-M	11	0.0034	-245.1		PPB	22 Dec 2017 15:27:20	U	SMPL
CCV39 - 1	CCV39	17171	4.9840	-245.1		PPB	22 Dec 2017 15:29:27	U	SMPL
CCB39 - 1	CCB39	-83	-0.0239	-245.1		PPB	22 Dec 2017 15:31:34	U	SMPL
I7010-02 - 1	122017-DBRI-F-U-M	20	0.0060	-245.1		PPB	22 Dec 2017 15:33:42	U	SMPL
I7010-03 - 1	122017-TIDO-F-D-B	24	0.0072	-245.1		PPB	22 Dec 2017 15:35:49	U	SMPL
I7010-04 - 1	122017-TIDO-F-D-S	40	0.0118	-245.1		PPB	22 Dec 2017 15:37:57	U	SMPL
I7010-05 - 1	122017-TIDO-F-U-B	39	0.0116	-245.1		PPB	22 Dec 2017 15:40:04	U	SMPL
I7010-06 - 1	122017-TIDO-F-U-S	45	0.0133	-245.1		PPB	22 Dec 2017 15:42:11	U	SMPL
I7010-07 - 1	122017-JETT-F-D-B	42	0.0124	-245.1		PPB	22 Dec 2017 15:44:19	U	SMPL
I7010-08 - 1	122017-JETT-F-D-M	46	0.0136	-245.1		PPB	22 Dec 2017 15:46:27	U	SMPL
I7010-09 - 1	122017-JETT-F-D-S	36	0.0107	-245.1		PPB	22 Dec 2017 15:48:35	U	SMPL
I7010-10 - 1	122017-JETT-F-U-B	59	0.0174	-245.1		PPB	22 Dec 2017 15:50:44	U	SMPL
I7010-11 - 1	122017-JETT-F-U-M	54	0.0159	-245.1		PPB	22 Dec 2017 15:52:51	U	SMPL
CCV40 - 1	CCV40	16982	4.9292	-245.1		PPB	22 Dec 2017 15:54:58	U	SMPL
CCB40 - 1	CCB40	-83	-0.0239	-245.1		PPB	22 Dec 2017 15:57:05	U	SMPL
I7010-12 - 1	122017-JETT-F-U-S	-9	-0.0024	-245.1		PPB	22 Dec 2017 15:59:14	U	SMPL
I7010-12DUP - 1	122017-JETT-F-U-SDUP	10	0.0031	-245.1		PPB	22 Dec 2017 16:01:22	U	SMPL
I7010-12MS - 1	122017-JETT-F-U-SMS	14015	4.0680	-245.1		PPB	22 Dec 2017 16:03:29	U	SMPL
I7010-12MSD - 1	122017-JETT-F-U-SMSD	13937	4.0454	-245.1		PPB	22 Dec 2017 16:05:36	U	SMPL
I7042-02 - 1	EFFLUENT-METALS	363	0.1056	-245.1		PPB	22 Dec 2017 16:07:44	U	SMPL
CCV41 - 1	CCV41	16712	4.8508	-245.1		PPB	22 Dec 2017 16:09:51	U	SMPL
CCB41 - 1	CCB41	-90	-0.0259	-245.1		PPB	22 Dec 2017 16:11:59	U	SMPL
I7008-14LX5 - 1		5	0.0017	-245.1		PPB	22 Dec 2017 16:14:09	U	SMPL
I7008-14A - 1		14257	4.1382	-245.1		PPB	22 Dec 2017 16:16:17	U	SMPL
I7009-10LX5 - 1		-71	-0.0204	-245.1		PPB	22 Dec 2017 16:18:25	U	SMPL

LB92345 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Dec 2017 13:06:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
I7009-10A - 1		14403	4.1806		-245.1	PPB	22 Dec 2017 16:20:32	U	SMPL
I7010-12LX5 - 1		-59	-0.0169		-245.1	PPB	22 Dec 2017 16:22:40	U	SMPL
I7010-12A - 1		13876	4.0277		-245.1	PPB	22 Dec 2017 16:24:48	U	SMPL
CCV42 - 1	CCV42	16578	4.8119		-245.1	PPB	22 Dec 2017 16:31:24	U	SMPL
CCB42 - 1	CCB42	-84	-0.0241		-245.1	PPB	22 Dec 2017 16:33:32	U	SMPL