

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

12/22/2017 11:18


Test Total CN

Accepted 12/22/2017 11:18

Factor 1079
 Bias 0.004

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.2818	0.0000	
2	5.0PPBCN	0.009	5.7278	5.0000	
3	10PPBCN	0.013	9.7081	10.0000	
4	50PPBCN	0.050	49.7801	50.0000	
5	100PPBCN	0.096	99.8581	100.0000	
6	250PPBCN	0.235	249.1903	250.0000	
7	500PPBCN	0.468	500.4538	500.0000	

LB92351

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/22/2017 13:39

Test: Total CN ug/d.

Sample Id	Result	Dil. 1 +	Response	Errors

ICV1	97.147	0.0	0.094	
ICB1	0.573	0.0	0.004	
CCV1	245.322	0.0	0.231	
CCB1	0.658	0.0	0.004	
PB105283BL	0.569	0.0	0.004	
PB105283BS	98.198	0.0	0.095	
LOWPB105283	9.653	0.0	0.013	
HIGHPB105283	482.062	0.0	0.451	
I6677-09	0.366	0.0	0.004	
I6677-09DUP	0.131	0.0	0.004	
I6677-09MS	34.039	0.0	0.035	
I6677-09MSD	34.760	0.0	0.036	
I6677-11	0.439	0.0	0.004	
I7022-01	0.580	0.0	0.004	
CCV2	245.209	0.0	0.231	
CCB2	0.798	0.0	0.005	
I7022-03	1.432	0.0	0.005	
I7022-06	1.931	0.0	0.006	
I7030-01	3.683	0.0	0.007	
I7037-01	0.524	0.0	0.004	
CCV3	254.179	0.0	0.239	
CCB3	0.722	0.0	0.004	

N	22
Mean	68.772
SD	126.2677
CV%	183.60

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 22 13:24:32 2017

Fri Dec 22 13:37:25 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.2818	µg/l	12/22/2017 11:08:34	
5.OPPBCN	A	Total CN	P	5.7278	µg/l	12/22/2017 11:08:35	
10PPBCN	A	Total CN	P	9.7081	µg/l	12/22/2017 11:08:36	
50PPBCN	A	Total CN	P	49.7801	µg/l	12/22/2017 11:08:37	
100PPBCN	A	Total CN	P	99.8581	µg/l	12/22/2017 11:08:38	
250PPBCN	A	Total CN	P	249.1903	µg/l	12/22/2017 11:08:39	
500PPBCN	A	Total CN	P	500.4538	µg/l	12/22/2017 11:08:40	
ICV1	S	Total CN	P	97.1474	µg/l	12/22/2017 13:24:33	
ICB1	S	Total CN	P	0.5726	µg/l	12/22/2017 13:24:34	
CCV1	S	Total CN	P	245.322	µg/l	12/22/2017 13:24:37	
CCB1	S	Total CN	P	0.6576	µg/l	12/22/2017 13:24:39	
PB105283BL	S	Total CN	P	0.5693	µg/l	12/22/2017 13:24:40	
PB105283BS	S	Total CN	P	98.1978	µg/l	12/22/2017 13:24:41	
LOWPB105283	S	Total CN	P	9.6533	µg/l	12/22/2017 13:24:42	
HIGHPB105283	S	Total CN	P	482.0619	µg/l	12/22/2017 13:32:05	
I6677-09	S	Total CN	P	0.3662	µg/l	12/22/2017 13:32:06	
I6677-09DUP	S	Total CN	P	0.1308	µg/l	12/22/2017 13:32:07	
I6677-09MS	S	Total CN	P	34.0388	µg/l	12/22/2017 13:32:08	
I6677-09MSD	S	Total CN	P	34.7596	µg/l	12/22/2017 13:32:09	
I6677-11	S	Total CN	P	0.4393	µg/l	12/22/2017 13:32:10	
I7022-01	S	Total CN	P	0.5802	µg/l	12/22/2017 13:32:11	
CCV2	S	Total CN	P	245.2092	µg/l	12/22/2017 13:32:12	
CCB2	S	Total CN	P	0.7983	µg/l	12/22/2017 13:32:13	
I7022-03	S	Total CN	P	1.4318	µg/l	12/22/2017 13:32:14	
I7022-06	S	Total CN	P	1.9312	µg/l	12/22/2017 13:32:15	
I7030-01	S	Total CN	P	3.6831	µg/l	12/22/2017 13:37:20	
I7037-01	S	Total CN	P	0.5244	µg/l	12/22/2017 13:37:21	
CCV3	S	Total CN	P	254.1785	µg/l	12/22/2017 13:37:22	
CCB3	S	Total CN	P	0.7217	µg/l	12/22/2017 13:37:23	