



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: AMANDEEP

Run Number: LB92364

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP60558
Residual chlorine std, 0PPM	WP60552
Residual chlorine std, 0.1PPM	WP60553
Residual chlorine LOQ, 0.2PPM	WP60555
Residual chlorine std, 0.2PPM	WP60556
Residual chlorine std, and CCV , 0.4PPM	WP60557
Residual chlorine std, 1.6PPM	WP60559
Residual chlorine ICV-LCS, 0.4PPM	WP60560
Residual chlorine LOD, 0.05PPM	WP60554
Total Chlorine Powder Pillows	W1960

Intercept: 0.0029

Slope: 0.9880

Regression: 0.999979

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	12/22/2017	09:42
2	CAL2	0.1	1	0.000	0.100	0.100	12/22/2017	09:42
3	CAL3	0.2	1	0.000	0.200	0.200	12/22/2017	09:43
4	CAL4	0.4	1	0.000	0.400	0.400	12/22/2017	09:43
5	CAL5	0.8	1	0.000	0.800	0.800	12/22/2017	09:44
6	CAL6	1.6	1	0.000	1.580	1.580	12/22/2017	09:44



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: AMANDEEP

Run Number: LB92364

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.400	0.400	0.403	12/22/2017	09:47
2	ICB				1	0.000	0.000	0.000	-0.004	12/22/2017	10:05
3	CCV1			0.4	1	0.000	0.400	0.400	0.403	12/22/2017	10:08
4	CCB1				1	0.000	0.000	0.000	-0.004	12/22/2017	10:10
5	LB92364BL	5	5		1	0.000	0.000	0.000	-0.004	12/22/2017	10:13
6	LB92364BS	5	5	0.4	1	0.000	0.410	0.410	0.413	12/22/2017	10:16
7	I7033-01	20	20		1	0.000	0.000	0.000	-0.004	12/22/2017	10:19
8	I7033-02	2	10		1	0.000	0.010	0.010	0.006	12/22/2017	10:23
9	I7033-03	20	20		1	0.000	0.050	0.050	0.047	12/22/2017	10:26
10	I7033-04	5	10		1	0.000	0.040	0.040	0.036	12/22/2017	10:29
11	CCV2			0.4	1	0.000	0.400	0.400	0.403	12/22/2017	10:32
12	CCB2				1	0.000	0.000	0.000	-0.004	12/22/2017	10:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ResCL122217 WorkList ID : 106915 Date : 12/22/2017 9:28:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I7033-01	Residual Chlorine	Cool 4 deg C	THEL01	G22	1-POOL-WATER	12/21/2017	SM4500 Cl G
	Water	I7033-02	Residual Chlorine	Cool 4 deg C	THEL01	G22	2-BTTOOM-OF-POOL-WALL-	12/21/2017	SM4500 Cl G
	Water	I7033-03	Residual Chlorine	Cool 4 deg C	THEL01	G22	3-POOL-WALL-PATCH-AREA	12/21/2017	SM4500 Cl G
	Water	I7033-04	Residual Chlorine	Cool 4 deg C	THEL01	G22	4-WATER-FOUND-ON-THE-S	12/21/2017	SM4500 Cl G

Date/Time 12/24/12 9:30 AM
 Received by: AS
 Relinquished by: CS

Date/Time 12/22/12 12:00 PM
 Received by: AS
 Relinquished by: AS