


7470A

LB 92370
INSTRUMENT ID: CX2

Linear

 μ Abs.:

37232



A= 0.0000e+000

B= 2.6914e-004

C= 3.4237e-002

Rho= 0.9999444

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	0	0.000	0				
0.2	0.200	0.220	0.020	689	0.0 %	689				
2.5	2.500	2.477	-0.023	9075	0.0 %	9075				
5.0	5.000	4.945	-0.055	18246	0.0 %	18246				
7.5	7.500	7.469	-0.031	27625	0.0 %	27625				
10.0	10.000	10.055	0.055	37232	0.0 %	37232				

LB92370 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Dec 2017 09:50:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7470A	PPB	26 Dec 2017 10:04:13	S	Std
0.2 - 1	50.2	689	-	0.2000	7470A	PPB	26 Dec 2017 10:06:22	S	Std
2.5 - 1	52.5	9075	-	2.5000	7470A	PPB	26 Dec 2017 10:08:31	S	Std
5.0 - 1	55	18246	-	5.0000	7470A	PPB	26 Dec 2017 10:10:39	S	Std
7.5 - 1	57.5	27625	-	7.5000	7470A	PPB	26 Dec 2017 10:12:47	S	Std
10.0 - 1	50	37232	-	10.0000	7470A	PPB	26 Dec 2017 10:14:55	S	Std
ICV22 - 1	ICV22	14998	4.0709	-	7470A	PPB	26 Dec 2017 10:17:30	U	SMPL
ICB22 - 1	ICB22	-109	0.0049	-	7470A	PPB	26 Dec 2017 10:19:36	U	SMPL
CCV43 - 1	CCV43	18296	4.9585	-	7470A	PPB	26 Dec 2017 10:21:44	U	SMPL
CCB43 - 1	CCB43	-110	0.0046	-	7470A	PPB	26 Dec 2017 10:23:50	U	SMPL
CRA - 1	CRA	711	0.2256	-	7470A	PPB	26 Dec 2017 10:25:58	U	SMPL
HighStd - 1	HighStd	36668	9.9032	-	7470A	PPB	26 Dec 2017 10:28:04	U	SMPL
ChkStd - 1	ChkStd	25968	7.0234	-	7470A	PPB	26 Dec 2017 10:30:11	U	SMPL
PB105313BL - 1	PBW	-189	-0.0166	-	7470A	PPB	26 Dec 2017 10:32:20	U	SMPL
PB105313BS - 1	LCSW	13816	3.7527	-	7470A	PPB	26 Dec 2017 10:34:29	U	SMPL
PB105313TB - 1	PB105268TB	-93	0.0092	-	7470A	PPB	26 Dec 2017 10:36:37	U	SMPL
I6677-10 - 1	OR-02-122117-A	2	0.0348	-	7470A	PPB	26 Dec 2017 10:38:45	U	SMPL
I6677-12 - 1	OR-02-122117-B	15	0.0383	-	7470A	PPB	26 Dec 2017 10:40:54	U	SMPL
I7022-02 - 1	352	8	0.0364	-	7470A	PPB	26 Dec 2017 10:43:02	U	SMPL
I7022-04 - 1	72-11938	8	0.0364	-	7470A	PPB	26 Dec 2017 10:45:09	U	SMPL
CCV44 - 1	CCV44	18538	5.0236	-	7470A	PPB	26 Dec 2017 10:47:16	U	SMPL
CCB44 - 1	CCB44	-132	-0.0013	-	7470A	PPB	26 Dec 2017 10:49:23	U	SMPL
I7022-07 - 1	160	13	0.0377	-	7470A	PPB	26 Dec 2017 10:51:30	U	SMPL
I7037-02 - 1	VNJ-221-OILY-STONES	9	0.0367	-	7470A	PPB	26 Dec 2017 10:53:37	U	SMPL
I7037-02DUP - 1	VNJ-221-OILY-STONESDUP	18	0.0391	-	7470A	PPB	26 Dec 2017 10:55:43	U	SMPL
I7037-02MS - 1	VNJ-221-OILY-STONESMS	14038	3.8125	-	7470A	PPB	26 Dec 2017 10:57:50	U	SMPL
I7037-02MSD - 1	VNJ-221-OILY-STONESMSD	14462	3.9266	-	7470A	PPB	26 Dec 2017 10:59:58	U	SMPL
I7037-11 - 1	OIL-DEBRIS	-85	0.0114	-	7470A	PPB	26 Dec 2017 11:02:05	U	SMPL
I7037-02LX5 - 1		3	0.0350	-	7470A	PPB	26 Dec 2017 11:04:14	U	SMPL
I7037-02A - 1		13054	3.5476	-	7470A	PPB	26 Dec 2017 11:06:22	U	SMPL
CCV45 - 1	CCV45	18170	4.9246	-	7470A	PPB	26 Dec 2017 11:08:31	U	SMPL
CCB45 - 1	CCB45	-121	0.0017	-	7470A	PPB	26 Dec 2017 11:10:38	U	SMPL