

7471B


LB 92372

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44595



A= 0.0000e+000

B= 2.2696e-004

C= 8.0724e-004

Rho= 0.9997077

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	32	0.000	32				
0.2	0.200	0.210	0.010	923	0.0 %	923	0			
2.5	2.500	2.559	0.059	11272	0.0 %	11272				
5.0	5.000	4.969	-0.031	21889	0.0 %	21889				
7.5	7.500	7.332	-0.168	32302	0.0 %	32302				
10.0	10.000	10.122	0.122	44595	0.0 %	44595				

LB92372 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Dec 2017 12:42:18

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	32	-	0.0000	7471B	PPB	26 Dec 2017 14:04:55	S	Std
0.2 - 1	52.5	923	-	0.2000	7471B	PPB	26 Dec 2017 14:07:13	S	Std
2.5 - 1	55	11272	-	2.5000	7471B	PPB	26 Dec 2017 14:09:30	S	Std
5.0 - 1	57.5	21889	-	5.0000	7471B	PPB	26 Dec 2017 14:11:48	S	Std
7.5 - 1	510	32302	-	7.5000	7471B	PPB	26 Dec 2017 14:14:05	S	Std
10.0 - 1		44595	-	10.0000	7471B	PPB	26 Dec 2017 14:16:22	S	Std
ICV23 - 1	ICV23	16939	3.8452	-7471B	PPB	PPB	26 Dec 2017 14:19:12	U	SMPL
ICB23 - 1	ICB23	-98	-0.0214	-7471B	PPB	PPB	26 Dec 2017 14:21:28	U	SMPL
CCV46 - 1	CCV46	21939	4.9800	-7471B	PPB	PPB	26 Dec 2017 14:23:43	U	SMPL
CCB46 - 1	CCB46	-124	-0.0273	-7471B	PPB	PPB	26 Dec 2017 14:25:59	U	SMPL
CRA - 1	CRA	873	0.1989	-7471B	PPB	PPB	26 Dec 2017 14:28:16	U	SMPL
HighStd - 1	HighStd	44885	10.1878	-7471B	PPB	PPB	26 Dec 2017 14:30:32	U	SMPL
ChkStd - 1	ChkStd	30204	6.8558	-7471B	PPB	PPB	26 Dec 2017 14:32:48	U	SMPL
PB105292BL - 1	PBS	-196	-0.0437	-7471B	PPB	PPB	26 Dec 2017 14:35:06	U	SMPL
PB105292BS - 1	LCSS	15706	3.5654	-7471B	PPB	PPB	26 Dec 2017 14:37:24	U	SMPL
I7021-01 - 1	DB-SW1-7	8643	1.9624	-7471B	PPB	PPB	26 Dec 2017 14:39:42	U	SMPL
I7021-02 - 1	DB-SW1-15	3844	0.8732	-7471B	PPB	PPB	26 Dec 2017 14:41:59	U	SMPL
I7022-01 - 1	352	25248	5.7310	-7471B	PPB	PPB	26 Dec 2017 14:44:16	U	SMPL
I7022-03 - 1	72-11938	29935	6.7948	-7471B	PPB	PPB	26 Dec 2017 14:46:34	U	SMPL
I7022-06 - 1	160	19781	4.4902	-7471B	PPB	PPB	26 Dec 2017 14:48:51	U	SMPL
CCV47 - 1	CCV47	21553	4.8924	-7471B	PPB	PPB	26 Dec 2017 14:51:08	U	SMPL
CCB47 - 1	CCB47	-131	-0.0289	-7471B	PPB	PPB	26 Dec 2017 14:53:25	U	SMPL
I7034-02 - 1	TB-01(4-5)	26963	6.1202	-7471B	PPB	PPB	26 Dec 2017 14:55:41	U	SMPL
I7034-04 - 1	TB-02(0-2)	4898	1.1124	-7471B	PPB	PPB	26 Dec 2017 14:57:58	U	SMPL
I7034-05 - 1	TB-03(3-5)	4956	1.1256	-7471B	PPB	PPB	26 Dec 2017 15:00:15	U	SMPL
I7034-06 - 1	TB-04(0-2)	4576	1.0394	-7471B	PPB	PPB	26 Dec 2017 15:02:31	U	SMPL
I7034-08 - 1	TB-05(5-6)	6870	1.5600	-7471B	PPB	PPB	26 Dec 2017 15:04:48	U	SMPL
I7034-09 - 1	TB-07(0-2)	5509	1.2511	-7471B	PPB	PPB	26 Dec 2017 15:07:05	U	SMPL
I7034-09DUP - 1	TB-07(0-2)DUP	4549	1.0332	-7471B	PPB	PPB	26 Dec 2017 15:09:22	U	SMPL
I7034-09MS - 1	TB-07(0-2)MS	20674	4.6929	-7471B	PPB	PPB	26 Dec 2017 15:11:39	U	SMPL
I7034-09MSD - 1	TB-07(0-2)MSD	20298	4.6076	-7471B	PPB	PPB	26 Dec 2017 15:13:57	U	SMPL
I7034-10 - 1	TB-08(0-2)	25749	5.8447	-7471B	PPB	PPB	26 Dec 2017 15:16:13	U	SMPL
CCV48 - 1	CCV48	21668	4.9185	-7471B	PPB	PPB	26 Dec 2017 15:18:30	U	SMPL
CCB48 - 1	CCB48	-142	-0.0314	-7471B	PPB	PPB	26 Dec 2017 15:20:48	U	SMPL
I7034-11 - 1	TB-09(0-2)	6652	1.5105	-7471B	PPB	PPB	26 Dec 2017 15:23:05	U	SMPL
I7034-12 - 1	TB-09(6-8)	7093	1.6106	-7471B	PPB	PPB	26 Dec 2017 15:25:21	U	SMPL
I7034-13 - 1	TB-11(0-2)	7176	1.6294	-7471B	PPB	PPB	26 Dec 2017 15:27:38	U	SMPL
I7034-15 - 1	TB-12(5.5-7.5)	15337	3.4816	-7471B	PPB	PPB	26 Dec 2017 15:29:54	U	SMPL
I7037-01 - 1	VNJ-221-OILY-STONES	158	0.0367	-7471B	PPB	PPB	26 Dec 2017 15:32:11	U	SMPL
I7037-04 - 1	OIL-DEBRIS	57	0.0137	-7471B	PPB	PPB	26 Dec 2017 15:34:28	U	SMPL
I7043-01 - 1	106-3DS(90-92)	105	0.0246	-7471B	PPB	PPB	26 Dec 2017 15:36:45	U	SMPL
I7043-02 - 1	106-3DS(112-114)	167	0.0387	-7471B	PPB	PPB	26 Dec 2017 15:39:03	U	SMPL
PB105299BL - 1	PBS	41	0.0101	-7471B	PPB	PPB	26 Dec 2017 15:41:20	U	SMPL
PB105299BS - 1	LCSS	15515	3.5220	-7471B	PPB	PPB	26 Dec 2017 15:43:36	U	SMPL
CCV49 - 1	CCV49	21948	4.9821	-7471B	PPB	PPB	26 Dec 2017 15:45:53	U	SMPL
CCB49 - 1	CCB49	-107	-0.0235	-7471B	PPB	PPB	26 Dec 2017 15:48:10	U	SMPL
6677-09 - 1	OR-02-122117-A	1538	0.3499	-7471B	PPB	PPB	26 Dec 2017 15:50:26	U	SMPL
6677-11 - 1	OR-02-122117-B	2154	0.4897	-7471B	PPB	PPB	26 Dec 2017 15:52:42	U	SMPL
6684-05 - 1	BU-01-122217	1329	0.3024	-7471B	PPB	PPB	26 Dec 2017 15:54:59	U	SMPL
6686-05 - 1	TR-04-122217	930	0.2119	-7471B	PPB	PPB	26 Dec 2017 15:57:15	U	SMPL
7023-01 - 1	LB-1	4110	0.9336	-7471B	PPB	PPB	26 Dec 2017 15:59:32	U	SMPL
7023-02 - 1	LB-2	3893	0.8844	-7471B	PPB	PPB	26 Dec 2017 16:01:49	U	SMPL
7023-03 - 1	LB-3	15124	3.4333	-7471B	PPB	PPB	26 Dec 2017 16:04:06	U	SMPL
7023-04 - 1	LB-DUP	7284	1.6540	-7471B	PPB	PPB	26 Dec 2017 16:06:23	U	SMPL
7023-05 - 1	LB-4	16753	3.8030	-7471B	PPB	PPB	26 Dec 2017 16:08:41	U	SMPL
7023-05DUP - 1	LB-4DUP	16638	3.7769	-7471B	PPB	PPB	26 Dec 2017 16:13:39	U	SMPL
CCV50 - 1	CCV50	21851	4.9600	-7471B	PPB	PPB	26 Dec 2017 16:15:56	U	SMPL
CCB50 - 1	CCB50	-118	-0.0260	-7471B	PPB	PPB	26 Dec 2017 16:18:12	U	SMPL
7023-06 - 1	I7023-05MS	37198	8.4431	-7471B	PPB	PPB	26 Dec 2017 16:23:44	U	SMPL
7023-07 - 1	I7023-05MSD	34198	7.7623	-7471B	PPB	PPB	26 Dec 2017 16:26:00	U	SMPL
7023-08 - 1	LB-5	3917	0.8898	-7471B	PPB	PPB	26 Dec 2017 16:28:18	U	SMPL
7023-09 - 1	LB-6	1127	0.2566	-7471B	PPB	PPB	26 Dec 2017 16:30:36	U	SMPL
7049-01 - 1	RT-3074-RT-917	10681	2.4249	-7471B	PPB	PPB	26 Dec 2017 16:32:52	U	SMPL
PB105309BL - 1	PBS	-68	-0.0146	-7471B	PPB	PPB	26 Dec 2017 16:35:09	U	SMPL
PB105309BS - 1	LCSS	15077	3.4226	-7471B	PPB	PPB	26 Dec 2017 16:37:26	U	SMPL
7056-01 - 1	B-9(0-10)COMP	35546	8.0682	-7471B	PPB	PPB	26 Dec 2017 16:39:44	U	SMPL
7064-01 - 1	SAMPLE-1	6131	1.3923	-7471B	PPB	PPB	26 Dec 2017 16:42:02	U	SMPL
7064-03 - 1	SAMPLE-2	15227	3.4567	-7471B	PPB	PPB	26 Dec 2017 16:44:20	U	SMPL
CCV51 - 1	CCV51	22062	5.0079	-7471B	PPB	PPB	26 Dec 2017 16:46:37	U	SMPL
CCB51 - 1	CCB51	-145	-0.0321	-7471B	PPB	PPB	26 Dec 2017 16:48:53	U	SMPL
7064-05 - 1	SAMPLE-3	16818	3.8178	-7471B	PPB	PPB	26 Dec 2017 16:51:10	U	SMPL
7064-07 - 1	SAMPLE-4	12349	2.8035	-7471B	PPB	PPB	26 Dec 2017 16:53:27	U	SMPL
7064-09 - 1	SAMPLE-5	13332	3.0266	-7471B	PPB	PPB	26 Dec 2017 16:55:45	U	SMPL
7064-09DUP - 1	SAMPLE-5DUP	12561	2.8516	-7471B	PPB	PPB	26 Dec 2017 16:58:02	U	SMPL
7064-09MS - 1	SAMPLE-5MS	32495	7.3758	-7471B	PPB	PPB	26 Dec 2017 17:00:18	U	SMPL
7064-09MSD - 1	SAMPLE-5MSD	30539	6.9318	-7471B	PPB	PPB	26 Dec 2017 17:02:35	U	SMPL
7064-11 - 1	SAMPLE-6	13447	3.0527	-7471B	PPB	PPB	26 Dec 2017 17:04:54	U	SMPL
7064-13 - 1	SAMPLE-7	11910	2.7039	-7471B	PPB	PPB	26 Dec 2017 17:07:14	U	SMPL
7064-15 - 1	SAMPLE-8	12524	2.8432	-7471B	PPB	PPB	26 Dec 2017 17:09:32	U	SMPL
7064-17 - 1	SAMPLE-9	8904	2.0216	-7471B	PPB	PPB	26 Dec 2017 17:11:48	U	SMPL
CV52 - 1	CCV52	22031	5.0009	-7471B	PPB	PPB	26 Dec 2017 17:14:05	U	SMPL

LB92372 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Dec 2017 12:42:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB52 - 1	CCB52	-91	-0.0198		-7471B	PPB	26 Dec 2017 17:16:22	U	SMPL
I7064-19 - 1	SAMPLE-10	15523	3.5239		-7471B	PPB	26 Dec 2017 17:18:40	U	SMPL
I7034-09LX5 - 1		989	0.2253		-7471B	PPB	26 Dec 2017 17:20:57	U	SMPL
I7034-09A - 1		22025	4.9995		-7471B	PPB	26 Dec 2017 17:23:14	U	SMPL
I7023-05LX5 - 1		3065	0.6964		-7471B	PPB	26 Dec 2017 17:25:31	U	SMPL
I7023-05A - 1		33392	7.5793		-7471B	PPB	26 Dec 2017 17:27:49	U	SMPL
I7064-09LX5 - 1		2459	0.5589		-7471B	PPB	26 Dec 2017 17:30:06	U	SMPL
I7064-09A - 1		29653	6.7308		-7471B	PPB	26 Dec 2017 17:32:25	U	SMPL
CCV53 - 1	CCV53	21779	4.9437		-7471B	PPB	26 Dec 2017 17:34:42	U	SMPL
CCB53 - 1	CCB53	-119	-0.0262		-7471B	PPB	26 Dec 2017 17:37:02	U	SMPL