



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92394

Temp.: 23.04 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.02 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/27/2017	10:20
2	CAL2	1	Water			6.99	12/27/2017	10:22
3	CAL3	1	Water			10.02	12/27/2017	10:24
4	ICV	1	Water			7.00	12/27/2017	10:26
5	CCV1	1	Water			2.02	12/27/2017	10:28
6	I6685-06	1	Solid	20	20	7.20	12/27/2017	10:30
7	I6685-06DUP	1	Solid	20	20	7.21	12/27/2017	10:32
8	I6685-08	1	Solid	20	20	7.24	12/27/2017	10:34
9	I6685-10	1	Solid	20	20	8.65	12/27/2017	10:36
10	I6685-12	1	Solid	20	20	7.50	12/27/2017	10:38
11	I7077-05	1	Solid	20	20	5.92	12/27/2017	10:40
12	I7077-06	1	Solid	20	20	7.73	12/27/2017	10:42
13	I7078-04	1	Solid	20	20	9.18	12/27/2017	10:44
14	CCV2	1	Water			12.02	12/27/2017	10:46