


7470A

LB92395
INSTRUMENT ID: CY2

Linear

 μ Abs.:

34434



A= 0.0000e+000

B= 2.9184e-004

C= 2.4639e-003

Rho= 0.9999280

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	26	0.000	26				
0.2	0.200	0.180	-0.020	608	0.0 %	608				
2.5	2.500	2.548	0.048	8724	0.0 %	8724				
5.0	5.000	4.988	-0.012	17083	0.0 %	17083				
7.5	7.500	7.422	-0.078	25423	0.0 %	25423				
10.0	10.000	10.052	0.052	34434	0.0 %	34434				

LB92395 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Dec 2017 11:37:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	27 Dec 2017 11:51:45	S	Std
0.2 - 1	50.2	608	-	0.2000	7470A	PPB	27 Dec 2017 11:53:54	S	Std
2.5 - 1	52.5	8724	-	2.5000	7470A	PPB	27 Dec 2017 11:56:02	S	Std
5.0 - 1	55	17083	-	5.0000	7470A	PPB	27 Dec 2017 11:58:10	S	Std
7.5 - 1	57.5	25423	-	7.5000	7470A	PPB	27 Dec 2017 12:00:18	S	Std
10.0 - 1	310	34434	-	10.0000	7470A	PPB	27 Dec 2017 12:02:26	S	Std
ICV24 - 1	ICV24	13013	3.8002	-	7470A	PPB	27 Dec 2017 12:04:57	U	SMPL
ICB24 - 1	ICB24	-86	-0.0226	-	7470A	PPB	27 Dec 2017 12:07:03	U	SMPL
CCV54 - 1	CCV54	17497	5.1088	-	7470A	PPB	27 Dec 2017 12:09:10	U	SMPL
CCB54 - 1	CCB54	-90	-0.0238	-	7470A	PPB	27 Dec 2017 12:11:16	U	SMPL
CRA - 1	CRA	621	0.1837	-	7470A	PPB	27 Dec 2017 12:13:24	U	SMPL
HighStd - 1	HighStd	34722	10.1357	-	7470A	PPB	27 Dec 2017 12:15:30	U	SMPL
ChkStd - 1	ChkStd	24307	7.0962	-	7470A	PPB	27 Dec 2017 12:17:37	U	SMPL
PB105346BL - 1	PBW	-151	-0.0416	-	7470A	PPB	27 Dec 2017 12:19:47	U	SMPL
PB105346BS - 1	LCSW	13228	3.8629	-	7470A	PPB	27 Dec 2017 12:21:55	U	SMPL
PB105346TB - 1	PB105307TB	-48	-0.0115	-	7470A	PPB	27 Dec 2017 12:24:03	U	SMPL
I6684-06 - 1	BU-01-122217	19	0.0080	-	7470A	PPB	27 Dec 2017 12:26:11	U	SMPL
I6686-06 - 1	TR-04-122217	9	0.0051	-	7470A	PPB	27 Dec 2017 12:28:20	U	SMPL
I7049-03 - 1	RT-3074-RT-917	21	0.0086	-	7470A	PPB	27 Dec 2017 12:30:28	U	SMPL
I7056-03 - 1	B-9(0-10)COMP	59	0.0197	-	7470A	PPB	27 Dec 2017 12:32:35	U	SMPL
CCV55 - 1	CCV55	17275	5.0440	-	7470A	PPB	27 Dec 2017 12:34:41	U	SMPL
CCB55 - 1	CCB55	-104	-0.0279	-	7470A	PPB	27 Dec 2017 12:36:48	U	SMPL
I7061-01 - 1	TEC-COMP1	17	0.0074	-	7470A	PPB	27 Dec 2017 12:38:55	U	SMPL
I7069-11 - 1	SB-06-COMP	79	0.0255	-	7470A	PPB	27 Dec 2017 12:41:02	U	SMPL
I7069-12 - 1	SB-25-COMP	16	0.0071	-	7470A	PPB	27 Dec 2017 12:43:08	U	SMPL
I7069-13 - 1	SB-24-COMP	29	0.0109	-	7470A	PPB	27 Dec 2017 12:45:16	U	SMPL
I7069-14 - 1	SB-23-COMP	23	0.0092	-	7470A	PPB	27 Dec 2017 12:47:23	U	SMPL
I7069-15 - 1	SB-22-COMP	42	0.0147	-	7470A	PPB	27 Dec 2017 12:49:31	U	SMPL
I7069-15DUP - 1	SB-22-COMPDUP	2	0.0030	-	7470A	PPB	27 Dec 2017 12:51:39	U	SMPL
I7069-15MS - 1	SB-22-COMPMS	13180	3.8489	-	7470A	PPB	27 Dec 2017 12:53:47	U	SMPL
I7069-15MSD - 1	SB-22-COMPMSD	13252	3.8699	-	7470A	PPB	27 Dec 2017 12:55:55	U	SMPL
PB105347BL - 1	PBW	-77	-0.0200	-	7470A	PPB	27 Dec 2017 12:58:02	U	SMPL
CCV56 - 1	CCV56	17258	5.0390	-	7470A	PPB	27 Dec 2017 13:00:09	U	SMPL
CCB56 - 1	CCB56	-90	-0.0238	-	7470A	PPB	27 Dec 2017 13:02:16	U	SMPL
PB105347BS - 1	LCSW	13924	4.0660	-	7470A	PPB	27 Dec 2017 13:04:24	U	SMPL
I7053-01 - 1	MW-18	141	0.0436	-	7470A	PPB	27 Dec 2017 13:06:31	U	SMPL
I7053-02 - 1	MW-3R	1694	0.4968	-	7470A	PPB	27 Dec 2017 13:08:39	U	SMPL
I7053-03 - 1	MW-12R	516	0.1531	-	7470A	PPB	27 Dec 2017 13:10:46	U	SMPL
I7053-04 - 1	MW-12RF	285	0.0856	-	7470A	PPB	27 Dec 2017 13:12:53	U	SMPL
7053-04DUP - 1	MW-12RFDUP	265	0.0798	-	7470A	PPB	27 Dec 2017 13:15:01	U	SMPL
7053-04MS - 1	MW-12RFMS	14532	4.2435	-	7470A	PPB	27 Dec 2017 13:17:09	U	SMPL
7053-04MSD - 1	MW-12RFMSD	14452	4.2201	-	7470A	PPB	27 Dec 2017 13:19:17	U	SMPL
CCV57 - 1	CCV57	17262	5.0402	-	7470A	PPB	27 Dec 2017 13:21:26	U	SMPL
CCB57 - 1	CCB57	-108	-0.0291	-	7470A	PPB	27 Dec 2017 13:23:33	U	SMPL
7069-15LX5 - 1		-9	-0.0002	-	7470A	PPB	27 Dec 2017 13:25:41	U	SMPL
7069-15A - 1		13699	4.0004	-	7470A	PPB	27 Dec 2017 13:27:48	U	SMPL
7053-04LX5 - 1		-43	-0.0101	-	7470A	PPB	27 Dec 2017 13:29:55	U	SMPL
7053-04A - 1		14836	4.3322	-	7470A	PPB	27 Dec 2017 13:32:03	U	SMPL
CCV58 - 1	CCV58	17332	5.0606	-	7470A	PPB	27 Dec 2017 13:34:10	U	SMPL
CCB58 - 1	CCB58	-112	-0.0302	-	7470A	PPB	27 Dec 2017 13:36:17	U	SMPL