



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92396

Temp.: 23.04 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/27/2017	10:20
2	CAL2	1	Water			6.99	12/27/2017	10:22
3	CAL3	1	Water			10.02	12/27/2017	10:24
4	ICV	1	Water			7.00	12/27/2017	10:26
5	CCV1	1	Water			2.01	12/27/2017	11:20
6	I7080-01	1	Water			4.45	12/27/2017	11:22
7	I7080-01DUP	1	Water			4.46	12/27/2017	11:24
8	I7080-02	1	Water			4.57	12/27/2017	11:26
9	I7080-03	1	Water			5.04	12/27/2017	11:28
10	CCV2	1	Water			12.01	12/27/2017	11:30