


245.1

LB 92403  
INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

33788



A= 0.0000e+000

B= 2.9462e-004

C= -1.4270e-003

Rho= 0.9999124

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | -0.019 | -0.019 | -59   | 0.000      | -59   |       |       |       |       |
| 0.05   | 0.050  | 0.053  | 0.003  | 184   | 0.0 %      | 184   |       |       |       |       |
| 0.2    | 0.200  | 0.229  | 0.029  | 783   | 0.0 %      | 783   |       |       |       |       |
| 2.5    | 2.500  | 2.490  | -0.010 | 8456  | 0.0 %      | 8456  |       |       |       |       |
| 5.0    | 5.000  | 4.943  | -0.057 | 16782 | 0.0 %      | 16782 |       |       |       |       |
| 7.5    | 7.500  | 7.601  | 0.101  | 25803 | 0.0 %      | 25803 |       |       |       |       |
| 10.0   | 10.000 | 9.953  | -0.047 | 33788 | 0.0 %      | 33788 |       |       |       |       |

## LB92403 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Dec 2017 14:17:39

| Sample ID       | Extended ID          | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|----------------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                   | -59        | -       | 0.0000   | 245.1  | PPB   | 27 Dec 2017 14:54:58 | S    | Std  |
| 0.05 - 1        | 50.5                 | 184        | -       | 0.0500   | 245.1  | PPB   | 27 Dec 2017 14:57:06 | S    | Std  |
| 0.2 - 1         | 50.2                 | 783        | -       | 0.2000   | 245.1  | PPB   | 27 Dec 2017 14:59:15 | S    | Std  |
| 2.5 - 1         | 52.5                 | 8456       | -       | 2.5000   | 245.1  | PPB   | 27 Dec 2017 15:01:23 | S    | Std  |
| 5.0 - 1         | 55                   | 16782      | -       | 5.0000   | 245.1  | PPB   | 27 Dec 2017 15:03:31 | S    | Std  |
| 7.5 - 1         | 57.5                 | 25803      | -       | 7.5000   | 245.1  | PPB   | 27 Dec 2017 15:05:38 | S    | Std  |
| 10.0 - 1        | 510                  | 33788      | -       | 10.0000  | 245.1  | PPB   | 27 Dec 2017 15:07:47 | S    | Std  |
| ICV25 - 1       | ICV25                | 13155      | 3.8743  | -245.1   |        | PPB   | 27 Dec 2017 15:10:27 | U    | SMPL |
| ICB25 - 1       | ICB25                | -91        | -0.0282 | -245.1   |        | PPB   | 27 Dec 2017 15:12:33 | U    | SMPL |
| CCV59 - 1       | CCV59                | 16796      | 4.9470  | -245.1   |        | PPB   | 27 Dec 2017 15:14:41 | U    | SMPL |
| CCB59 - 1       | CCB59                | -106       | -0.0327 | -245.1   |        | PPB   | 27 Dec 2017 15:16:47 | U    | SMPL |
| CRA - 1         | CRA                  | 749        | 0.2192  | -245.1   |        | PPB   | 27 Dec 2017 15:18:55 | U    | SMPL |
| HighStd - 1     | HighStd              | 33644      | 9.9108  | -245.1   |        | PPB   | 27 Dec 2017 15:21:02 | U    | SMPL |
| ChkStd - 1      | ChkStd               | 23736      | 6.9917  | -245.1   |        | PPB   | 27 Dec 2017 15:23:09 | U    | SMPL |
| PB105348BL - 1  | PBW                  | -113       | -0.0347 | -245.1   |        | PPB   | 27 Dec 2017 15:25:18 | U    | SMPL |
| PB105348BS - 1  | LCSW                 | 13412      | 3.9500  | -245.1   |        | PPB   | 27 Dec 2017 15:31:25 | U    | SMPL |
| I6878-02 - 1    | TW-3-DEP             | 287        | 0.0831  | -245.1   |        | PPB   | 27 Dec 2017 15:33:32 | U    | SMPL |
| I7058-01 - 1    | 122117-TIDO-F-D-B    | 64         | 0.0174  | -245.1   |        | PPB   | 27 Dec 2017 15:35:40 | U    | SMPL |
| I7058-02 - 1    | 122117-TIDO-F-D-S    | 51         | 0.0136  | -245.1   |        | PPB   | 27 Dec 2017 15:37:48 | U    | SMPL |
| I7058-03 - 1    | 122117-TIDO-F-U-B    | 59         | 0.0160  | -245.1   |        | PPB   | 27 Dec 2017 15:39:57 | U    | SMPL |
| I7058-04 - 1    | 122117-TIDO-F-U-S    | 24         | 0.0056  | -245.1   |        | PPB   | 27 Dec 2017 15:42:03 | U    | SMPL |
| CCV60 - 1       | CCV60                | 16880      | 4.9718  | -245.1   |        | PPB   | 27 Dec 2017 15:44:10 | U    | SMPL |
| CCB60 - 1       | CCB60                | -78        | -0.0244 | -245.1   |        | PPB   | 27 Dec 2017 15:46:17 | U    | SMPL |
| I7058-05 - 1    | 122117-SIDI-F-D-B    | 58         | 0.0157  | -245.1   |        | PPB   | 27 Dec 2017 15:48:24 | U    | SMPL |
| I7058-06 - 1    | 122117-SIDI-F-D-S    | 42         | 0.0109  | -245.1   |        | PPB   | 27 Dec 2017 15:50:31 | U    | SMPL |
| I7058-07 - 1    | 122117-SIDI-F-U-B    | 22         | 0.0051  | -245.1   |        | PPB   | 27 Dec 2017 15:52:38 | U    | SMPL |
| I7058-08 - 1    | 122117-SIDI-F-U-S    | 31         | 0.0077  | -245.1   |        | PPB   | 27 Dec 2017 15:54:45 | U    | SMPL |
| I7058-09 - 1    | 122117-JETT-F-D-B    | 23         | 0.0053  | -245.1   |        | PPB   | 27 Dec 2017 15:56:52 | U    | SMPL |
| I7058-10 - 1    | 122117-JETT-F-D-M    | 49         | 0.0130  | -245.1   |        | PPB   | 27 Dec 2017 15:59:00 | U    | SMPL |
| I7058-11 - 1    | 122117-JETT-F-D-S    | 55         | 0.0148  | -245.1   |        | PPB   | 27 Dec 2017 16:01:08 | U    | SMPL |
| I7058-12 - 1    | 122117-JETT-F-U-B    | 76         | 0.0210  | -245.1   |        | PPB   | 27 Dec 2017 16:03:16 | U    | SMPL |
| I7058-13 - 1    | 122117-JETT-F-U-M    | 51         | 0.0136  | -245.1   |        | PPB   | 27 Dec 2017 16:05:25 | U    | SMPL |
| I7058-14 - 1    | 122117-JETT-F-U-S    | 37         | 0.0095  | -245.1   |        | PPB   | 27 Dec 2017 16:07:32 | U    | SMPL |
| CCV61 - 1       | CCV61                | 16848      | 4.9624  | -245.1   |        | PPB   | 27 Dec 2017 16:09:39 | U    | SMPL |
| CCB61 - 1       | CCB61                | -73        | -0.0229 | -245.1   |        | PPB   | 27 Dec 2017 16:11:46 | U    | SMPL |
| I7058-14DUP - 1 | 122117-JETT-F-U-SDUP | 47         | 0.0124  | -245.1   |        | PPB   | 27 Dec 2017 16:15:47 | U    | SMPL |
| I7058-14MS - 1  | 122117-JETT-F-U-SMS  | 14611      | 4.3033  | -245.1   |        | PPB   | 27 Dec 2017 16:17:53 | U    | SMPL |
| I7058-14MSD - 1 | 122117-JETT-F-U-SMSD | 14373      | 4.2332  | -245.1   |        | PPB   | 27 Dec 2017 16:20:00 | U    | SMPL |
| I7058-14LX5 - 1 |                      | -63        | -0.0200 | -245.1   |        | PPB   | 27 Dec 2017 16:22:07 | U    | SMPL |
| I7058-14A - 1   |                      | 15327      | 4.5142  | -245.1   |        | PPB   | 27 Dec 2017 16:24:14 | U    | SMPL |
| CCV62 - 1       | CCV62                | 16798      | 4.9476  | -245.1   |        | PPB   | 27 Dec 2017 16:26:22 | U    | SMPL |
| CCB62 - 1       | CCB62                | -88        | -0.0274 | -245.1   |        | PPB   | 27 Dec 2017 16:28:30 | U    | SMPL |