



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92413

Temp.: 23.02 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.02	12/28/2017	09:50
2	CAL2	1	Water			6.99	12/28/2017	09:52
3	CAL3	1	Water			10.01	12/28/2017	09:54
4	ICV	1	Water			7.00	12/28/2017	09:56
5	CCV1	1	Water			2.01	12/28/2017	09:58
6	I7082-01	1	Solid	20	20	7.89	12/28/2017	10:00
7	I7083-01	1	Solid	20	20	8.01	12/28/2017	10:02
8	I7083-02	1	Solid	20	20	7.16	12/28/2017	10:04
9	I7083-02DUP	1	Solid	20	20	7.17	12/28/2017	10:06
10	I7083-03	1	Solid	20	20	6.82	12/28/2017	10:08
11	CCV2	1	Water			12.01	12/28/2017	10:10