


7470A

LB92414
INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.9097e-004

C= -2.3173e-002

Rho= 0.9997955

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	139	0.000	139				
0.2	0.200	0.160	-0.040	628	0.0 %	628				
2.5	2.500	2.431	-0.069	8435	0.0 %	8435				
5.0	5.000	5.132	0.132	17717	0.0 %	17717				
7.5	7.500	7.545	0.045	26011	0.0 %	26011				
10.0	10.000	9.915	-0.085	34154	0.0 %	34154				

LB92414 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Dec 2017 10:37:18

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	139	-	0.0000	7470A	PPB	28 Dec 2017 10:45:09	S	Std
0.2 - 1	50.2	628	-	0.2000	7470A	PPB	28 Dec 2017 10:47:17	S	Std
2.5 - 1	52.5	8435	-	2.5000	7470A	PPB	28 Dec 2017 10:49:26	S	Std
5.0 - 1	55	17717	-	5.0000	7470A	PPB	28 Dec 2017 10:51:34	S	Std
7.5 - 1	57.5	26011	-	7.5000	7470A	PPB	28 Dec 2017 10:53:42	S	Std
10.0 - 1	510	34154	-	10.0000	7470A	PPB	28 Dec 2017 10:58:22	S	Std
ICV26 - 1	ICV26	12773	3.6934	-	7470A	PPB	28 Dec 2017 11:00:53	U	SMPL
ICB26 - 1	ICB26	21	-0.0171	-	7470A	PPB	28 Dec 2017 11:03:00	U	SMPL
CCV63 - 1	CCV63	17612	5.1014	-	7470A	PPB	28 Dec 2017 11:05:07	U	SMPL
CCB63 - 1	CCB63	4	-0.0220	-	7470A	PPB	28 Dec 2017 11:07:13	U	SMPL
CRA - 1	CRA	624	0.1584	-	7470A	PPB	28 Dec 2017 11:09:21	U	SMPL
HighStd - 1	HighStd	34415	9.9906	-	7470A	PPB	28 Dec 2017 11:11:28	U	SMPL
ChkStd - 1	ChkStd	24553	7.1211	-	7470A	PPB	28 Dec 2017 11:13:35	U	SMPL
PB105374BL - 1	PBW	-123	-0.0590	-	7470A	PPB	28 Dec 2017 11:15:44	U	SMPL
PB105374BS - 1	LCSW	13554	3.9207	-	7470A	PPB	28 Dec 2017 11:17:54	U	SMPL
PB105374TB - 1	PB105357TB	-48	-0.0371	-	7470A	PPB	28 Dec 2017 11:20:01	U	SMPL
I6685-06 - 1	AU-05-122617-A	39	-0.0118	-	7470A	PPB	28 Dec 2017 11:22:10	U	SMPL
I6685-06DUP - 1	AU-05-122617-ADUP	56	-0.0069	-	7470A	PPB	28 Dec 2017 11:24:18	U	SMPL
I6685-06MS - 1	AU-05-122617-AMS	12787	3.6975	-	7470A	PPB	28 Dec 2017 11:26:26	U	SMPL
I6685-06MSD - 1	AU-05-122617-AMSD	13169	3.8086	-	7470A	PPB	28 Dec 2017 11:28:33	U	SMPL
CCV64 - 1	CCV64	17382	5.0345	-	7470A	PPB	28 Dec 2017 11:30:40	U	SMPL
CCB64 - 1	CCB64	16	-0.0185	-	7470A	PPB	28 Dec 2017 11:32:46	U	SMPL
I6685-08 - 1	AU-05-122617-B	168	0.0257	-	7470A	PPB	28 Dec 2017 11:34:54	U	SMPL
I6685-10 - 1	AU-05-122617-C	167	0.0254	-	7470A	PPB	28 Dec 2017 11:37:00	U	SMPL
I6685-12 - 1	AU-05-122617-D	187	0.0312	-	7470A	PPB	28 Dec 2017 11:39:07	U	SMPL
I6685-06LX5 - 1		73	-0.0019	-	7470A	PPB	28 Dec 2017 11:41:14	U	SMPL
I6685-06A - 1		13634	3.9439	-	7470A	PPB	28 Dec 2017 11:43:21	U	SMPL
CCV65 - 1	CCV65	17668	5.1177	-	7470A	PPB	28 Dec 2017 11:45:29	U	SMPL
CCB65 - 1	CCB65	12	-0.0197	-	7470A	PPB	28 Dec 2017 11:47:37	U	SMPL