

LIB 92418


7471A

INSTRUMENT 10: CV1

Linear

μ Abs.:

43135



A= 0.0000e+000

B= 2.3022e-004

C= 9.1024e-003

Rho= 0.9998661

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	73	0.000	73				
0.2	0.200	0.209	0.009	868	0.0 %	868				
2.5	2.500	2.430	-0.070	10517	0.0 %	10517				
5.0	5.000	4.984	-0.016	21609	0.0 %	21609				
7.5	7.500	7.611	0.111	33021	0.0 %	33021	0			
10.0	10.000	9.940	-0.060	43135	0.0 %	43135				

LB92418 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Dec 2017 14:16:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	73	-	0.0000	7471A	PPB	28 Dec 2017 14:19:17	S	Std
0.2 - 1	50.2	868	-	0.2000	7471A	PPB	28 Dec 2017 14:21:35	S	Std
2.5 - 1	52.5	10517	-	2.5000	7471A	PPB	28 Dec 2017 14:23:53	S	Std
5.0 - 1	55	21609	-	5.0000	7471A	PPB	28 Dec 2017 14:26:10	S	Std
7.5 - 1	57.5	33021	-	7.5000	7471A	PPB	28 Dec 2017 14:28:27	S	Std
10.0 - 1	60	43135	-	10.0000	7471A	PPB	28 Dec 2017 14:30:45	S	Std
ICV27 - 1	ICV27	16837	3.8853	-	7471A	PPB	28 Dec 2017 14:33:45	U	SMPL
ICB27 - 1	ICB27	-69	-0.0068	-	7471A	PPB	28 Dec 2017 14:36:01	U	SMPL
CCV66 - 1	CCV66	21808	5.0298	-	7471A	PPB	28 Dec 2017 14:38:18	U	SMPL
CCB66 - 1	CCB66	-88	-0.0112	-	7471A	PPB	28 Dec 2017 14:40:33	U	SMPL
CRA - 1	CRA	829	0.2000	-	7471A	PPB	28 Dec 2017 14:42:50	U	SMPL
HighStd - 1	HighStd	43672	10.0633	-	7471A	PPB	28 Dec 2017 14:45:06	U	SMPL
ChkStd - 1	ChkStd	29955	6.9054	-	7471A	PPB	28 Dec 2017 14:47:22	U	SMPL
PB105376BL - 1	PBS	-71	-0.0072	-	7471A	PPB	28 Dec 2017 14:49:40	U	SMPL
PB105376BS - 1	LCSS	15317	3.5354	-	7471A	PPB	28 Dec 2017 14:51:58	U	SMPL
I6685-05 - 1	AU-05-122617-A	2230	0.5225	-	7471A	PPB	28 Dec 2017 14:59:05	U	SMPL
I6685-05DUP - 1	AU-05-122617-ADUP	2223	0.5209	-	7471A	PPB	28 Dec 2017 15:01:22	U	SMPL
I6685-05MS - 1	AU-05-122617-AMS	17356	4.0048	-	7471A	PPB	28 Dec 2017 15:03:39	U	SMPL
I6685-05MSD - 1	AU-05-122617-AMSD	17329	3.9986	-	7471A	PPB	28 Dec 2017 15:05:57	U	SMPL
I6685-07 - 1	AU-05-122617-B	1276	0.3029	-	7471A	PPB	28 Dec 2017 15:08:14	U	SMPL
CCV67 - 1	CCV67	21624	4.9874	-	7471A	PPB	28 Dec 2017 15:10:29	U	SMPL
CCB67 - 1	CCB67	-101	-0.0141	-	7471A	PPB	28 Dec 2017 15:12:45	U	SMPL
I6685-09 - 1	AU-05-122617-C	2126	0.4986	-	7471A	PPB	28 Dec 2017 15:15:02	U	SMPL
I6685-11 - 1	AU-05-122617-D	2473	0.5784	-	7471A	PPB	28 Dec 2017 15:17:18	U	SMPL
I7030-01 - 1	CS-TP-BANK-4	4690	1.0888	-	7471A	PPB	28 Dec 2017 15:19:34	U	SMPL
I7078-02 - 1	PB-05(0-8)	8994	2.0797	-	7471A	PPB	28 Dec 2017 15:21:50	U	SMPL
I6685-05LX5 - 1		422	0.1063	-	7471A	PPB	28 Dec 2017 15:24:07	U	SMPL
I6685-05A - 1		18604	4.2921	-	7471A	PPB	28 Dec 2017 15:26:23	U	SMPL
CCV68 - 1	CCV68	21647	4.9927	-	7471A	PPB	28 Dec 2017 15:28:41	U	SMPL
CCB68 - 1	CCB68	-81	-0.0095	-	7471A	PPB	28 Dec 2017 15:30:58	U	SMPL