



### Analytical Summary Report

Analysis Method: 9045C  
Parameter: Corrosivity  
Run Number: LB92429  
BalanceID: WC SC-4

Analyst By : Sweta  
Supervisor Review By : apatel  
Temp.: 23.06 °C  
SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.02 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/28/2017	17:30
2	CAL2	1	Water			6.99	12/28/2017	17:32
3	CAL3	1	Water			10.01	12/28/2017	17:34
4	ICV	1	Water			7.00	12/28/2017	17:36
5	CCV1	1	Water			2.02	12/28/2017	17:38
6	I6686-08	1	Solid	20	20	8.72	12/28/2017	17:40
7	I7092-02	1	Solid	20	20	6.42	12/28/2017	17:42
8	I7092-02DUP	1	Solid	20	20	6.44	12/28/2017	17:44
9	CCV2	1	Water			12.01	12/28/2017	17:46