


7470A

LB92434
INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.8483e-004

C= 3.3093e-003

Rho= 0.9999909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	16	0.000	16				
0.2	0.200	0.208	0.008	720	0.0 %	720				
2.5	2.500	2.484	-0.016	8710	0.0 %	8710				
5.0	5.000	5.005	0.005	17559	0.0 %	17559				
7.5	7.500	7.474	-0.026	26229	0.0 %	26229				
10.0	10.000	10.021	0.021	35169	0.0 %	35169				

LB92434 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Dec 2017 11:24:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	16	-	0.0000	7470A	PPB	29 Dec 2017 11:24:19	S	Std
0.2 - 1	S02	720	-	0.2000	7470A	PPB	29 Dec 2017 11:26:28	S	Std
2.5 - 1	S2.5	8710	-	2.5000	7470A	PPB	29 Dec 2017 11:28:36	S	Std
5.0 - 1	S5	17559	-	5.0000	7470A	PPB	29 Dec 2017 11:30:45	S	Std
7.5 - 1	S7.5	26229	-	7.5000	7470A	PPB	29 Dec 2017 11:32:52	S	Std
10.0 - 1	S10	35169	-	10.0000	7470A	PPB	29 Dec 2017 11:35:01	S	Std
ICV28 - 1	ICV28	13432	3.8292	-	7470A	PPB	29 Dec 2017 11:40:05	U	SMPL
ICB28 - 1	ICB28	-67	-0.0158	-	7470A	PPB	29 Dec 2017 11:42:11	U	SMPL
CCV69 - 1	CCV69	17482	4.9828	-	7470A	PPB	29 Dec 2017 11:44:19	U	SMPL
CCB69 - 1	CCB69	-97	-0.0243	-	7470A	PPB	29 Dec 2017 11:46:26	U	SMPL
CRA - 1	CRA	712	0.2061	-	7470A	PPB	29 Dec 2017 11:48:33	U	SMPL
HighStd - 1	HighStd	34404	9.8027	-	7470A	PPB	29 Dec 2017 11:50:40	U	SMPL
ChkStd - 1	ChkStd	24181	6.8909	-	7470A	PPB	29 Dec 2017 11:52:47	U	SMPL
PB105394BL - 1	PBW	-153	-0.0403	-	7470A	PPB	29 Dec 2017 11:54:55	U	SMPL
PB105394BS - 1	LCSW	12581	3.5868	-	7470A	PPB	29 Dec 2017 11:57:04	U	SMPL
PB105394TB - 1	PB105375TB	-66	-0.0155	-	7470A	PPB	29 Dec 2017 11:59:12	U	SMPL
I7077-05 - 1	SB-07-COMP	3	0.0042	-	7470A	PPB	29 Dec 2017 12:01:20	U	SMPL
I7077-06 - 1	SB-21-COMP	3	0.0042	-	7470A	PPB	29 Dec 2017 12:03:28	U	SMPL
I7078-04 - 1	PB-05(0-8)	27	0.0110	-	7470A	PPB	29 Dec 2017 12:05:37	U	SMPL
I7082-01 - 1	OWS-1	0	0.0033	-	7470A	PPB	29 Dec 2017 12:07:43	U	SMPL
CCV70 - 1	CCV70	17665	5.0349	-	7470A	PPB	29 Dec 2017 12:09:50	U	SMPL
CCB70 - 1	CCB70	-84	-0.0206	-	7470A	PPB	29 Dec 2017 12:11:56	U	SMPL
I7083-01 - 1	OWS-3	20	0.0090	-	7470A	PPB	29 Dec 2017 12:14:04	U	SMPL
I7083-02 - 1	OWS-2	114	0.0358	-	7470A	PPB	29 Dec 2017 12:16:11	U	SMPL
I7083-03 - 1	OWS-1	114	0.0358	-	7470A	PPB	29 Dec 2017 12:18:17	U	SMPL
I7083-03DUP - 1	OWS-1DUP	98	0.0312	-	7470A	PPB	29 Dec 2017 12:20:24	U	SMPL
I7083-03MS - 1	OWS-1MS	13474	3.8412	-	7470A	PPB	29 Dec 2017 12:22:32	U	SMPL
I7083-03MSD - 1	OWS-1MSD	13408	3.8224	-	7470A	PPB	29 Dec 2017 12:24:39	U	SMPL
I7083-03LX5 - 1		-46	-0.0098	-	7470A	PPB	29 Dec 2017 12:26:47	U	SMPL
I7083-03A - 1		13938	3.9733	-	7470A	PPB	29 Dec 2017 12:28:56	U	SMPL
CCV71 - 1	CCV71	17743	5.0571	-	7470A	PPB	29 Dec 2017 12:31:05	U	SMPL
CCB71 - 1	CCB71	-80	-0.0195	-	7470A	PPB	29 Dec 2017 12:33:12	U	SMPL