


7471B

LB 92440
INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2912e-004

C= 1.5504e-003

Rho= 0.9997826

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	28	0.000	28				
0.2	0.200	0.204	0.004	884	0.0 %	884	0			
2.5	2.500	2.471	-0.029	10779	0.0 %	10779				
5.0	5.000	4.948	-0.052	21590	0.0 %	21590				
7.5	7.500	7.656	0.156	33407	0.0 %	33407				
10.0	10.000	9.912	-0.088	43256	0.0 %	43256				

LB92440 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Dec 2017 14:05:57					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.0000	7471B	PPB	29 Dec 2017 14:09:44	S	Std
0.2 - 1	50.2	884	-	0.2000	7471B	PPB	29 Dec 2017 14:12:02	S	Std
2.5 - 1	52.5	10779	-	2.5000	7471B	PPB	29 Dec 2017 14:14:19	S	Std
5.0 - 1	55	21590	-	5.0000	7471B	PPB	29 Dec 2017 14:16:37	S	Std
7.5 - 1	57.5	33407	-	7.5000	7471B	PPB	29 Dec 2017 14:18:54	S	Std
10.0 - 1	510	43256	-	10.0000	7471B	PPB	29 Dec 2017 14:21:11	S	Std
ICV29 - 1	ICV29	16015	3.6710	-	7471B	PPB	29 Dec 2017 14:26:28	U	SMPL
ICB29 - 1	ICB29	-96	-0.0204	-	7471B	PPB	29 Dec 2017 14:28:43	U	SMPL
CCV72 - 1	CCV72	21769	4.9893	-	7471B	PPB	29 Dec 2017 14:30:59	U	SMPL
CCB72 - 1	CCB72	-117	-0.0253	-	7471B	PPB	29 Dec 2017 14:33:15	U	SMPL
CRA - 1	CRA	833	0.1924	-	7471B	PPB	29 Dec 2017 14:35:32	U	SMPL
HighStd - 1	HighStd	43421	9.9503	-	7471B	PPB	29 Dec 2017 14:37:48	U	SMPL
ChkStd - 1	ChkStd	29948	6.8633	-	7471B	PPB	29 Dec 2017 14:40:04	U	SMPL
PB105415BL - 1	PBS	-164	-0.0360	-	7471B	PPB	29 Dec 2017 14:42:22	U	SMPL
PB105415BS - 1	LCSS	15044	3.4485	-	7471B	PPB	29 Dec 2017 14:44:40	U	SMPL
I6686-07 - 1	TR-05-122817	657	0.1521	-	7471B	PPB	29 Dec 2017 14:46:57	U	SMPL
I6696-05 - 1	CS-TP102-A	4296	0.9859	-	7471B	PPB	29 Dec 2017 14:49:14	U	SMPL
I6696-06 - 1	CS-TP102-B(1-2)	22152	5.0771	-	7471B	PPB	29 Dec 2017 14:51:32	U	SMPL
I6696-07 - 1	CS-TP102-B(2+)	23405	5.3642	-	7471B	PPB	29 Dec 2017 14:53:50	U	SMPL
I6696-08 - 1	CS-TP-BLANK-1	5840	1.3396	-	7471B	PPB	29 Dec 2017 14:56:07	U	SMPL
CCV73 - 1	CCV73	21942	5.0290	-	7471B	PPB	29 Dec 2017 14:58:24	U	SMPL
CCB73 - 1	CCB73	-113	-0.0243	-	7471B	PPB	29 Dec 2017 15:00:40	U	SMPL
I7091-01 - 1	RO-CHRT25337	115	0.0279	-	7471B	PPB	29 Dec 2017 15:02:57	U	SMPL
I7093-01 - 1	B-6-0-2	46113	10.5671	-	7471B	PPB	29 Dec 2017 15:05:13	U	SMPL
I7093-02 - 1	B-4-0-2	12117	2.7778	-	7471B	PPB	29 Dec 2017 15:07:28	U	SMPL
I7093-03 - 1	B-10-0-2	38698	8.8682	-	7471B	PPB	29 Dec 2017 15:09:47	U	SMPL
I7093-04 - 1	B-9-0-2	13669	3.1334	-	7471B	PPB	29 Dec 2017 15:12:04	U	SMPL
I7093-05 - 1	B-3-0-2	5427	1.2450	-	7471B	PPB	29 Dec 2017 15:14:23	U	SMPL
I7093-06 - 1	B-2-0-2	1670	0.3842	-	7471B	PPB	29 Dec 2017 15:16:40	U	SMPL
I7093-07 - 1	B-8-0-2	13247	3.0367	-	7471B	PPB	29 Dec 2017 15:18:57	U	SMPL
I7093-08 - 1	B-1-0-2	6844	1.5697	-	7471B	PPB	29 Dec 2017 15:21:15	U	SMPL
I7095-01 - 1	DRUMS-1-4	1449	0.3335	-	7471B	PPB	29 Dec 2017 15:23:31	U	SMPL
CCV74 - 1	CCV74	21786	4.9932	-	7471B	PPB	29 Dec 2017 15:25:48	U	SMPL
CCB74 - 1	CCB74	-111	-0.0239	-	7471B	PPB	29 Dec 2017 15:28:04	U	SMPL
I7100-01 - 1	EPE-105-1(0-5)	5747	1.3183	-	7471B	PPB	29 Dec 2017 15:35:12	U	SMPL
I7100-01DUP - 1	EPE-105-1(0-5)DUP	5762	1.3218	-	7471B	PPB	29 Dec 2017 15:37:28	U	SMPL
I7100-01MS - 1	EPE-105-1(0-5)MS	21378	4.8997	-	7471B	PPB	29 Dec 2017 15:42:56	U	SMPL
I7100-01MSD - 1	EPE-105-1(0-5)MSD	23602	5.4093	-	7471B	PPB	29 Dec 2017 15:47:53	U	SMPL
I7100-02 - 1	EPE-105-1(5-10)	1086	0.2504	-	7471B	PPB	29 Dec 2017 15:50:09	U	SMPL
I7100-03 - 1	EPE-106-2(0-5)	25722	5.8951	-	7471B	PPB	29 Dec 2017 15:52:27	U	SMPL
I7100-04 - 1	EPE-106-2(5-10)	54006	12.3756	-	7471B	PPB	29 Dec 2017 15:54:44	U	SMPL
I7100-01LX5 - 1		829	0.1915	-	7471B	PPB	29 Dec 2017 15:57:02	U	SMPL
I7100-01A - 1		21517	4.9316	-	7471B	PPB	29 Dec 2017 15:59:22	U	SMPL
7093-01DLX2 - 1		22336	5.1192	-	7471B	PPB	29 Dec 2017 16:01:38	U	SMPL
CCV75 - 1	CCV75	21931	5.0264	-	7471B	PPB	29 Dec 2017 16:03:56	U	SMPL
CCB75 - 1	CCB75	-138	-0.0301	-	7471B	PPB	29 Dec 2017 16:06:13	U	SMPL
7100-04DLX2 - 1		26548	6.0843	-	7471B	PPB	29 Dec 2017 16:08:30	U	SMPL
CCV76 - 1	CCV76	21857	5.0095	-	7471B	PPB	29 Dec 2017 16:10:46	U	SMPL
CCB76 - 1	CCB76	-162	-0.0356	-	7471B	PPB	29 Dec 2017 16:13:04	U	SMPL