

7470A


LB 92448

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35100



A= 0.0000e+000

B= 2.8331e-004

C= 7.0454e-002

Rho= 0.9996747

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	-110	0.000	-110				
0.2	0.200	0.255	0.055	652	0.0 %	652				
2.5	2.500	2.474	-0.026	8482	0.0 %	8482				
5.0	5.000	4.810	-0.190	16729	0.0 %	16729				
7.5	7.500	7.607	0.107	26603	0.0 %	26603				
10.0	10.000	10.015	0.015	35100	0.0 %	35100				

LB92448 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Jan 2018 11:00:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-110	-	0.0000	7470A	PPB	02 Jan 2018 11:29:48	S	Std
0.2 - 1	50.2	652	-	0.2000	7470A	PPB	02 Jan 2018 11:31:57	S	Std
2.5 - 1	52.5	8482	-	2.5000	7470A	PPB	02 Jan 2018 11:34:05	S	Std
5.0 - 1	55	16729	-	5.0000	7470A	PPB	02 Jan 2018 11:36:13	S	Std
7.5 - 1	57.5	26603	-	7.5000	7470A	PPB	02 Jan 2018 11:38:21	S	Std
10.0 - 1	57.5	35100	-	10.0000	7470A	PPB	02 Jan 2018 11:40:29	S	Std
ICV30 - 1	ICV30	13722	3.9581	-7470A	PPB	PPB	02 Jan 2018 11:43:26	U	SMPL
ICB30 - 1	ICB30	-119	0.0367	-7470A	PPB	PPB	02 Jan 2018 11:45:33	U	SMPL
CCV77 - 1	CCV77	16960	4.8754	-7470A	PPB	PPB	02 Jan 2018 11:47:40	U	SMPL
CCB77 - 1	CCB77	-98	0.0427	-7470A	PPB	PPB	02 Jan 2018 11:49:47	U	SMPL
CRA - 1	CRA	635	0.2504	-7470A	PPB	PPB	02 Jan 2018 11:54:15	U	SMPL
HighStd - 1	HighStd	35849	10.2269	-7470A	PPB	PPB	02 Jan 2018 11:56:22	U	SMPL
ChkStd - 1	ChkStd	23735	6.7948	-7470A	PPB	PPB	02 Jan 2018 11:58:28	U	SMPL
PB105409BL - 1	PBW	-150	0.0280	-7470A	PPB	PPB	02 Jan 2018 12:00:38	U	SMPL
PB105409BS - 1	LCSW	13885	4.0042	-7470A	PPB	PPB	02 Jan 2018 12:02:47	U	SMPL
PB105409TB - 1	PB105395TB	-101	0.0418	-7470A	PPB	PPB	02 Jan 2018 12:04:54	U	SMPL
I6686-08 - 1	TR-05-122817	4	0.0716	-7470A	PPB	PPB	02 Jan 2018 12:07:02	U	SMPL
I6686-08DUP - 1	TR-05-122817DUP	36	0.0807	-7470A	PPB	PPB	02 Jan 2018 12:09:11	U	SMPL
I6686-08MS - 1	TR-05-122817MS	13710	3.9547	-7470A	PPB	PPB	02 Jan 2018 12:11:19	U	SMPL
I6686-08MSD - 1	TR-05-122817MSD	12852	3.7116	-7470A	PPB	PPB	02 Jan 2018 12:13:26	U	SMPL
CCV78 - 1	CCV78	16984	4.8822	-7470A	PPB	PPB	02 Jan 2018 12:15:33	U	SMPL
CCB78 - 1	CCB78	-120	0.0365	-7470A	PPB	PPB	02 Jan 2018 12:17:40	U	SMPL
I7091-02 - 1	RO-CHRT25337	-15	0.0662	-7470A	PPB	PPB	02 Jan 2018 12:19:47	U	SMPL
I7092-02 - 1	LAR-1169	284	0.1509	-7470A	PPB	PPB	02 Jan 2018 12:21:54	U	SMPL
I7095-02 - 1	DRUMS-1-4	44	0.0829	-7470A	PPB	PPB	02 Jan 2018 12:24:01	U	SMPL
6686-08LX5 - 1		24	0.0773	-7470A	PPB	PPB	02 Jan 2018 12:27:48	U	SMPL
6686-08A - 1		13816	3.9847	-7470A	PPB	PPB	02 Jan 2018 12:29:55	U	SMPL
CCV79 - 1	CCV79	17138	4.9258	-7470A	PPB	PPB	02 Jan 2018 12:32:03	U	SMPL
CCB79 - 1	CCB79	-126	0.0348	-7470A	PPB	PPB	02 Jan 2018 12:34:11	U	SMPL