



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92449

Temp.: 21.07 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/29/2017	16:20
2	CAL2	1	Water			6.99	12/29/2017	16:22
3	CAL3	1	Water			10.01	12/29/2017	16:24
4	ICV	1	Water			7.00	12/29/2017	16:26
5	CCV1	1	Water			2.01	12/29/2017	16:28
6	I6680-22	1	Solid	20	20	8.22	12/29/2017	16:30
7	I6680-22DUP	1	Solid	20	20	8.24	12/29/2017	16:32
8	I6680-24	1	Solid	20	20	8.30	12/29/2017	16:34
9	I6680-26	1	Solid	20	20	8.33	12/29/2017	16:36
10	I6680-28	1	Solid	20	20	8.61	12/29/2017	16:38
11	I6682-08	1	Solid	20	20	8.71	12/29/2017	16:40
12	CCV2	1	Water			12.01	12/29/2017	16:42