



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92451

Temp.: 21.07 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/02/2018	13:05
2	CAL2	1	Water			6.99	01/02/2018	13:07
3	CAL3	1	Water			10.01	01/02/2018	13:09
4	ICV	1	Water			7.00	01/02/2018	13:11
5	CCV1	1	Water			2.01	01/02/2018	13:13
6	I7105-02	1	Solid	20	20	7.62	01/02/2018	13:15
7	I7105-02DUP	1	Solid	20	20	7.64	01/02/2018	13:17
8	I7105-04	1	Solid	20	20	7.81	01/02/2018	13:19
9	I7105-06	1	Solid	20	20	9.93	01/02/2018	13:21
10	CCV2	1	Water			12.01	01/02/2018	13:23

LB 92451

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I7105 WorkList ID : 107096 Date : 1/2/2018 12:44:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/29/2017	Solid	I7105-02	Corrosivity	Cool 4 deg C	PSEG01	F52	RT4644	12/29/2017	9045C
12/29/2017	Solid	I7105-04	Corrosivity	Cool 4 deg C	PSEG01	F52	RT4651	12/29/2017	9045C
12/29/2017	Solid	I7105-06	Corrosivity	Cool 4 deg C	PSEG01	F52	RT1500	12/29/2017	9045C

Date/Time 1/2/18 12:55 PM

Received by: SP

Relinquished by: 

Date/Time 1/2/18 2:00 PM

Received by: 

Relinquished by: SP