


7471B

LB92454
INSTRUMENT ID: CV2

Linear

 μ Abs.:

45043



A= 0.0000e+000

B= 2.2162e-004

C= 3.1515e-002

Rho= 0.9999078

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	41	0.000	41				
0.2	0.200	0.233	0.033	911	0.0 %	911	0			
2.5	2.500	2.438	-0.062	10860	0.0 %	10860				
5.0	5.000	4.924	-0.076	22076	0.0 %	22076				
7.5	7.500	7.550	0.050	33926	0.0 %	33926				
10.0	10.000	10.014	0.014	45043	0.0 %	45043				

LB92454 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jan 2018 14:43:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	02 Jan 2018 14:47:38	S	Std
0.2 - 1	50.2	911	-	0.2000	7471B	PPB	02 Jan 2018 14:49:56	S	Std
2.5 - 1	52.5	10860	-	2.5000	7471B	PPB	02 Jan 2018 14:52:13	S	Std
5.0 - 1	55	22076	-	5.0000	7471B	PPB	02 Jan 2018 14:54:31	S	Std
7.5 - 1	57.5	33926	-	7.5000	7471B	PPB	02 Jan 2018 14:56:48	S	Std
10.0 - 1	510	45043	-	10.0000	7471B	PPB	02 Jan 2018 15:01:40	S	Std
ICV31 - 1	ICV31	16947	3.7872	-	7471B	PPB	02 Jan 2018 15:04:19	U	SMPL
ICB31 - 1	ICB31	-139	0.0007	-	7471B	PPB	02 Jan 2018 15:06:35	U	SMPL
CCV80 - 1	CCV80	22233	4.9587	-	7471B	PPB	02 Jan 2018 15:08:51	U	SMPL
CCB80 - 1	CCB80	-114	0.0063	-	7471B	PPB	02 Jan 2018 15:11:07	U	SMPL
CRA - 1	CRA	895	0.2299	-	7471B	PPB	02 Jan 2018 15:13:23	U	SMPL
HighStd - 1	HighStd	45546	10.1252	-	7471B	PPB	02 Jan 2018 15:15:39	U	SMPL
ChkStd - 1	ChkStd	30798	6.8568	-	7471B	PPB	02 Jan 2018 15:17:55	U	SMPL
PB105417BL - 1	PBS	-155	-0.0028	-	7471B	PPB	02 Jan 2018 15:20:13	U	SMPL
PB105417BS - 1	LCSS	15660	3.5020	-	7471B	PPB	02 Jan 2018 15:22:32	U	SMPL
I6680-21 - 1	CL-01-122917-A	2274	0.5355	-	7471B	PPB	02 Jan 2018 15:24:48	U	SMPL
I6680-23 - 1	CL-01-122917-B	1844	0.4402	-	7471B	PPB	02 Jan 2018 15:27:06	U	SMPL
I6680-25 - 1	CL-01-122917-C	2059	0.4878	-	7471B	PPB	02 Jan 2018 15:29:23	U	SMPL
I6680-27 - 1	CL-01-122917-D	1663	0.4001	-	7471B	PPB	02 Jan 2018 15:31:41	U	SMPL
I6682-07 - 1	NB-08-122917	1932	0.4597	-	7471B	PPB	02 Jan 2018 15:33:56	U	SMPL
CCV81 - 1	CCV81	21649	4.8293	-	7471B	PPB	02 Jan 2018 15:36:12	U	SMPL
CCB81 - 1	CCB81	-172	-0.0066	-	7471B	PPB	02 Jan 2018 15:38:28	U	SMPL
I7102-01 - 1	BISP-SB-01	1265	0.3119	-	7471B	PPB	02 Jan 2018 15:40:45	U	SMPL
I7102-02 - 1	BISP-SB-02	1262	0.3112	-	7471B	PPB	02 Jan 2018 15:43:01	U	SMPL
I7102-02DUP - 1	BISP-SB-02DUP	1171	0.2910	-	7471B	PPB	02 Jan 2018 15:45:16	U	SMPL
I7102-03 - 1	I7102-02MS	17660	3.9453	-	7471B	PPB	02 Jan 2018 15:47:33	U	SMPL
I7102-04 - 1	I7102-02MSD	16993	3.7974	-	7471B	PPB	02 Jan 2018 15:49:49	U	SMPL
I7102-05 - 1	BISP-SB-03	164	0.0679	-	7471B	PPB	02 Jan 2018 15:52:06	U	SMPL
I7102-06 - 1	BISP-SB-04	858	0.2217	-	7471B	PPB	02 Jan 2018 15:54:23	U	SMPL
I7102-07 - 1	BISP-SB-05	912	0.2336	-	7471B	PPB	02 Jan 2018 15:56:41	U	SMPL
I7102-02LX5 - 1		221	0.0805	-	7471B	PPB	02 Jan 2018 15:58:59	U	SMPL
I7102-02A - 1		17792	3.9745	-	7471B	PPB	02 Jan 2018 16:01:14	U	SMPL
CCV82 - 1	CCV82	21896	4.8840	-	7471B	PPB	02 Jan 2018 16:03:31	U	SMPL
CCB82 - 1	CCB82	-134	0.0018	-	7471B	PPB	02 Jan 2018 16:05:47	U	SMPL