

7470A


LB92464

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34571



A= 0.0000e+000

B= 2.8994e-004

C= 5.6729e-003

Rho= 0.9999864

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	19	0.000	19				
0.2	0.200	0.200	0.000	671	0.0 %	671				
2.5	2.500	2.502	0.002	8611	0.0 %	8611				
5.0	5.000	4.991	-0.009	17195	0.0 %	17195				
7.5	7.500	7.466	-0.034	25730	0.0 %	25730				
10.0	10.000	10.029	0.029	34571	0.0 %	34571				

LB92464 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jan 2018 10:38:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	19	-	0.00007470A	PPB		03 Jan 2018 10:51:19	S	Std
0.2 - 1	50.2	671	-	0.20007470A	PPB		03 Jan 2018 10:53:28	S	Std
2.5 - 1	52.5	8611	-	2.50007470A	PPB		03 Jan 2018 10:55:36	S	Std
5.0 - 1	55	17195	-	5.00007470A	PPB		03 Jan 2018 10:57:44	S	Std
7.5 - 1	57.5	25730	-	7.50007470A	PPB		03 Jan 2018 10:59:52	S	Std
10.0 - 1	510	34571	-	10.00007470A	PPB		03 Jan 2018 11:02:00	S	Std
ICV32 - 1	ICV32	13016	3.7795	-7470A	PPB		03 Jan 2018 11:04:49	U	SMPL
ICB32 - 1	ICB32	-97	-0.0225	-7470A	PPB		03 Jan 2018 11:06:55	U	SMPL
CCV83 - 1	CCV83	17447	5.0643	-7470A	PPB		03 Jan 2018 11:09:01	U	SMPL
CCB83 - 1	CCB83	-98	-0.0227	-7470A	PPB		03 Jan 2018 11:11:08	U	SMPL
CRA - 1	CRA	687	0.2049	-7470A	PPB		03 Jan 2018 11:13:15	U	SMPL
HighStd - 1	HighStd	34871	10.1162	-7470A	PPB		03 Jan 2018 11:15:22	U	SMPL
ChkStd - 1	ChkStd	24217	7.0272	-7470A	PPB		03 Jan 2018 11:17:29	U	SMPL
PB105433BL - 1	PBW	-186	-0.0483	-7470A	PPB		03 Jan 2018 11:19:38	U	SMPL
PB105433BS - 1	LCSW	13143	3.8164	-7470A	PPB		03 Jan 2018 11:21:48	U	SMPL
PB105433TB - 1	PB105423TB	-114	-0.0274	-7470A	PPB		03 Jan 2018 11:23:55	U	SMPL
I6680-22 - 1	CL-01-122917-A	-20	-0.0001	-7470A	PPB		03 Jan 2018 11:26:03	U	SMPL
I6680-22DUP - 1	CL-01-122917-ADUP	18	0.0109	-7470A	PPB		03 Jan 2018 11:28:11	U	SMPL
I6680-22MS - 1	CL-01-122917-AMS	13176	3.8259	-7470A	PPB		03 Jan 2018 11:30:20	U	SMPL
I6680-22MSD - 1	CL-01-122917-AMSD	13319	3.8674	-7470A	PPB		03 Jan 2018 11:32:27	U	SMPL
CCV84 - 1	CCV84	17292	5.0193	-7470A	PPB		03 Jan 2018 11:34:34	U	SMPL
CCB84 - 1	CCB84	-113	-0.0271	-7470A	PPB		03 Jan 2018 11:36:40	U	SMPL
I6680-24 - 1	CL-01-122917-B	-22	-0.0007	-7470A	PPB		03 Jan 2018 11:38:48	U	SMPL
I6680-26 - 1	CL-01-122917-C	-10	0.0028	-7470A	PPB		03 Jan 2018 11:40:54	U	SMPL
I6680-28 - 1	CL-01-122917-D	2	0.0063	-7470A	PPB		03 Jan 2018 11:43:01	U	SMPL
I6682-08 - 1	NB-08-122917	-6	0.0039	-7470A	PPB		03 Jan 2018 11:45:08	U	SMPL
I6680-22LX5 - 1		26	0.0132	-7470A	PPB		03 Jan 2018 11:47:16	U	SMPL
I6680-22A - 1		14027	4.0727	-7470A	PPB		03 Jan 2018 11:49:24	U	SMPL
CCV85 - 1	CCV85	17317	5.0266	-7470A	PPB		03 Jan 2018 11:51:31	U	SMPL
CCB85 - 1	CCB85	-112	-0.0268	-7470A	PPB		03 Jan 2018 11:53:40	U	SMPL