

7471B


LB92472

INSTRUMENT ID: CV2

Linear

 μ Abs.:

43850



A= 0.0000e+000

B= 2.2916e-004

C= 1.6350e-002

Rho= 0.9999265

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	40	0.000	40				
0.2	0.200	0.202	0.002	808	0.0 %	808	0			
2.5	2.500	2.497	-0.003	10823	0.0 %	10823				
5.0	5.000	4.996	-0.004	21729	0.0 %	21729				
7.5	7.500	7.416	-0.084	32289	0.0 %	32289				
10.0	10.000	10.065	0.065	43850	0.0 %	43850				

LB92472 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jan 2018 14:27:12

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	40	-	0.00007471B	PPB	03 Jan 2018 14:28:45	S	Std
0.2 - 1	50.2	808	-	0.20007471B	PPB	03 Jan 2018 14:31:03	S	Std
2.5 - 1	52.5	10823	-	2.50007471B	PPB	03 Jan 2018 14:33:20	S	Std
5.0 - 1	55	21729	-	5.00007471B	PPB	03 Jan 2018 14:35:38	S	Std
7.5 - 1	57.5	32289	-	7.50007471B	PPB	03 Jan 2018 14:37:55	S	Std
10.0 - 1	516	43850	-	10.00007471B	PPB	03 Jan 2018 14:42:48	S	Std
ICV33 - 1	ICV33	15862	3.6513	-7471B	PPB	03 Jan 2018 14:45:44	U	SMPL
ICB33 - 1	ICB33	-105	-0.0077	-7471B	PPB	03 Jan 2018 14:48:00	U	SMPL
CCV86 - 1	CCV86	21953	5.0471	-7471B	PPB	03 Jan 2018 14:50:15	U	SMPL
CCB86 - 1	CCB86	-82	-0.0024	-7471B	PPB	03 Jan 2018 14:52:31	U	SMPL
CRA - 1	CRA	781	0.1953	-7471B	PPB	03 Jan 2018 14:54:48	U	SMPL
HighStd - 1	HighStd	44102	10.1227	-7471B	PPB	03 Jan 2018 14:57:04	U	SMPL
ChkStd - 1	ChkStd	29924	6.8737	-7471B	PPB	03 Jan 2018 14:59:20	U	SMPL
PB105434BL - 1	PBS	-151	-0.0183	-7471B	PPB	03 Jan 2018 15:01:38	U	SMPL
PB105434BS - 1	LCSS	18139	4.1731	-7471B	PPB	03 Jan 2018 15:03:55	U	SMPL
I7105-01 - 1	RT4644	2432	0.5737	-7471B	PPB	03 Jan 2018 15:06:13	U	SMPL
I7105-03 - 1	RT4651	10895	2.5130	-7471B	PPB	03 Jan 2018 15:08:30	U	SMPL
I7105-05 - 1	RT1500	3670	0.8574	-7471B	PPB	03 Jan 2018 15:10:47	U	SMPL
I7106-01 - 1	EPE-106-2(25-27)	58	0.0296	-7471B	PPB	03 Jan 2018 15:13:05	U	SMPL
J1003-01 - 1	B-10(0-10)COMP	37521	8.6146	-7471B	PPB	03 Jan 2018 15:15:21	U	SMPL
CCV87 - 1	CCV87	21634	4.9740	-7471B	PPB	03 Jan 2018 15:17:37	U	SMPL
CCB87 - 1	CCB87	-90	-0.0043	-7471B	PPB	03 Jan 2018 15:19:55	U	SMPL
J1019-01 - 1	EO-01-010218-A	36962	8.4865	-7471B	PPB	03 Jan 2018 15:22:11	U	SMPL
J1019-01DUP - 1	EO-01-010218-ADUP	33298	7.6469	-7471B	PPB	03 Jan 2018 15:24:27	U	SMPL
J1019-01MS - 1	EO-01-010218-AMS	53395	12.2523	-7471B	PPB	03 Jan 2018 15:29:28	U	SMPL
J1019-01MSD - 1	EO-01-010218-AMSD	52331	12.0085	-7471B	PPB	03 Jan 2018 15:31:45	U	SMPL
J1019-03 - 1	EO-01-010218-B	2215	0.5239	-7471B	PPB	03 Jan 2018 15:34:04	U	SMPL
J1022-01 - 1	SU-03-010218	1264	0.3060	-7471B	PPB	03 Jan 2018 15:36:23	U	SMPL
J1025-01 - 1	EPE-105-1(40-42)	97	0.0386	-7471B	PPB	03 Jan 2018 15:38:40	U	SMPL
J1025-02 - 1	EPE-105-1(50-52)	144	0.0493	-7471B	PPB	03 Jan 2018 15:40:57	U	SMPL
J1025-03 - 1	EPE-105-1(60-62)	409	0.1101	-7471B	PPB	03 Jan 2018 15:43:15	U	SMPL
J1025-04 - 1	EPE-105-1(70-72)	446	0.1186	-7471B	PPB	03 Jan 2018 15:45:31	U	SMPL
CCV88 - 1	CCV88	21877	5.0297	-7471B	PPB	03 Jan 2018 15:47:48	U	SMPL
CCB88 - 1	CCB88	-94	-0.0052	-7471B	PPB	03 Jan 2018 15:50:04	U	SMPL
J1025-05 - 1	EPE-105-1(80-82)	568	0.1465	-7471B	PPB	03 Jan 2018 15:52:21	U	SMPL
J1025-06 - 1	EPE-106-2(30-32)	218	0.0663	-7471B	PPB	03 Jan 2018 15:54:38	U	SMPL
J1025-07 - 1	EPE-106-2(70-72)	718	0.1809	-7471B	PPB	03 Jan 2018 15:56:54	U	SMPL
J1025-08 - 1	EPE-106-2(75-77)	702	0.1772	-7471B	PPB	03 Jan 2018 15:59:11	U	SMPL
J1025-09 - 1	EPE-106-2(65-67)	385	0.1046	-7471B	PPB	03 Jan 2018 16:01:27	U	SMPL
J1019-01LX5 - 1		7566	1.7502	-7471B	PPB	03 Jan 2018 16:03:44	U	SMPL
J1019-01A - 1		51748	11.8749	-7471B	PPB	03 Jan 2018 16:06:02	U	SMPL
CCV89 - 1	CCV89	21890	5.0327	-7471B	PPB	03 Jan 2018 16:08:19	U	SMPL
CCB89 - 1	CCB89	-130	-0.0134	-7471B	PPB	03 Jan 2018 16:10:39	U	SMPL
J1019-01DLX2 - 1		18963	4.3619	-7471B	PPB	03 Jan 2018 16:12:56	U	SMPL
J1019-01DUPDLX2 - 1		17401	4.0040	-7471B	PPB	03 Jan 2018 16:15:12	U	SMPL
J1019-01MSDLX2 - 1		27722	6.3691	-7471B	PPB	03 Jan 2018 16:17:29	U	SMPL
J1019-01MSDDLX2 - 1		27469	6.3111	-7471B	PPB	03 Jan 2018 16:19:46	U	SMPL
J1019-01LDLX10 - 1		3540	0.8276	-7471B	PPB	03 Jan 2018 16:22:04	U	SMPL
J1019-01ADLX2 - 1		26979	6.1988	-7471B	PPB	03 Jan 2018 16:24:22	U	SMPL
CCV90 - 1	CCV90	21827	5.0182	-7471B	PPB	03 Jan 2018 16:26:38	U	SMPL
CCB90 - 1	CCB90	-97	-0.0059	-7471B	PPB	03 Jan 2018 16:28:57	U	SMPL