



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92480

Temp.: 21.03 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

True Value of CCV4 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/04/2018	09:45
2	CAL2	1	Water			6.99	01/04/2018	09:46
3	CAL3	1	Water			10.02	01/04/2018	09:47
4	ICV	1	Water			7.00	01/04/2018	09:48
5	CCV1	1	Water			2.01	01/04/2018	09:49
6	J1011-02	1	Solid	20	20	9.55	01/04/2018	09:51
7	J1011-02DUP	1	Solid	20	20	9.54	01/04/2018	09:53
8	J1011-04	1	Solid	20	20	8.50	01/04/2018	09:55
9	J1011-06	1	Solid	20	20	8.36	01/04/2018	09:57
10	J1011-08	1	Solid	20	20	6.28	01/04/2018	09:59
11	J1012-02	1	Solid	20	20	9.05	01/04/2018	10:01
12	J1012-04	1	Solid	20	20	9.45	01/04/2018	10:03
13	J1012-06	1	Solid	20	20	8.86	01/04/2018	10:05
14	J1012-08	1	Solid	20	20	8.66	01/04/2018	10:07
15	J1012-10	1	Solid	20	20	9.60	01/04/2018	10:09
16	CCV2	1	Water			12.01	01/04/2018	10:11
17	J1012-12	1	Solid	20	20	8.93	01/04/2018	10:13
18	J1012-14	1	Solid	20	20	9.26	01/04/2018	10:15
19	J1012-16	1	Solid	20	20	8.85	01/04/2018	10:17
20	J1016-02	1	Solid	20	20	8.06	01/04/2018	10:19

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J1027-02	1	Solid	20	20	8.41	01/04/2018	10:19
22	J1027-04	1	Solid	20	20	9.20	01/04/2018	10:21
23	J1027-06	1	Solid	20	20	8.93	01/04/2018	10:23
24	J1032-01	1	Solid	20	20	7.79	01/04/2018	10:25
25	J1032-02	1	Solid	20	20	7.76	01/04/2018	10:27
26	J1032-03	1	Solid	20	20	7.88	01/04/2018	10:29
27	CCV3	1	Water			2.01	01/04/2018	10:31
28	J1032-04	1	Solid	20	20	7.20	01/04/2018	10:33
29	CCV4	1	Water			12.01	01/04/2018	10:35

WORKLIST(Hardcopy Internal Chain)

LB92480

Worklist Name : CORROSIVITY010418

Worklist ID : 107161

Date : 1/4/2018 9:35:09 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/03/2018	Solid	J1011-02	Corrosivity	Cool 4 deg C	PSEG05	F51	HR-06-010318-A	01/03/2018	9045C
01/03/2018	Solid	J1011-04	Corrosivity	Cool 4 deg C	PSEG05	F51	HR-06-010318-B	01/03/2018	9045C
01/03/2018	Solid	J1011-06	Corrosivity	Cool 4 deg C	PSEG05	F51	HR-06-010318-C	01/03/2018	9045C
01/03/2018	Solid	J1011-08	Corrosivity	Cool 4 deg C	PSEG05	J52	HR-05-010318	01/03/2018	9045C
01/03/2018	Solid	J1012-02	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-A	01/03/2018	9045C
01/03/2018	Solid	J1012-04	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-B	01/03/2018	9045C
01/03/2018	Solid	J1012-06	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-C	01/03/2018	9045C
01/03/2018	Solid	J1012-08	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-D	01/03/2018	9045C
01/03/2018	Solid	J1012-10	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-E	01/03/2018	9045C
01/03/2018	Solid	J1012-12	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-F	01/03/2018	9045C
01/03/2018	Solid	J1012-14	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-G	01/03/2018	9045C
01/03/2018	Solid	J1012-16	Corrosivity	Cool 4 deg C	PSEG05	F52	JC-03-010318-H	01/03/2018	9045C
01/03/2018	Solid	J1016-02	Corrosivity	Cool 4 deg C	PSEG05	J52	OK-03-010318	01/03/2018	9045C
01/03/2018	Solid	J1027-02	Corrosivity	Cool 4 deg C	PSEG01	F52	RBR251608	01/03/2018	9045C
01/03/2018	Solid	J1027-04	Corrosivity	Cool 4 deg C	PSEG01	F52	RBR251588	01/03/2018	9045C
01/03/2018	Solid	J1027-06	Corrosivity	Cool 4 deg C	PSEG01	F52	RBR251678	01/03/2018	9045C
01/02/2018	Solid	J1032-01	Corrosivity	Cool 4 deg C	CLEA11	J22	OWS-1	01/02/2018	9045C
01/02/2018	Solid	J1032-02	Corrosivity	Cool 4 deg C	CLEA11	J22	OWS-2	01/02/2018	9045C
01/02/2018	Solid	J1032-03	Corrosivity	Cool 4 deg C	CLEA11	J22	OWS-3	01/02/2018	9045C
01/02/2018	Solid	J1032-04	Corrosivity	Cool 4 deg C	CLEA11	J22	OWS-4	01/02/2018	9045C

Date/Time

01/04/18 9:40 AM

Received by:

SP

Relinquished by:

SP

Date/Time

01/04/18 10:50 AM

Received by:

SP

Relinquished by:

SP