


7471B

LB92489
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43372



A= 0.0000e+000

B= 2.3002e-004

C= -8.3098e-003

Rho= 0.9999803

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	43	0.000	43				
0.2	0.200	0.197	-0.003	892	0.0 %	892	0			
2.5	2.500	2.483	-0.017	10833	0.0 %	10833				
5.0	5.000	5.007	0.007	21803	0.0 %	21803				
7.5	7.500	7.543	0.043	32830	0.0 %	32830				
10.0	10.000	9.968	-0.032	43372	0.0 %	43372				

LB92489 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jan 2018 12:05:02

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.00007471B	PPB		05 Jan 2018 12:13:15	S	Std
0.2 - 1	50.2	892	-	0.20007471B	PPB		05 Jan 2018 12:15:33	S	Std
2.5 - 1	52.5	10833	-	2.50007471B	PPB		05 Jan 2018 12:17:51	S	Std
5.0 - 1	55	21803	-	5.00007471B	PPB		05 Jan 2018 12:20:09	S	Std
7.5 - 1	57.5	32830	-	7.50007471B	PPB		05 Jan 2018 12:22:26	S	Std
10.0 - 1	510	43372	-	10.00007471B	PPB		05 Jan 2018 12:24:44	S	Std
ICV35 - 1	ICV35	17096	3.9241	-7471B	PPB		05 Jan 2018 12:27:29	U	SMPL
ICB35 - 1	ICB35	-93	-0.0297	-7471B	PPB		05 Jan 2018 12:29:44	U	SMPL
CCV01 - 1	CCV01	21613	4.9631	-7471B	PPB		05 Jan 2018 12:34:31	U	SMPL
CCB01 - 1	CCB01	-107	-0.0329	-7471B	PPB		05 Jan 2018 12:36:47	U	SMPL
CRA - 1	CRA	874	0.1927	-7471B	PPB		05 Jan 2018 12:39:03	U	SMPL
HighStd - 1	HighStd	42915	9.8629	-7471B	PPB		05 Jan 2018 12:41:19	U	SMPL
ChkStd - 1	ChkStd	29339	6.7402	-7471B	PPB		05 Jan 2018 12:43:35	U	SMPL
PB105482BL - 1	PBS	-171	-0.0476	-7471B	PPB		05 Jan 2018 12:45:54	U	SMPL
PB105482BS - 1	LCSS	15780	3.6214	-7471B	PPB		05 Jan 2018 12:48:13	U	SMPL
J1011-01 - 1	HR-06-010318-A	10752	2.4649	-7471B	PPB		05 Jan 2018 12:50:30	U	SMPL
J1011-03 - 1	HR-06-010318-B	1431	0.3208	-7471B	PPB		05 Jan 2018 12:52:48	U	SMPL
J1011-05 - 1	HR-06-010318-C	3165	0.7197	-7471B	PPB		05 Jan 2018 12:55:06	U	SMPL
J1011-05DUP - 1	HR-06-010318-CDUP	3040	0.6909	-7471B	PPB		05 Jan 2018 12:57:24	U	SMPL
J1011-05MS - 1	HR-06-010318-CMS	19174	4.4021	-7471B	PPB		05 Jan 2018 12:59:40	U	SMPL
CCV02 - 1	CCV02	19817	4.5500	-7471B	PPB		05 Jan 2018 13:01:57	U	SMPL
CCB02 - 1	CCB02	-139	-0.0403	-7471B	PPB		05 Jan 2018 13:04:13	U	SMPL
J1011-05MSD - 1	HR-06-010318-CMSD	19263	4.4225	-7471B	PPB		05 Jan 2018 13:06:30	U	SMPL
J1011-07 - 1	HR-05-010318	2172	0.4913	-7471B	PPB		05 Jan 2018 13:08:46	U	SMPL
J1012-01 - 1	JC-03-010318-A	4381	0.9994	-7471B	PPB		05 Jan 2018 13:11:03	U	SMPL
J1012-03 - 1	JC-03-010318-B	3052	0.6937	-7471B	PPB		05 Jan 2018 13:13:20	U	SMPL
J1012-05 - 1	JC-03-010318-C	3514	0.8000	-7471B	PPB		05 Jan 2018 13:15:37	U	SMPL
J1012-07 - 1	JC-03-010318-D	2521	0.5716	-7471B	PPB		05 Jan 2018 13:17:54	U	SMPL
J1012-09 - 1	JC-03-010318-E	2810	0.6380	-7471B	PPB		05 Jan 2018 13:20:11	U	SMPL
J1012-11 - 1	JC-03-010318-F	2751	0.6245	-7471B	PPB		05 Jan 2018 13:22:29	U	SMPL
J1012-13 - 1	JC-03-010318-G	3038	0.6905	-7471B	PPB		05 Jan 2018 13:24:47	U	SMPL
J1012-15 - 1	JC-03-010318-H	1028	0.2281	-7471B	PPB		05 Jan 2018 13:27:03	U	SMPL
CCV03 - 1	CCV03	19755	4.5357	-7471B	PPB		05 Jan 2018 13:29:20	U	SMPL
CCB03 - 1	CCB03	-106	-0.0327	-7471B	PPB		05 Jan 2018 13:31:37	U	SMPL
J1016-01 - 1	OK-03-010318	1713	0.3857	-7471B	PPB		05 Jan 2018 13:33:54	U	SMPL
J1027-01 - 1	RBR251608	1310	0.2930	-7471B	PPB		05 Jan 2018 13:36:11	U	SMPL
J1027-03 - 1	RBR251588	5186	1.1846	-7471B	PPB		05 Jan 2018 13:38:28	U	SMPL
J1027-05 - 1	RBR251678	48220	11.0832	-7471B	PPB		05 Jan 2018 13:40:44	U	SMPL
PB105483BL - 1	PBS	-272	-0.0709	-7471B	PPB		05 Jan 2018 13:43:01	U	SMPL
PB105483BS - 1	LCSS	14245	3.2683	-7471B	PPB		05 Jan 2018 13:45:20	U	SMPL
J1033-01 - 1	B-1(5)	17144	3.9351	-7471B	PPB		05 Jan 2018 13:47:37	U	SMPL
J1033-01DUP - 1	B-1(5)DUP	16946	3.8896	-7471B	PPB		05 Jan 2018 13:52:56	U	SMPL
J1033-01MS - 1	B-1(5)MS	35160	8.0792	-7471B	PPB		05 Jan 2018 13:55:14	U	SMPL
J1033-01MSD - 1	B-1(5)MSD	33039	7.5913	-7471B	PPB		05 Jan 2018 13:57:30	U	SMPL
CCV04 - 1	CCV04	20244	4.6482	-7471B	PPB		05 Jan 2018 14:05:36	U	SMPL
CCB04 - 1	CCB04	-114	-0.0345	-7471B	PPB		05 Jan 2018 14:07:53	U	SMPL
J1033-02 - 1	B-1(10)	27326	6.2772	-7471B	PPB		05 Jan 2018 14:10:10	U	SMPL
J1033-03 - 1	B-2(5)	29020	6.6668	-7471B	PPB		05 Jan 2018 14:12:27	U	SMPL
J1033-04 - 1	B-2(10)	20798	4.7756	-7471B	PPB		05 Jan 2018 14:14:45	U	SMPL
J1033-05 - 1	B-3(5)	37132	8.5327	-7471B	PPB		05 Jan 2018 14:17:02	U	SMPL
J1033-06 - 1	B-3(10)	215645	49.5941	-7471B	PPB		05 Jan 2018 14:19:19	U	SMPL
J1011-05LX5 - 1		574	0.1237	-7471B	PPB		05 Jan 2018 14:25:44	U	SMPL
J1011-05A - 1		19846	4.5566	-7471B	PPB		05 Jan 2018 14:28:01	U	SMPL
J1033-01LX5 - 1		3595	0.8186	-7471B	PPB		05 Jan 2018 14:30:19	U	SMPL
J1033-01A - 1		35687	8.2004	-7471B	PPB		05 Jan 2018 14:32:37	U	SMPL
J1027-05DLX2 - 1		24093	5.5335	-7471B	PPB		05 Jan 2018 14:34:53	U	SMPL
CCV05 - 1	CCV05	19825	4.5518	-7471B	PPB		05 Jan 2018 14:37:11	U	SMPL
CCB05 - 1	CCB05	-146	-0.0419	-7471B	PPB		05 Jan 2018 14:39:29	U	SMPL
J1033-06DLX10 - 1		23200	5.3281	-7471B	PPB		05 Jan 2018 14:41:46	U	SMPL
CCV06 - 1	CCV06	19771	4.5394	-7471B	PPB		05 Jan 2018 14:44:03	U	SMPL
CCB06 - 1	CCB06	-123	-0.0366	-7471B	PPB		05 Jan 2018 14:46:19	U	SMPL