


7470A

LB92494
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36324



A= 0.0000e+000

B= 2.7765e-004

C= 1.5654e-003

Rho= 0.9996996

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	54	0.000	54				
0.2	0.200	0.215	0.015	770	0.0 %	770				
2.5	2.500	2.458	-0.042	8847	0.0 %	8847				
5.0	5.000	5.096	0.096	18349	0.0 %	18349				
7.5	7.500	7.327	-0.173	26383	0.0 %	26383				
10.0	10.000	10.087	0.087	36324	0.0 %	36324				

LB92494 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jan 2018 15:15:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	05 Jan 2018 15:17:03	S	Std
0.2 - 1	50.2	770	-	0.2000	7470A	PPB	05 Jan 2018 15:19:12	S	Std
2.5 - 1	52.5	8847	-	2.5000	7470A	PPB	05 Jan 2018 15:21:21	S	Std
5.0 - 1	55	18349	-	5.0000	7470A	PPB	05 Jan 2018 15:23:29	S	Std
7.5 - 1	57.5	26383	-	7.5000	7470A	PPB	05 Jan 2018 15:25:38	S	Std
10.0 - 1	510	36324	-	10.0000	7470A	PPB	05 Jan 2018 15:33:02	S	Std
ICV36 - 1	ICV36	13610	3.7804	-7470A	PPB	05 Jan 2018 15:36:47	U	SMPL	
ICB36 - 1	ICB36	-90	-0.0234	-7470A	PPB	05 Jan 2018 15:38:53	U	SMPL	
CCV07 - 1	CCV07	18200	5.0548	-7470A	PPB	05 Jan 2018 15:41:00	U	SMPL	
CCB07 - 1	CCB07	-124	-0.0329	-7470A	PPB	05 Jan 2018 15:43:07	U	SMPL	
CRA - 1	CRA	730	0.2043	-7470A	PPB	05 Jan 2018 15:45:15	U	SMPL	
HighStd - 1	HighStd	36577	10.1573	-7470A	PPB	05 Jan 2018 15:47:22	U	SMPL	
ChkStd - 1	ChkStd	23773	6.6022	-7470A	PPB	05 Jan 2018 15:49:29	U	SMPL	
PB105487BL - 1	PBW	-178	-0.0479	-7470A	PPB	05 Jan 2018 15:51:38	U	SMPL	
PB105487BS - 1	LCSW	12950	3.5972	-7470A	PPB	05 Jan 2018 15:53:47	U	SMPL	
PB105487TB - 1	PB105467TB	-62	-0.0156	-7470A	PPB	05 Jan 2018 15:55:55	U	SMPL	
J1011-02 - 1	HR-06-010318-A	9	0.0041	-7470A	PPB	05 Jan 2018 15:58:03	U	SMPL	
J1011-02DUP - 1	HR-06-010318-ADUP	0	0.0016	-7470A	PPB	05 Jan 2018 16:00:12	U	SMPL	
J1011-02MS - 1	HR-06-010318-AMS	12820	3.5611	-7470A	PPB	05 Jan 2018 16:02:21	U	SMPL	
J1011-02MSD - 1	HR-06-010318-AMSD	12595	3.4986	-7470A	PPB	05 Jan 2018 16:04:27	U	SMPL	
CCV08 - 1	CCV08	17967	4.9902	-7470A	PPB	05 Jan 2018 16:06:34	U	SMPL	
CCB08 - 1	CCB08	-101	-0.0265	-7470A	PPB	05 Jan 2018 16:08:41	U	SMPL	
J1011-04 - 1	HR-06-010318-B	15	0.0057	-7470A	PPB	05 Jan 2018 16:10:50	U	SMPL	
J1011-06 - 1	HR-06-010318-C	19	0.0068	-7470A	PPB	05 Jan 2018 16:12:56	U	SMPL	
J1011-08 - 1	HR-05-010318	18	0.0066	-7470A	PPB	05 Jan 2018 16:15:03	U	SMPL	
J1012-02 - 1	JC-03-010318-A	6	0.0032	-7470A	PPB	05 Jan 2018 16:17:11	U	SMPL	
J1012-04 - 1	JC-03-010318-B	15	0.0057	-7470A	PPB	05 Jan 2018 16:19:18	U	SMPL	
J1012-06 - 1	JC-03-010318-C	1	0.0018	-7470A	PPB	05 Jan 2018 16:21:26	U	SMPL	
J1012-08 - 1	JC-03-010318-D	-18	-0.0034	-7470A	PPB	05 Jan 2018 16:23:35	U	SMPL	
J1012-10 - 1	JC-03-010318-E	6	0.0032	-7470A	PPB	05 Jan 2018 16:25:43	U	SMPL	
J1012-12 - 1	JC-03-010318-F	13	0.0052	-7470A	PPB	05 Jan 2018 16:27:52	U	SMPL	
J1012-14 - 1	JC-03-010318-G	15	0.0057	-7470A	PPB	05 Jan 2018 16:29:59	U	SMPL	
CCV09 - 1	CCV09	17634	4.8977	-7470A	PPB	05 Jan 2018 16:32:07	U	SMPL	
CCB09 - 1	CCB09	-89	-0.0231	-7470A	PPB	05 Jan 2018 16:34:14	U	SMPL	
J1012-16 - 1	JC-03-010318-H	-45	-0.0109	-7470A	PPB	05 Jan 2018 16:36:23	U	SMPL	
J1016-02 - 1	OK-03-010318	-20	-0.0040	-7470A	PPB	05 Jan 2018 16:38:30	U	SMPL	
J1017-02 - 1	4463	31	0.0102	-7470A	PPB	05 Jan 2018 16:40:38	U	SMPL	
J1017-09 - 1	4462	34	0.0110	-7470A	PPB	05 Jan 2018 16:42:45	U	SMPL	
J1017-10 - 1	4422	46	0.0143	-7470A	PPB	05 Jan 2018 16:44:53	U	SMPL	
J1027-02 - 1	RBR251608	-8	-0.0007	-7470A	PPB	05 Jan 2018 16:47:01	U	SMPL	
J1027-04 - 1	RBR251588	12	0.0049	-7470A	PPB	05 Jan 2018 16:49:09	U	SMPL	
J1027-06 - 1	RBR251678	19	0.0068	-7470A	PPB	05 Jan 2018 16:51:17	U	SMPL	
J1032-01 - 1	OVS-1	-32	-0.0073	-7470A	PPB	05 Jan 2018 16:53:26	U	SMPL	
J1032-02 - 1	OVS-2	-13	-0.0020	-7470A	PPB	05 Jan 2018 16:55:34	U	SMPL	
CCV10 - 1	CCV10	17781	4.9385	-7470A	PPB	05 Jan 2018 16:57:41	U	SMPL	
CCB10 - 1	CCB10	-87	-0.0226	-7470A	PPB	05 Jan 2018 16:59:49	U	SMPL	
J1032-03 - 1	OVS-3	-14	-0.0023	-7470A	PPB	05 Jan 2018 17:01:57	U	SMPL	
J1032-04 - 1	OVS-4	0	0.0016	-7470A	PPB	05 Jan 2018 17:04:05	U	SMPL	
J1011-02LX5 - 1		-4	0.0005	-7470A	PPB	05 Jan 2018 17:06:12	U	SMPL	
J1011-02A - 1		13344	3.7066	-7470A	PPB	05 Jan 2018 17:08:20	U	SMPL	
CCV11 - 1	CCV11	17993	4.9974	-7470A	PPB	05 Jan 2018 17:10:27	U	SMPL	
CCB11 - 1	CCB11	-126	-0.0334	-7470A	PPB	05 Jan 2018 17:12:35	U	SMPL	