



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92495

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	01/05/2018	14:45
2	J1017-02		1	>100.0	>212.0	01/05/2018	15:10
3	J1017-02DUP		1	>100.0	>212.0	01/05/2018	15:35

Result = (Celsius * 1.8) + 32

LB92495

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP1518

WorkList ID : 107191

Date : 1/5/2018 2:25:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/10/2018	Solid	J1017-02	Flash Point	Cool 4 deg C	PSEG02	J52	4463	01/03/2018	1010A

Date/Time 1/5/18 2:30 PM
Received by: CP
Relinquished by: KS

Date/Time 1/5/18 15:40
Received by: CP
Relinquished by: KS