



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92505

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 97.0

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/08/2018	09:15
2	CAL2	1	Water			6.99	01/08/2018	09:17
3	CAL3	1	Water			10.01	01/08/2018	09:19
4	ICV	1	Water			7.00	01/08/2018	09:21
5	CCV1	1	Water			2.01	01/08/2018	09:23
6	J1043-17	1	Solid	20	20	8.69	01/08/2018	09:25
7	J1043-17DUP	1	Solid	20	20	8.68	01/08/2018	09:27
8	J1043-18	1	Solid	20	20	7.48	01/08/2018	09:29
9	J1043-19	1	Solid	20	20	8.68	01/08/2018	09:31
10	J1043-20	1	Solid	20	20	4.83	01/08/2018	09:33
11	J1043-21	1	Solid	20	20	6.05	01/08/2018	09:35
12	J1043-22	1	Solid	20	20	7.90	01/08/2018	09:37
13	J1043-23	1	Solid	20	20	6.05	01/08/2018	09:39
14	J1043-24	1	Solid	20	20	6.01	01/08/2018	09:41
15	CCV2	1	Water			12.01	01/08/2018	09:43

WORKLIST(Hardcopy Internal Chain)

LB 92505

Worklist Name : corrosivity010818

Worklist ID : 107202

Date : 1/8/2018 8:46:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/02/2018	Solid	J1043-17	Corrosivity	Cool 4 deg C	LIR001	J11	SB-10-COMP	01/02/2018	9045C
01/02/2018	Solid	J1043-18	Corrosivity	Cool 4 deg C	LIR001	J11	SB-20-COMP	01/02/2018	9045C
01/03/2018	Solid	J1043-19	Corrosivity	Cool 4 deg C	LIR001	J11	SB-09-COMP	01/03/2018	9045C
01/03/2018	Solid	J1043-20	Corrosivity	Cool 4 deg C	LIR001	J11	SB-05-COMP	01/03/2018	9045C
01/03/2018	Solid	J1043-21	Corrosivity	Cool 4 deg C	LIR001	J11	SB-04-COMP	01/03/2018	9045C
01/03/2018	Solid	J1043-22	Corrosivity	Cool 4 deg C	LIR001	J11	SB-01-COMP	01/03/2018	9045C
01/03/2018	Solid	J1043-23	Corrosivity	Cool 4 deg C	LIR001	J11	SB-02-COMP	01/03/2018	9045C
01/03/2018	Solid	J1043-24	Corrosivity	Cool 4 deg C	LIR001	J11	SB-03-COMP	01/03/2018	9045C

Date/Time

01/08/18 8:55 AM

Received by:

SP

Relinquished by:

CP

Date/Time

01/08/18 10:30 AM

Received by:

CP

Relinquished by:

SP