


245.1

LB92511
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33995



A= 0.0000e+000

B= 2.9235e-004

C= -1.8362e-002

Rho= 0.9998671

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	27	0.000	27				
0.05	0.050	0.041	-0.009	204	0.0 %	204				
0.2	0.200	0.172	-0.028	650	0.0 %	650				
2.5	2.500	2.471	-0.029	8514	0.0 %	8514				
5.0	5.000	5.130	0.130	17611	0.0 %	17611				
7.5	7.500	7.525	0.025	25803	0.0 %	25803				
10.0	10.000	9.921	-0.079	33999	0.0 %	33999				

LB92511 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Jan 2018 13:33:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	245.1	PPB	08 Jan 2018 14:20:47	S	Std
0.05 - 1	50.05	204	-	0.0500	245.1	PPB	08 Jan 2018 14:22:56	S	Std
0.2 - 1	50.2	650	-	0.2000	245.1	PPB	08 Jan 2018 14:25:05	S	Std
2.5 - 1	52.5	8514	-	2.5000	245.1	PPB	08 Jan 2018 14:27:14	S	Std
5.0 - 1	55	17611	-	5.0000	245.1	PPB	08 Jan 2018 14:29:22	S	Std
7.5 - 1	57.5	25803	-	7.5000	245.1	PPB	08 Jan 2018 14:31:30	S	Std
10.0 - 1	510	33999	-	10.0000	245.1	PPB	08 Jan 2018 14:33:38	S	Std
ICV37 - 1	ICV37	13213	3.8445	-245.1		PPB	08 Jan 2018 14:37:02	U	SMPL
ICB37 - 1	ICB37	-56	-0.0347	-245.1		PPB	08 Jan 2018 14:39:08	U	SMPL
CCV12 - 1	CCV12	17625	5.1344	-245.1		PPB	08 Jan 2018 14:41:15	U	SMPL
CCB12 - 1	CCB12	-106	-0.0494	-245.1		PPB	08 Jan 2018 14:43:21	U	SMPL
CRA - 1	CRA	630	0.1658	-245.1		PPB	08 Jan 2018 14:45:29	U	SMPL
HighStd - 1	HighStd	34499	10.0675	-245.1		PPB	08 Jan 2018 14:47:36	U	SMPL
ChkStd - 1	ChkStd	23959	6.9861	-245.1		PPB	08 Jan 2018 14:49:43	U	SMPL
PB105500BL - 1	PBW	15	-0.0140	-245.1		PPB	08 Jan 2018 14:54:45	U	SMPL
PB105500BS - 1	LCSW	13653	3.9731	-245.1		PPB	08 Jan 2018 14:56:52	U	SMPL
J1029-01 - 1	COMP	-63	-0.0368	-245.1		PPB	08 Jan 2018 14:59:00	U	SMPL
J1029-01DUP - 1	COMPDUP	3	-0.0175	-245.1		PPB	08 Jan 2018 15:01:09	U	SMPL
J1029-01MS - 1	COMPMS	13500	3.9284	-245.1		PPB	08 Jan 2018 15:03:17	U	SMPL
J1029-01MSD - 1	COMPMSD	13926	4.0529	-245.1		PPB	08 Jan 2018 15:05:26	U	SMPL
CCV13 - 1	CCV13	17679	5.1501	-245.1		PPB	08 Jan 2018 15:07:33	U	SMPL
CCB13 - 1	CCB13	-118	-0.0529	-245.1		PPB	08 Jan 2018 15:09:40	U	SMPL
J1029-01LX5 - 1		7	-0.0163	-245.1		PPB	08 Jan 2018 15:11:48	U	SMPL
J1029-01A - 1		13767	4.0065	-245.1		PPB	08 Jan 2018 15:13:55	U	SMPL
CCV14 - 1	CCV14	17838	5.1966	-245.1		PPB	08 Jan 2018 15:16:02	U	SMPL
CCB14 - 1	CCB14	-91	-0.0450	-245.1		PPB	08 Jan 2018 15:18:09	U	SMPL