



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92512

Temp.: 22.4 °C

SLOPE.: 97.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/08/2018	12:55
2	CAL2	1	Water			6.99	01/08/2018	12:57
3	CAL3	1	Water			10.01	01/08/2018	12:59
4	ICV	1	Water			7.00	01/08/2018	13:01
5	CCV1	1	Water			2.01	01/08/2018	13:03
6	J1056-01	1	Water			4.33	01/08/2018	13:05
7	J1056-01DUP	1	Water			4.34	01/08/2018	13:07
8	CCV2	1	Water			12.01	01/08/2018	13:10

WORKLIST(Hardcopy Internal Chain)

LB92512

WorkList Name : Ph010818

WorkList ID : 107221

Date : 1/8/2018 12:32:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/03/2018	Water	J1056-01	pH	Cool 4 deg C	STER11	J22	POOLWATER	01/03/2018	SM 4500-PH B

Date/Time 01/08/18 12:40 PM

Received by: SP

Relinquished by:

Date/Time 01/08/18 12:30 PM

Received by: SP

Relinquished by: