

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 01/09/2018

DATE & TIME IN: 01/09/2018 09:05

OVEN TEMP: 103 °C

Run Number: LB92521

DATE & TIME OUT: 01/09/2018 13:30

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB92521BL	LB92521BL	1.8486	1.8486	100	1.8486	1.8486	1.8486	0	0
2	LB92521BS	LB92521BS	1.8395	1.8395	100	1.8915	1.8915	1.8915	0.052	520
3	J1054-03	FINAL-EFFLUENT-19C	1.8398	1.8398	100	1.8528	1.8528	1.8528	0.013	130
4	J1054-03DUP	FINAL-EFFLUENT-19CDUP	1.8362	1.8362	100	1.849	1.849	1.849	0.0128	128

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS010918

WorkList ID : 107228

Date : 1/9/2018 8:06:26 AM

2092521

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/14/2018	Water	J1054-03	TSS	Cool 4 deg C	M&MM01	J22	FINAL-EFFUENT-19C	01/08/2018	SM2540 D

Date/Time 01/09/18, 8:55
 Received by: AS
 Relinquished by: Cp

Date/Time 01/09/18, 9:15
 Received by: Cp
 Relinquished by: AS