


7470A

LB92524
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36191



A= 0.0000e+000

B= 2.7813e-004

C= 1.0997e-002

Rho= 0.9998939

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	16	0.000	16				
0.2	0.200	0.207	0.007	706	0.0 %	706				
2.5	2.500	2.510	0.010	8986	0.0 %	8986				
5.0	5.000	4.995	-0.005	17921	0.0 %	17921				
7.5	7.500	7.395	-0.105	26549	0.0 %	26549				
10.0	10.000	10.077	0.077	36191	0.0 %	36191				

LB92524 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Jan 2018 11:05:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7470A	PPB	09 Jan 2018 11:10:42	S	Std
0.2 - 1	50.2	706	-	0.2000	7470A	PPB	09 Jan 2018 11:12:51	S	Std
2.5 - 1	52.5	8986	-	2.5000	7470A	PPB	09 Jan 2018 11:14:59	S	Std
5.0 - 1	55	17921	-	5.0000	7470A	PPB	09 Jan 2018 11:17:07	S	Std
7.5 - 1	57.5	26549	-	7.5000	7470A	PPB	09 Jan 2018 11:19:14	S	Std
10.0 - 1	59.0	36191	-	10.0000	7470A	PPB	09 Jan 2018 11:23:52	S	Std
ICV38 - 1	ICV38	13532	3.7746	-	7470A	PPB	09 Jan 2018 11:26:23	U	SMPL
ICB38 - 1	ICB38	-89	-0.0138	-	7470A	PPB	09 Jan 2018 11:28:29	U	SMPL
CCV15 - 1	CCV15	18277	5.0943	-	7470A	PPB	09 Jan 2018 11:30:37	U	SMPL
CCB15 - 1	CCB15	-103	-0.0176	-	7470A	PPB	09 Jan 2018 11:32:43	U	SMPL
CRA - 1	CRA	714	0.2096	-	7470A	PPB	09 Jan 2018 11:34:51	U	SMPL
HighStd - 1	HighStd	35945	10.0083	-	7470A	PPB	09 Jan 2018 11:36:57	U	SMPL
ChkStd - 1	ChkStd	25473	7.0957	-	7470A	PPB	09 Jan 2018 11:39:04	U	SMPL
PB105514BL - 1	PBW	-165	-0.0349	-	7470A	PPB	09 Jan 2018 11:41:13	U	SMPL
PB105514BS - 1	LCSW	12985	3.6225	-	7470A	PPB	09 Jan 2018 11:43:22	U	SMPL
PB105514TB - 1	PB105505TB	-49	-0.0026	-	7470A	PPB	09 Jan 2018 11:45:29	U	SMPL
J1043-17 - 1	SB-10-COMP	14	0.0149	-	7470A	PPB	09 Jan 2018 11:47:37	U	SMPL
J1043-17DUP - 1	SB-10-COMPDUP	48	0.0243	-	7470A	PPB	09 Jan 2018 11:49:45	U	SMPL
J1043-17MS - 1	SB-10-COMPMS	13455	3.7532	-	7470A	PPB	09 Jan 2018 11:51:54	U	SMPL
J1043-17MSD - 1	SB-10-COMPMSD	13463	3.7554	-	7470A	PPB	09 Jan 2018 11:54:01	U	SMPL
CCV16 - 1	CCV16	17909	4.9920	-	7470A	PPB	09 Jan 2018 11:56:07	U	SMPL
CCB16 - 1	CCB16	-111	-0.0199	-	7470A	PPB	09 Jan 2018 11:58:14	U	SMPL
J1043-18 - 1	SB-20-COMP	3	0.0118	-	7470A	PPB	09 Jan 2018 12:00:22	U	SMPL
J1043-19 - 1	SB-09-COMP	48	0.0243	-	7470A	PPB	09 Jan 2018 12:02:29	U	SMPL
J1043-20 - 1	SB-05-COMP	40	0.0221	-	7470A	PPB	09 Jan 2018 12:04:36	U	SMPL
J1043-21 - 1	SB-04-COMP	6	0.0127	-	7470A	PPB	09 Jan 2018 12:06:43	U	SMPL
J1043-22 - 1	SB-01-COMP	5	0.0124	-	7470A	PPB	09 Jan 2018 12:08:50	U	SMPL
J1043-23 - 1	SB-02-COMP	14	0.0149	-	7470A	PPB	09 Jan 2018 12:10:58	U	SMPL
J1043-24 - 1	SB-03-COMP	47	0.0241	-	7470A	PPB	09 Jan 2018 12:13:06	U	SMPL
PB105515BL - 1	PBW	47	0.0241	-	7470A	PPB	09 Jan 2018 12:15:14	U	SMPL
PB105515BS - 1	LCSW	11991	3.3460	-	7470A	PPB	09 Jan 2018 12:17:22	U	SMPL
J1049-07 - 1	MW-A	806	0.2352	-	7470A	PPB	09 Jan 2018 12:19:29	U	SMPL
CCV17 - 1	CCV17	18285	5.0965	-	7470A	PPB	09 Jan 2018 12:21:37	U	SMPL
CCB17 - 1	CCB17	-81	-0.0115	-	7470A	PPB	09 Jan 2018 12:23:43	U	SMPL
J1049-08 - 1	MW-B	386	0.1184	-	7470A	PPB	09 Jan 2018 12:25:51	U	SMPL
J1049-09 - 1	MW-C	625	0.1848	-	7470A	PPB	09 Jan 2018 12:27:58	U	SMPL
J1049-10 - 1	MW-E	62	0.0282	-	7470A	PPB	09 Jan 2018 12:30:06	U	SMPL
J1049-11 - 1	MW-F	3356	0.9444	-	7470A	PPB	09 Jan 2018 12:32:12	U	SMPL
J1049-11DUP - 1	MW-FDUP	3488	0.9811	-	7470A	PPB	09 Jan 2018 12:34:20	U	SMPL
J1049-11MS - 1	MW-FMS	18649	5.1978	-	7470A	PPB	09 Jan 2018 12:36:27	U	SMPL
J1049-11MSD - 1	MW-FMSD	17397	4.8496	-	7470A	PPB	09 Jan 2018 12:38:35	U	SMPL
CCV18 - 1	CCV18	18214	5.0768	-	7470A	PPB	09 Jan 2018 12:40:44	U	SMPL
CCB18 - 1	CCB18	-104	-0.0179	-	7470A	PPB	09 Jan 2018 12:42:53	U	SMPL
J1043-17LX5 - 1		-67	-0.0076	-	7470A	PPB	09 Jan 2018 12:45:01	U	SMPL
J1043-17A - 1		13714	3.8252	-	7470A	PPB	09 Jan 2018 12:47:07	U	SMPL
J1049-11LX5 - 1		504	0.1512	-	7470A	PPB	09 Jan 2018 12:49:14	U	SMPL
J1049-11A - 1		18453	5.1433	-	7470A	PPB	09 Jan 2018 12:51:22	U	SMPL
CCV19 - 1	CCV19	18091	5.0426	-	7470A	PPB	09 Jan 2018 12:53:29	U	SMPL
CCB19 - 1	CCB19	-110	-0.0196	-	7470A	PPB	09 Jan 2018 12:55:36	U	SMPL