

7471B


LB92531

INSTRUMENT ID: CN2

Linear

 μ Abs.:

41872



A= 0.0000e+000

B= 2.3865e-004

C= -2.1636e-002

Rho= 0.9999271

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	37	0.000	37				
0.2	0.200	0.190	-0.010	887	0.0 %	887	0			
2.5	2.500	2.473	-0.027	10451	0.0 %	10451				
5.0	5.000	5.098	0.098	21454	0.0 %	21454				
7.5	7.500	7.481	-0.019	31436	0.0 %	31436				
10.0	10.000	9.971	-0.029	41872	0.0 %	41872				

LB92531 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jan 2018 14:33:49

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	37	-	0.00007471B	PPB	09 Jan 2018 14:36:23	S	Std
0.2 - 1	50.2	887	-	0.20007471B	PPB	09 Jan 2018 14:38:41	S	Std
2.5 - 1	52.5	10451	-	2.50007471B	PPB	09 Jan 2018 14:40:59	S	Std
5.0 - 1	55	21454	-	5.00007471B	PPB	09 Jan 2018 14:43:16	S	Std
7.5 - 1	57.5	31436	-	7.50007471B	PPB	09 Jan 2018 14:45:33	S	Std
10.0 - 1	510	41872	-	10.00007471B	PPB	09 Jan 2018 14:50:27	S	Std
ICV39 - 1	ICV39	16494	3.9147	-7471B	PPB	09 Jan 2018 14:53:07	U	SMPL
ICB39 - 1	ICB39	-99	-0.0453	-7471B	PPB	09 Jan 2018 14:55:22	U	SMPL
CCV20 - 1	CCV20	21488	5.1065	-7471B	PPB	09 Jan 2018 14:57:38	U	SMPL
CCB20 - 1	CCB20	-63	-0.0367	-7471B	PPB	09 Jan 2018 14:59:54	U	SMPL
CRA - 1	CRA	854	0.1822	-7471B	PPB	09 Jan 2018 15:02:10	U	SMPL
HighStd - 1	HighStd	42166	10.0414	-7471B	PPB	09 Jan 2018 15:04:26	U	SMPL
ChkStd - 1	ChkStd	28980	6.8945	-7471B	PPB	09 Jan 2018 15:06:42	U	SMPL
PB105522BL - 1	PBS	-92	-0.0436	-7471B	PPB	09 Jan 2018 15:08:59	U	SMPL
PB105522BS - 1	LCSS	15787	3.7460	-7471B	PPB	09 Jan 2018 15:11:18	U	SMPL
J1018-01 - 1	OR-03-010818	1949	0.4435	-7471B	PPB	09 Jan 2018 15:13:35	U	SMPL
J1022-04 - 1	SU-04-010818	790	0.1669	-7471B	PPB	09 Jan 2018 15:15:52	U	SMPL
J1023-01 - 1	TR-04-010817	676	0.1397	-7471B	PPB	09 Jan 2018 15:18:09	U	SMPL
J1044-04 - 1	SP-COMP-A	4465	1.0439	-7471B	PPB	09 Jan 2018 15:20:27	U	SMPL
J1044-06 - 1	SP-COMP-B	2750	0.6347	-7471B	PPB	09 Jan 2018 15:22:42	U	SMPL
CCV21 - 1	CCV21	21489	5.1068	-7471B	PPB	09 Jan 2018 15:24:58	U	SMPL
CCB21 - 1	CCB21	-84	-0.0417	-7471B	PPB	09 Jan 2018 15:27:13	U	SMPL
J1044-08 - 1	SP-COMP-C	10365	2.4520	-7471B	PPB	09 Jan 2018 15:29:29	U	SMPL
J1044-10 - 1	SP-COMP-D	2838	0.6557	-7471B	PPB	09 Jan 2018 15:31:45	U	SMPL
J1050-01 - 1	EPE-105-1(90-92)	898	0.1927	-7471B	PPB	09 Jan 2018 15:34:01	U	SMPL
J1050-01DUP - 1	EPE-105-1(90-92)DUP	934	0.2013	-7471B	PPB	09 Jan 2018 15:36:17	U	SMPL
J1050-01MS - 1	EPE-105-1(90-92)MS	17644	4.1891	-7471B	PPB	09 Jan 2018 15:38:34	U	SMPL
J1050-01MSD - 1	EPE-105-1(90-92)MSD	17023	4.0409	-7471B	PPB	09 Jan 2018 15:40:51	U	SMPL
J1050-02 - 1	EPE-105-1(105-107)	61	-0.0071	-7471B	PPB	09 Jan 2018 15:43:08	U	SMPL
J1050-03 - 1	EPE-106-2(85-87)	244	0.0366	-7471B	PPB	09 Jan 2018 15:45:25	U	SMPL
J1050-01LX5 - 1		189	0.0235	-7471B	PPB	09 Jan 2018 15:49:27	U	SMPL
J1050-01A - 1		18013	4.2772	-7471B	PPB	09 Jan 2018 15:51:42	U	SMPL
CCV22 - 1	CCV22	21545	5.1201	-7471B	PPB	09 Jan 2018 15:53:58	U	SMPL
CCB22 - 1	CCB22	-108	-0.0474	-7471B	PPB	09 Jan 2018 15:56:15	U	SMPL