



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92541

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/09/2018	17:20
2	CAL2	1	Water			6.99	01/09/2018	17:22
3	CAL3	1	Water			10.01	01/09/2018	17:24
4	ICV	1	Water			7.00	01/09/2018	17:26
5	CCV1	1	Water			2.01	01/09/2018	17:28
6	J1013-02	1	Solid	20	20	8.63	01/09/2018	17:30
7	J1013-02DUP	1	Solid	20	20	8.62	01/09/2018	17:32
8	J1064-15	1	Solid	20	20	8.77	01/09/2018	17:34
9	J1067-03	1	Solid	20	20	8.10	01/09/2018	17:36
10	CCV2	1	Water			12.01	01/09/2018	17:37

WORKLIST(Hardcopy Internal Chain)

LB92541

WorkList Name : corrosivity01918

WorkList ID : 107264

Date : 1/9/2018 4:48:38 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/09/2018	Solid	J1013-02	Corrosivity	Cool 4 deg C	PSEG05	J52	NB-08-010918	01/09/2018	9045C
01/09/2018	Solid	J1064-15	Corrosivity	Cool 4 deg C	PSEG02	J52	OILY-STONES	01/09/2018	9045C
01/09/2018	Solid	J1067-03	Corrosivity	Cool 4 deg C	PSEG01	J52	CHRT-27007	01/09/2018	9045C

Date/Time

01/09/18 5:16 PM

Received by:

SP

Relinquished by:

CP

Date/Time

01/09/18 6:00 PM

Received by:

CP

Relinquished by:

SP