


7470A

LB 92545
INSTRUMENT ID: CY1

Linear

 μ Abs.:

35144



A= 0.0000e+000

B= 2.8340e-004

C= 4.1821e-002

Rho= 0.9999315

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	-21	0.000	-21				
0.2	0.200	0.220	0.020	628	0.0 %	628				
2.5	2.500	2.407	-0.093	8345	0.0 %	8345				
5.0	5.000	5.027	0.027	17591	0.0 %	17591				
7.5	7.500	7.509	0.009	26349	0.0 %	26349				
10.0	10.000	10.002	0.002	35144	0.0 %	35144				

LB92545 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jan 2018 12:25:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7470A	PPB	10 Jan 2018 12:42:22	S	Std
0.2 - 1	50.2	628	-	0.2000	7470A	PPB	10 Jan 2018 12:44:30	S	Std
2.5 - 1	52.5	8345	-	2.5000	7470A	PPB	10 Jan 2018 12:46:38	S	Std
5.0 - 1	55	17591	-	5.0000	7470A	PPB	10 Jan 2018 12:48:46	S	Std
7.5 - 1	57.5	26349	-	7.5000	7470A	PPB	10 Jan 2018 12:53:18	S	Std
10.0 - 1	510	35144	-	10.0000	7470A	PPB	10 Jan 2018 12:55:26	S	Std
ICV40 - 1	ICV40	13912	3.9844	-7470A	PPB	PPB	10 Jan 2018 12:59:02	U	SMPL
ICB40 - 1	ICB40	22	0.0481	-7470A	PPB	PPB	10 Jan 2018 13:01:08	U	SMPL
CCV23 - 1	CCV23	18092	5.1690	-7470A	PPB	PPB	10 Jan 2018 13:03:14	U	SMPL
CCB23 - 1	CCB23	40	0.0532	-7470A	PPB	PPB	10 Jan 2018 13:05:21	U	SMPL
CRA - 1	CRA	651	0.2263	-7470A	PPB	PPB	10 Jan 2018 13:07:28	U	SMPL
HighStd - 1	HighStd	35392	10.0718	-7470A	PPB	PPB	10 Jan 2018 13:09:34	U	SMPL
ChkStd - 1	ChkStd	24317	6.9332	-7470A	PPB	PPB	10 Jan 2018 13:11:41	U	SMPL
PB105544BL - 1	PBW	5	0.0432	-7470A	PPB	PPB	10 Jan 2018 13:13:49	U	SMPL
PB105544BS - 1	LCSW	12857	3.6854	-7470A	PPB	PPB	10 Jan 2018 13:15:58	U	SMPL
PB105544TB - 1	PB105524TB	77	0.0636	-7470A	PPB	PPB	10 Jan 2018 13:18:05	U	SMPL
J1018-02 - 1	OR-03-010818	174	0.0911	-7470A	PPB	PPB	10 Jan 2018 13:20:13	U	SMPL
J1018-02DUP - 1	OR-03-010818DUP	185	0.0942	-7470A	PPB	PPB	10 Jan 2018 13:22:21	U	SMPL
J1018-02MS - 1	OR-03-010818MS	13409	3.8419	-7470A	PPB	PPB	10 Jan 2018 13:24:30	U	SMPL
J1018-02MSD - 1	OR-03-010818MSD	13539	3.8787	-7470A	PPB	PPB	10 Jan 2018 13:26:36	U	SMPL
CCV24 - 1	CCV24	17701	5.0582	-7470A	PPB	PPB	10 Jan 2018 13:28:42	U	SMPL
CCB24 - 1	CCB24	13	0.0455	-7470A	PPB	PPB	10 Jan 2018 13:30:49	U	SMPL
J1022-05 - 1	SU-04-010818	150	0.0843	-7470A	PPB	PPB	10 Jan 2018 13:32:56	U	SMPL
J1023-02 - 1	TR-04-010817	167	0.0891	-7470A	PPB	PPB	10 Jan 2018 13:35:03	U	SMPL
J1044-05 - 1	SP-COMP-A	168	0.0894	-7470A	PPB	PPB	10 Jan 2018 13:37:10	U	SMPL
J1044-07 - 1	SP-COMP-B	201	0.0988	-7470A	PPB	PPB	10 Jan 2018 13:39:17	U	SMPL
J1044-09 - 1	SP-COMP-C	177	0.0920	-7470A	PPB	PPB	10 Jan 2018 13:41:25	U	SMPL
J1044-11 - 1	SP-COMP-D	163	0.0880	-7470A	PPB	PPB	10 Jan 2018 13:43:33	U	SMPL
J1057-01 - 1	OWS-1	161	0.0874	-7470A	PPB	PPB	10 Jan 2018 13:45:41	U	SMPL
J1057-02 - 1	OWS-2B	283	0.1220	-7470A	PPB	PPB	10 Jan 2018 13:47:49	U	SMPL
J1057-03 - 1	OWS-2A	192	0.0962	-7470A	PPB	PPB	10 Jan 2018 13:49:58	U	SMPL
J1057-04 - 1	OWS-2C	185	0.0942	-7470A	PPB	PPB	10 Jan 2018 13:52:05	U	SMPL
CCV25 - 1	CCV25	17633	5.0389	-7470A	PPB	PPB	10 Jan 2018 13:54:11	U	SMPL
CCB25 - 1	CCB25	-88	0.0169	-7470A	PPB	PPB	10 Jan 2018 13:56:18	U	SMPL
J1062-11 - 1	SB-26-COMP	39	0.0529	-7470A	PPB	PPB	10 Jan 2018 13:58:26	U	SMPL
J1062-12 - 1	SB-27-COMP	-4	0.0407	-7470A	PPB	PPB	10 Jan 2018 14:00:33	U	SMPL
J1062-13 - 1	SB-28-COMP	19	0.0472	-7470A	PPB	PPB	10 Jan 2018 14:02:40	U	SMPL
J1062-14 - 1	SB-29-COMP	173	0.0908	-7470A	PPB	PPB	10 Jan 2018 14:04:47	U	SMPL
J1062-15 - 1	SB-30-COMP	159	0.0869	-7470A	PPB	PPB	10 Jan 2018 14:06:54	U	SMPL
J1018-02LX5 - 1		36	0.0520	-7470A	PPB	PPB	10 Jan 2018 14:09:01	U	SMPL
J1018-02A - 1		14253	4.0811	-7470A	PPB	PPB	10 Jan 2018 14:11:09	U	SMPL
CCV26 - 1	CCV26	17260	4.9332	-7470A	PPB	PPB	10 Jan 2018 14:13:17	U	SMPL
CCB26 - 1	CCB26	-10	0.0390	-7470A	PPB	PPB	10 Jan 2018 14:15:25	U	SMPL