



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92558

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.0

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/11/2018	13:45
2	CAL2	1	Water			7.00	01/11/2018	13:47
3	CAL3	1	Water			10.01	01/11/2018	13:49
4	ICV	1	Water			7.00	01/11/2018	13:51
5	CCV1	1	Water			2.01	01/11/2018	13:53
6	J1078-05	1	Solid	20	20	8.88	01/11/2018	13:55
7	J1078-05DUP	1	Solid	20	20	8.89	01/11/2018	13:57
8	J1088-04	1	Solid	20	20	5.90	01/11/2018	13:59
9	J1098-01	1	Solid	20	20	9.16	01/11/2018	14:01
10	J1098-02	1	Solid	20	20	7.47	01/11/2018	14:03
11	J1098-03	1	Solid	20	20	6.76	01/11/2018	14:05
12	J1098-04	1	Solid	20	20	6.48	01/11/2018	14:07
13	J1100-01	1	Solid	20	20	5.88	01/11/2018	14:09
14	CCV2	1	Water			12.02	01/11/2018	14:12

WORKLIST(Hardcopy Internal Chain)

LB92558

WorkList Name : corrosivity011118

WorkList ID : 107309

Date : 1/11/2018 8:55:53 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2018	Solid	J1078-05	Corrosivity	Cool 4 deg C	WSP001	K22	PB-6-OW(8-32)	01/08/2018	9045C
01/10/2018	Solid	J1088-04	Corrosivity	Cool 4 deg C	PSEG01	J52	TP-COMP	01/10/2018	9045C
01/09/2018	Solid	J1098-01	Corrosivity	Cool 4 deg C	CLEA11	K61	OWS-1	01/09/2018	9045C
01/09/2018	Solid	J1098-02	Corrosivity	Cool 4 deg C	CLEA11	K61	OWS-2	01/09/2018	9045C
01/09/2018	Solid	J1098-03	Corrosivity	Cool 4 deg C	CLEA11	K61	OWS-3	01/09/2018	9045C
01/09/2018	Solid	J1098-04	Corrosivity	Cool 4 deg C	CLEA11	K61	OWS-4	01/09/2018	9045C
01/10/2018	Solid	J1100-01	Corrosivity	Cool 4 deg C	CLEA11	K23	OWS-1	01/10/2018	9045C

Date/Time

11/11/18 1:30 PM

Received by:

SP

Relinquished by:

SP

Date/Time

11/11/18 2:30 PM

Received by:

SP

Relinquished by:

SP