


7471B

LB94562
INSTRUMENT ID: CV2

Linear

 μ Abs.:

45821



A= 0.0000e+000

B= 2.1982e-004

C= -1.8407e-002

Rho= 0.9998922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	35	0.000	35				
0.2	0.200	0.185	-0.015	925	0.0 %	925	0			
2.5	2.500	2.564	0.064	11749	0.0 %	11749				
5.0	5.000	5.009	0.009	22869	0.0 %	22869				
7.5	7.500	7.399	-0.101	33744	0.0 %	33744				
10.0	10.000	10.054	0.054	45821	0.0 %	45821				

LB92562 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Jan 2018 12:50:01

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	7471B	PPB	11 Jan 2018 12:52:08	S	Std
0.2 - 1	50.2	925	-	0.2000	7471B	PPB	11 Jan 2018 12:54:25	S	Std
2.5 - 1	52.5	11749	-	2.5000	7471B	PPB	11 Jan 2018 12:56:43	S	Std
5.0 - 1	55	22869	-	5.0000	7471B	PPB	11 Jan 2018 12:59:00	S	Std
7.5 - 1	57.5	33744	-	7.5000	7471B	PPB	11 Jan 2018 13:01:17	S	Std
10.0 - 1	50	45821	-	10.0000	7471B	PPB	11 Jan 2018 13:06:13	S	Std
ICV43 - 1	ICV43	17256	3.7748	-	7471B	PPB	11 Jan 2018 13:09:03	U	SMPL
ICB43 - 1	ICB43	-90	-0.0382	-	7471B	PPB	11 Jan 2018 13:11:18	U	SMPL
CCV32 - 1	CCV32	21939	4.8042	-	7471B	PPB	11 Jan 2018 13:13:34	U	SMPL
CCB32 - 1	CCB32	-90	-0.0382	-	7471B	PPB	11 Jan 2018 13:15:50	U	SMPL
CRA - 1	CRA	899	0.1792	-	7471B	PPB	11 Jan 2018 13:18:06	U	SMPL
HighStd - 1	HighStd	45189	9.9149	-	7471B	PPB	11 Jan 2018 13:20:22	U	SMPL
ChkStd - 1	ChkStd	30799	6.7518	-	7471B	PPB	11 Jan 2018 13:22:38	U	SMPL
PB105562BL - 1	PBS	-163	-0.0542	-	7471B	PPB	11 Jan 2018 13:24:57	U	SMPL
PB105562BS - 1	LCSS	16014	3.5017	-	7471B	PPB	11 Jan 2018 13:27:14	U	SMPL
J1013-01 - 1	NB-08-010918	160	0.0168	-	7471B	PPB	11 Jan 2018 13:34:23	U	SMPL
J1013-01DUP - 1	NB-08-010918DUP	131	0.0104	-	7471B	PPB	11 Jan 2018 13:36:39	U	SMPL
J1013-01MS - 1	NB-08-010918MS	16618	3.6345	-	7471B	PPB	11 Jan 2018 13:38:57	U	SMPL
J1013-01MSD - 1	NB-08-010918MSD	16655	3.6427	-	7471B	PPB	11 Jan 2018 13:41:15	U	SMPL
J1064-14 - 1	OILY-STONES	129	0.0099	-	7471B	PPB	11 Jan 2018 13:43:31	U	SMPL
CCV33 - 1	CCV33	21935	4.8033	-	7471B	PPB	11 Jan 2018 13:45:47	U	SMPL
CCB33 - 1	CCB33	-111	-0.0428	-	7471B	PPB	11 Jan 2018 13:48:02	U	SMPL
J1067-01 - 1	CHRT-27007	2028	0.4274	-	7471B	PPB	11 Jan 2018 13:50:19	U	SMPL
J1071-01 - 1	NAA-WS	262	0.0392	-	7471B	PPB	11 Jan 2018 13:52:35	U	SMPL
J1072-01 - 1	PAINT-CHIPS-DRUM	3828	0.8231	-	7471B	PPB	11 Jan 2018 13:54:50	U	SMPL
J1073-01 - 1	OILY-DEBRIS-PAINT-CHIPS	651	0.1247	-	7471B	PPB	11 Jan 2018 13:57:06	U	SMPL
J1073-03 - 1	OIL-DRUM	52	-0.0070	-	7471B	PPB	11 Jan 2018 13:59:23	U	SMPL
J1078-02 - 1	PB-6-OW(8-32)	13474	2.9434	-	7471B	PPB	11 Jan 2018 14:01:39	U	SMPL
J1081-01 - 1	EPE-121-1(0-5)	4664	1.0068	-	7471B	PPB	11 Jan 2018 14:03:57	U	SMPL
J1081-02 - 1	EPE-121-1(10-12)	6936	1.5062	-	7471B	PPB	11 Jan 2018 14:06:14	U	SMPL
J1081-03 - 1	EPE-106-1(0-5)	15067	3.2936	-	7471B	PPB	11 Jan 2018 14:08:32	U	SMPL
J1081-04 - 1	EPE-106-1(5-6)	53	-0.0068	-	7471B	PPB	11 Jan 2018 14:10:48	U	SMPL
CCV34 - 1	CCV34	21328	4.6699	-	7471B	PPB	11 Jan 2018 14:13:05	U	SMPL
CCB34 - 1	CCB34	-76	-0.0351	-	7471B	PPB	11 Jan 2018 14:15:21	U	SMPL
J1082-01 - 1	EP-02	183	0.0218	-	7471B	PPB	11 Jan 2018 14:17:37	U	SMPL
J1082-02 - 1	EP-07	223	0.0306	-	7471B	PPB	11 Jan 2018 14:19:54	U	SMPL
J1088-01 - 1	TP-COMP	1250	0.2564	-	7471B	PPB	11 Jan 2018 14:22:10	U	SMPL
PB105563BL - 1	PBS	31	-0.0116	-	7471B	PPB	11 Jan 2018 14:24:26	U	SMPL
PB105563BS - 1	LCSS	15370	3.3602	-	7471B	PPB	11 Jan 2018 14:32:08	U	SMPL
J1076-01 - 1	WC-9871-01	67418	14.8012	-	7471B	PPB	11 Jan 2018 14:34:24	U	SMPL
J1076-02 - 1	WC-7317-03	13388	2.9245	-	7471B	PPB	11 Jan 2018 14:36:41	U	SMPL
J1076-03 - 1	WC-7317-02	37393	8.2012	-	7471B	PPB	11 Jan 2018 14:39:01	U	SMPL
J1076-04 - 1	WC-12R12-01	21240	4.6505	-	7471B	PPB	11 Jan 2018 14:41:19	U	SMPL
J1076-05 - 1	WC-14-01	42364	9.2939	-	7471B	PPB	11 Jan 2018 14:43:36	U	SMPL
CCV35 - 1	CCV35	21166	4.6342	-	7471B	PPB	11 Jan 2018 14:45:53	U	SMPL
CCB35 - 1	CCB35	-116	-0.0439	-	7471B	PPB	11 Jan 2018 14:48:12	U	SMPL
J1076-06 - 1	WC-14R30-01	30710	6.7322	-	7471B	PPB	11 Jan 2018 14:50:30	U	SMPL
J1076-07 - 1	WC-14R30-02	12958	2.8300	-	7471B	PPB	11 Jan 2018 14:52:46	U	SMPL
J1076-08 - 1	WC-12R12-02	34240	7.5081	-	7471B	PPB	11 Jan 2018 14:55:04	U	SMPL
J1076-09 - 1	WC-42-02	22566	4.9420	-	7471B	PPB	11 Jan 2018 14:57:20	U	SMPL
J1076-10 - 1	WC-43-02	19136	4.1880	-	7471B	PPB	11 Jan 2018 14:59:39	U	SMPL
J1076-11 - 1	WC-096613-01	45492	9.9815	-	7471B	PPB	11 Jan 2018 15:01:56	U	SMPL
J1076-12 - 1	WC-096613-02	34416	7.5468	-	7471B	PPB	11 Jan 2018 15:04:13	U	SMPL
J1076-13 - 1	WC-50-01	44874	9.8457	-	7471B	PPB	11 Jan 2018 15:06:33	U	SMPL
J1076-14 - 1	WC-14767779-01	38350	8.4116	-	7471B	PPB	11 Jan 2018 15:08:52	U	SMPL
J1076-15 - 1	WC-11R-01	22836	5.0013	-	7471B	PPB	11 Jan 2018 15:11:10	U	SMPL
CCV36 - 1	CCV36	21483	4.7039	-	7471B	PPB	11 Jan 2018 15:13:28	U	SMPL
CCB36 - 1	CCB36	-106	-0.0417	-	7471B	PPB	11 Jan 2018 15:15:45	U	SMPL
J1076-16 - 1	WC-53-01	19815	4.3373	-	7471B	PPB	11 Jan 2018 15:18:01	U	SMPL
J1076-17 - 1	WC-52-01	22002	4.8180	-	7471B	PPB	11 Jan 2018 15:20:17	U	SMPL
J1076-18 - 1	WC-42-01	54938	12.0579	-	7471B	PPB	11 Jan 2018 15:22:34	U	SMPL
J1076-19 - 1	WC-43-01	34049	7.4662	-	7471B	PPB	11 Jan 2018 15:24:51	U	SMPL
J1076-20 - 1	WC-50-02	70702	15.5231	-	7471B	PPB	11 Jan 2018 15:27:10	U	SMPL
J1076-20DUP - 1	WC-50-02DUP	62855	13.7982	-	7471B	PPB	11 Jan 2018 15:29:29	U	SMPL
J1076-20MS - 1	WC-50-02MS	88560	19.4486	-	7471B	PPB	11 Jan 2018 15:34:58	U	SMPL
J1076-20MSD - 1	WC-50-02MSD	87000	19.1057	-	7471B	PPB	11 Jan 2018 15:37:15	U	SMPL
J1076-01DLX2 - 1		32604	7.1485	-	7471B	PPB	11 Jan 2018 15:39:36	U	SMPL
J1076-18DLX2 - 1		27436	6.0125	-	7471B	PPB	11 Jan 2018 15:41:54	U	SMPL
CCV37 - 1	CCV37	20877	4.5707	-	7471B	PPB	11 Jan 2018 15:44:12	U	SMPL
CCB37 - 1	CCB37	-109	-0.0424	-	7471B	PPB	11 Jan 2018 15:46:31	U	SMPL
J1013-01LX5 - 1		27	-0.0125	-	7471B	PPB	11 Jan 2018 15:48:48	U	SMPL
J1013-01A - 1		16699	3.6523	-	7471B	PPB	11 Jan 2018 15:51:04	U	SMPL
J1076-20LX5 - 1		13634	2.9786	-	7471B	PPB	11 Jan 2018 15:53:21	U	SMPL
J1076-20A - 1		89563	19.6691	-	7471B	PPB	11 Jan 2018 15:58:33	U	SMPL
J1076-20DLX2 - 1		35228	7.7253	-	7471B	PPB	11 Jan 2018 16:00:49	U	SMPL
J1076-20DUPDLX2 - 1		32357	7.0942	-	7471B	PPB	11 Jan 2018 16:03:09	U	SMPL
J1076-20MSDLX2 - 1		44469	9.7567	-	7471B	PPB	11 Jan 2018 16:05:28	U	SMPL
J1076-20MSDDLX2 - 1		43981	9.6494	-	7471B	PPB	11 Jan 2018 16:07:47	U	SMPL
J1076-20ADLX2 - 1		45270	9.9327	-	7471B	PPB	11 Jan 2018 16:12:25	U	SMPL
J1076-20LDLX10 - 1		7102	1.5427	-	7471B	PPB	11 Jan 2018 16:15:01	U	SMPL
CV38 - 1	CCV38	21099	4.6195	-	7471B	PPB	11 Jan 2018 16:17:36	U	SMPL
CB38 - 1	CCB38	-101	-0.0406	-	7471B	PPB	11 Jan 2018 16:19:53	U	SMPL