



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92571

Temp.: 22.3 °C

SLOPE.: 98.0

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/11/2018	16:15
2	CAL2	1	Water			7.01	01/11/2018	16:17
3	CAL3	1	Water			10.01	01/11/2018	16:19
4	ICV	1	Water			7.00	01/11/2018	16:21
5	CCV1	1	Water			2.01	01/11/2018	16:23
6	J1099-01	1	Water			4.28	01/11/2018	16:25
7	J1099-01DUP	1	Water			4.29	01/11/2018	16:27
8	J1105-01	1	Water			4.36	01/11/2018	16:29
9	CCV2	1	Water			12.01	01/11/2018	16:31

WORKLIST(Hardcopy Internal Chain)

LB92571

Worklist Name : Ph011118

Worklist ID : 107332

Date : 1/11/2018 3:55:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2018	Water	J1099-01	pH	Cool 4 deg C	STER16	K22	1 EXCELL	01/08/2018	SM 4500-PH B
01/05/2018	Water	J1105-01	pH	Cool 4 deg C	STER16	K22	2 MAINCELL	01/05/2018	SM 4500-PH B

Date/Time

11/11/18 4:05 PM

Received by:

SY

Relinquished by:

CP

Date/Time

11/11/18 4:45 PM

Received by:

CP

Relinquished by:

SP